

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL01 Date Analyzed: 10/03/2024

Lab Sample No.: PSTDCCC050 Data File : PL092221.D Time Analyzed: 15:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.059	8.960	9.160	49.630	50.000	-0.7
Endrin	6.578	6.479	6.679	43.850	50.000	-12.3
gamma-BHC (Lindane)	4.331	4.232	4.432	49.210	50.000	-1.6
Heptachlor	4.919	4.820	5.020	48.340	50.000	-3.3
Heptachlor epoxide	5.688	5.589	5.789	48.700	50.000	-2.6
Methoxychlor	7.502	7.404	7.604	46.560	50.000	-6.9
Tetrachloro-m-xylene	3.542	3.444	3.644	51.160	50.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL01 Date Analyzed: 10/03/2024

Lab Sample No.: PSTDCCC050 Data File : PL092221.D Time Analyzed: 15:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.921	7.823	8.023	47.460	50.000	-5.1
Endrin	5.646	5.548	5.748	47.570	50.000	-4.9
gamma-BHC (Lindane)	3.614	3.516	3.716	53.910	50.000	7.8
Heptachlor	3.953	3.855	4.055	51.650	50.000	3.3
Heptachlor epoxide	4.736	4.638	4.838	52.500	50.000	5.0
Methoxychlor	6.619	6.521	6.721	46.900	50.000	-6.2
Tetrachloro-m-xylene	2.781	2.682	2.882	53.960	50.000	7.9