

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL02 Date Analyzed: 10/03/2024

Lab Sample No.: PSTDCCC050 Data File : PL092229.D Time Analyzed: 18:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.058	8.960	9.160	50.740	50.000	1.5
Endrin	6.577	6.479	6.679	44.550	50.000	-10.9
gamma-BHC (Lindane)	4.331	4.232	4.432	49.990	50.000	0.0
Heptachlor	4.920	4.820	5.020	48.750	50.000	-2.5
Heptachlor epoxide	5.688	5.589	5.789	51.430	50.000	2.9
Methoxychlor	7.502	7.404	7.604	46.940	50.000	-6.1
Tetrachloro-m-xylene	3.542	3.444	3.644	51.560	50.000	3.1

CALIBRATION VERIFICATION SUMMARY

Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL02 Date Analyzed: 10/03/2024

Lab Sample No.: PSTDCCC050 Data File : PL092229.D Time Analyzed: 18:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.920	7.823	8.023	50.700	50.000	1.4
Endrin	5.646	5.548	5.748	48.820	50.000	-2.4
gamma-BHC (Lindane)	3.614	3.516	3.716	54.820	50.000	9.6
Heptachlor	3.953	3.855	4.055	51.780	50.000	3.6
Heptachlor epoxide	4.736	4.638	4.838	53.650	50.000	7.3
Methoxychlor	6.618	6.521	6.721	48.280	50.000	-3.4
Tetrachloro-m-xylene	2.781	2.682	2.882	53.800	50.000	7.6