

Surrogate Summary

SDG No.: **P4258**

Client: **Roman E&G Corp**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL091953.D	PIBLK-PL091953.D	Decachlorobiphenyl	1	20	19.6	98		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	18.7	94		43	140
		Tetrachloro-m-xylene	2	20	18.3	92		77	126
I.BLK-PL092220.D	PIBLK-PL092220.D	Decachlorobiphenyl	1	20	20.1	100		43	140
		Tetrachloro-m-xylene	1	20	20.0	100		77	126
		Decachlorobiphenyl	2	20	18.8	94		43	140
		Tetrachloro-m-xylene	2	20	20.0	100		77	126
PB163891BL	PB163891BL	Decachlorobiphenyl	1	20	19.0	95		43	140
		Tetrachloro-m-xylene	1	20	17.9	90		77	126
		Decachlorobiphenyl	2	20	18.0	90		43	140
		Tetrachloro-m-xylene	2	20	17.1	86		77	126
PB163891BS	PB163891BS	Decachlorobiphenyl	1	20	19.3	96		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	18.5	93		43	140
		Tetrachloro-m-xylene	2	20	17.9	90		77	126
PB163851TB	PB163851TB	Decachlorobiphenyl	1	20	20.5	103		43	140
		Tetrachloro-m-xylene	1	20	19.5	97		77	126
		Decachlorobiphenyl	2	20	19.7	99		43	140
		Tetrachloro-m-xylene	2	20	19.2	96		77	126
P4258-04	Chamberlain Ave	Decachlorobiphenyl	1	20	9.58	48		43	140
		Tetrachloro-m-xylene	1	20	17.7	88		77	126
		Decachlorobiphenyl	2	20	9.39	47		43	140
		Tetrachloro-m-xylene	2	20	18.3	91		77	126
P4258-04MS	Chamberlain AveMS	Decachlorobiphenyl	1	20	13.9	69		43	140
		Tetrachloro-m-xylene	1	20	16.2	81		77	126
		Decachlorobiphenyl	2	20	13.4	67		43	140
		Tetrachloro-m-xylene	2	20	16.5	83		77	126
P4258-04MSD	Chamberlain AveMSD	Decachlorobiphenyl	1	20	13.9	70		43	140
		Tetrachloro-m-xylene	1	20	16.8	84		77	126
		Decachlorobiphenyl	2	20	13.4	67		43	140
		Tetrachloro-m-xylene	2	20	16.8	84		77	126
I.BLK-PL092228.D	PIBLK-PL092228.D	Decachlorobiphenyl	1	20	21.0	105		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	20.3	102		43	140
		Tetrachloro-m-xylene	2	20	19.9	100		77	126