

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ROMA02

Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG No.: P4258

Instrument ID: MSVOA_N Calibration Date/Time: 10/03/2024 08:43

Lab File ID: VN084283.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.654	0.623		-4.74	20
1,1-Dichloroethene	0.563	0.545		-3.2	20
2-Butanone	0.434	0.451		3.92	20
Carbon Tetrachloride	0.525	0.548		4.38	20
Chloroform	1.175	1.187		1.02	20
Benzene	1.492	1.529		2.48	20
1,2-Dichloroethane	0.506	0.529		4.55	20
Trichloroethene	0.348	0.350		0.57	20
Tetrachloroethene	0.348	0.348		0	20
Chlorobenzene	1.109	1.084	0.3	-2.25	20
1,2-Dichloroethane-d4	0.741	0.758		2.29	20
Dibromofluoromethane	0.330	0.342		3.64	20
Toluene-d8	1.213	1.304		7.5	20
4-Bromofluorobenzene	0.442	0.468		5.88	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.