

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4258
Client: Roman E&G Corp
Analytical Method: SW8260-Low **Datafile :** VN084286.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1003WBS01	Vinyl chloride	20	19.7	ug/L	99			65	117	
	1,1-Dichloroethene	20	19.6	ug/L	98			74	110	
	2-Butanone	100	98.2	ug/L	98			65	122	
	Carbon Tetrachloride	20	20.1	ug/L	101			77	113	
	Chloroform	20	20.6	ug/L	103			79	113	
	Benzene	20	20.2	ug/L	101			82	109	
	1,2-Dichloroethane	20	20.4	ug/L	102			80	115	
	Trichloroethene	20	20.0	ug/L	100			77	113	
	Tetrachloroethene	20	20.2	ug/L	101			67	123	
	Chlorobenzene	20	19.7	ug/L	99			82	109	