

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4258
Client: Roman E&G Corp
Analytical Method: SW8260-Low **Datafile :** VN084287.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1003WBSD01	Vinyl chloride	20	19.7	ug/L	99	0		65	117	20
	1,1-Dichloroethene	20	19.5	ug/L	98	0		74	110	20
	2-Butanone	100	100	ug/L	100	2		65	122	20
	Carbon Tetrachloride	20	20.3	ug/L	102	1		77	113	20
	Chloroform	20	20.7	ug/L	104	1		79	113	20
	Benzene	20	20.8	ug/L	104	3		82	109	20
	1,2-Dichloroethane	20	21.7	ug/L	109	7		80	115	20
	Trichloroethene	20	19.9	ug/L	100	0		77	113	20
	Tetrachloroethene	20	20.3	ug/L	102	1		67	123	20
	Chlorobenzene	20	20.4	ug/L	102	3		82	109	20