

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ROMA02
Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258
Lab File ID: VN084283.D Date Analyzed: 10/03/2024
Instrument ID: MSVOA_N Time Analyzed: 08:43
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	169381	8.22	285966	9.10	262806	11.87
UPPER LIMIT	338762	8.724	571932	9.6	525612	12.365
LOWER LIMIT	84690.5	7.724	142983	8.6	131403	11.365
EPA SAMPLE NO.						
Chamberlain Ave	164187	8.22	302561	9.10	265786	11.87
VN1003WBL01	163611	8.22	286487	9.10	249556	11.87
VN1003WBS01	164341	8.22	282487	9.10	252092	11.87
VN1003WBSD01	156141	8.22	268205	9.10	240791	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.