

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY090924

Review By	Mahesh Dadoda	Review On	9/10/2024 10:25:17 AM
Supervise By	Semsettin Yesilyurt	Supervise On	9/10/2024 10:27:41 AM
SubDirectory	VY090924	HP Acquire Method	MSVOA_Y HP Processing Method 82y090924s.m

STD. NAME	STD REF.#
Tune/Reschk	VP130172
Initial Calibration Stds	VP130173,VP130174,VP130175,VP130176,VP130177,VP130178
CCC	VP130180,VP130181
Internal Standard/PEM	VP128297
ICV/I.BLK	VP130179
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY019453.D	09 Sep 2024 09:20		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY019454.D	09 Sep 2024 09:52	Good for DOD	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY019455.D	09 Sep 2024 10:15	LR-02,13	SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY019456.D	09 Sep 2024 10:44	%D failed for Comp. #13 in 20 PPB	SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY019457.D	09 Sep 2024 11:07		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY019458.D	09 Sep 2024 11:29		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY019459.D	09 Sep 2024 11:53		SY/MD	Ok,M
8	VIBLK	VIBLK	VY019460.D	09 Sep 2024 12:16		SY/MD	Ok
9	VSTDICV050	ICVVY090924	VY019461.D	09 Sep 2024 13:05		SY/MD	Ok,M
10	VY0909SBL01	VY0909SBL01	VY019462.D	09 Sep 2024 13:34		SY/MD	Ok
11	VY0909SBS01	VY0909SBS01	VY019463.D	09 Sep 2024 13:57		SY/MD	Ok,M
12	VY0909SBS01	VY0909SBS01	VY019464.D	09 Sep 2024 14:26		SY/MD	Ok,M
13	P3891-01	AUD-SBG	VY019465.D	09 Sep 2024 15:10	Vial A	SY/MD	Ok
14	P3852-02	VNJ223	VY019466.D	09 Sep 2024 15:33	Vial A	SY/MD	Ok
15	P3852-04	ETGI356	VY019467.D	09 Sep 2024 15:57	Vial A Internal Standard Fail	SY/MD	ReRun
16	P3857-03RE	WC-C1-GRABRE	VY019468.D	09 Sep 2024 16:20	Internal Standard Fail Vial B	SY/MD	Confirms
17	P3892-01	ARS20-005	VY019469.D	09 Sep 2024 16:43	Internal Standard Fail Vial A	SY/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY090924

Review By	Mahesh Dadoda	Review On	9/10/2024 10:25:17 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/10/2024 10:27:41 AM		
SubDirectory	VY090924	HP Acquire Method	MSVOA_Y	HP Processing Method	82y090924s.m
STD. NAME	STD REF.#				
Tune/Reschk	VP130172				
Initial Calibration Stds	VP130173,VP130174,VP130175,VP130176,VP130177,VP130178				
CCC	VP130180,VP130181				
Internal Standard/PEM	VP128297				
ICV/I.BLK	VP130179				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P3905-05	COMP-1	VY019470.D	09 Sep 2024 17:07	Vial A	SY/MD	Ok
19	P3905-03	TP-3B	VY019471.D	09 Sep 2024 17:30	Vial A	SY/MD	Ok
20	P3905-01	TP-3A	VY019472.D	09 Sep 2024 17:54	Vial A	SY/MD	Ok
21	P3906-15	SU-701-VOC-03	VY019473.D	09 Sep 2024 18:17	Internal Standard Fail Vial A	SY/MD	ReRun
22	VSTDCCC050	VSTDCCC050EC	VY019474.D	09 Sep 2024 18:40		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100224

Review By	Semsettin Yesilyurt	Review On	10/3/2024 4:56:16 PM
Supervise By	Maresh Dadoda	Supervise On	10/3/2024 4:58:28 PM
SubDirectory	VY100224	HP Acquire Method	MSVOA_Y HP Processing Method 82y090924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP130628
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130629,VP130630 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY019755.D	02 Oct 2024 08:37		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY019756.D	02 Oct 2024 09:39		SY/MD	Ok,M
3	VY1002SBL01	VY1002SBL01	VY019757.D	02 Oct 2024 10:13		SY/MD	Ok
4	VY1002SBS01	VY1002SBS01	VY019758.D	02 Oct 2024 12:36		SY/MD	Not Ok
5	P4225-01	TR-04-092724	VY019759.D	02 Oct 2024 13:03	NOT PURGE vial-B	SY/MD	Not Ok
6	P4236-06RE	TP3-VOCRE	VY019760.D	02 Oct 2024 13:26	Internal Standard Fail vial-B	SY/MD	Confirms
7	P4236-10	TP1-VOC	VY019761.D	02 Oct 2024 13:49	NOT PURGE vial-B	SY/MD	Not Ok
8	P4236-14RE	TP2-VOCRE	VY019762.D	02 Oct 2024 14:13	Internal Standard Fail vial-B	SY/MD	Confirms
9	P4236-02	TP4-VOC	VY019763.D	02 Oct 2024 14:36	NOT PURGE vial-B	SY/MD	Not Ok
10	P4232-03	TP-R-VOC	VY019764.D	02 Oct 2024 15:00	vial-A	SY/MD	Ok
11	P4258-03	Chamberlain Ave	VY019765.D	02 Oct 2024 15:23	Internal Standard Fail vial-A	SY/MD	ReRun
12	P4242-03	TP-Q-VOC	VY019766.D	02 Oct 2024 15:47	Internal Standard Fail vial-A	SY/MD	ReRun
13	P4257-02	VNJ-240	VY019767.D	02 Oct 2024 16:10	vial-A	SY/MD	Ok
14	P4255-01	OR-03-100124	VY019768.D	02 Oct 2024 16:33	vial-A	SY/MD	Ok
15	VY1002SBS02	VY1002SBS02	VY019769.D	02 Oct 2024 16:56		SY/MD	Ok,M
16	VY1002SBSD02	VY1002SBSD02	VY019770.D	02 Oct 2024 17:19		SY/MD	Ok,M
17	P4254-02	MH-147-SLUDGE-VOC	VY019771.D	02 Oct 2024 17:42	Internal Standard Fail vial-A	SY/MD	Ok
18	P4254-01	MH-147-SLUDGE	VY019772.D	02 Oct 2024 18:06	Not purge vial-A	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100224

Review By	Semsettin Yesilyurt	Review On	10/3/2024 4:56:16 PM		
Supervise By	Mahesh Dadoda	Supervise On	10/3/2024 4:58:28 PM		
SubDirectory	VY100224	HP Acquire Method	MSVOA_Y	HP Processing Method	82y090924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP130628
CCC	VP130629,VP130630
Internal Standard/PEM	VP128297
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VIBLK	VIBLK	VY019773.D	02 Oct 2024 18:29		SY/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VY019774.D	02 Oct 2024 18:52		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100324

Review By	Semsettin Yesilyurt	Review On	10/4/2024 10:17:50 AM
Supervise By	Maresh Dadoda	Supervise On	10/4/2024 10:18:45 AM
SubDirectory	VY100324	HP Acquire Method	HP Processing Method

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP130673
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130675,VP130676 VP130432

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY019775.D	03 Oct 2024 08:34		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY019776.D	03 Oct 2024 09:44		SY/MD	Ok,M
3	VY1003SBL01	VY1003SBL01	VY019777.D	03 Oct 2024 10:17		SY/MD	Ok
4	VY1003SBS01	VY1003SBS01	VY019778.D	03 Oct 2024 11:01		SY/MD	Ok,M
5	VY1003SBSD01	VY1003SBSD01	VY019779.D	03 Oct 2024 11:29	Recovery fail	SY/MD	Not Ok
6	P4258-03RE	Chamberlain AveRE	VY019780.D	03 Oct 2024 12:10	Internal Standard Fail vial-B	SY/MD	Confirms
7	P4242-03	TP-Q-VOC	VY019781.D	03 Oct 2024 12:34	vial-B	SY/MD	Ok
8	P4277-01	BU-03-10022024	VY019782.D	03 Oct 2024 12:57	NOT PURGE vial-a	SY/MD	Not Ok
9	P4277-02	BU-03-10022024	VY019783.D	03 Oct 2024 13:21	Not purge vial-A	SY/MD	Not Ok
10	P4276-01	VNJ-208	VY019784.D	03 Oct 2024 13:44	vial-A	SY/MD	Ok
11	P4273-02	AU-05-100224	VY019785.D	03 Oct 2024 14:07	Internal Standard Fail vial-A	SY/MD	ReRun
12	P4275-01	TRE-1300	VY019786.D	03 Oct 2024 14:31	Need MeOH vial-a	SY/MD	Dilution
13	P4254-01	MH-147-SLUDGE	VY019787.D	03 Oct 2024 14:54	Internal standard fail	SY/MD	Ok
14	P4254-02RE	MH-147-SLUDGE-VOC	VY019788.D	03 Oct 2024 15:18	Internal standard fail;Confirm Concentration vial-B	SY/MD	Not Ok
15	P4270-03	IB-2-VOC	VY019789.D	03 Oct 2024 15:41	Confirm Concentration vial-a	SY/MD	Not Ok
16	VSTDCCC050	VSTDCCC050EC	VY019790.D	03 Oct 2024 16:41		SY/MD	Ok,M

M : Manual Integration