

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ROMA02
Lab Code: CHEM Case No.: P4258 SAS No.: P4258 SDG NO.: P4258
Lab File ID: VY019776.D Date Analyzed: 10/03/2024
Instrument ID: MSVOA_Y Time Analyzed: 09:44
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	277386	7.71	478984	8.62	402421	11.41
UPPER LIMIT	554772	8.213	957968	9.116	804842	11.914
LOWER LIMIT	138693	7.213	239492	8.116	201211	10.914
EPA SAMPLE NO.						
Chamberlain AveRE	82774 *	7.71	150343 *	8.62	119106 *	11.41
VY1003SBL01	267988	7.71	512122	8.62	450147	11.41
VY1003SBS01	265491	7.71	467884	8.62	391562	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.