

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard			
Chk Standard	MP82455,MP82456		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/18/24 11:34		Kareem	OK
2	S1	S01	CAL2	10/18/24 11:39		Kareem	OK
3	S2	S02	CAL3	10/18/24 11:44		Kareem	OK
4	S3	S03	CAL4	10/18/24 11:48		Kareem	OK
5	S4	S04	CAL5	10/18/24 11:52		Kareem	OK
6	S5	S05	CAL6	10/18/24 11:56		Kareem	OK
7	S6	S06	CAL7	10/18/24 12:01		Kareem	OK
8	ICV003	ICV003	ICV	10/18/24 12:05		Kareem	OK
9	ICB003	ICB003	ICB	10/18/24 12:11	Inst. paused for review	Kareem	OK
10	ICSA003	ICSA003	ICSA	10/18/24 12:15		Kareem	OK
11	ICSAB003	ICSAB003	ICSAB	10/18/24 13:03	Inst. paused for review	Kareem	OK
12	CCV012	CCV012	CCV	10/18/24 13:07		Kareem	OK
13	CCB012	CCB012	CCB	10/18/24 13:22	Inst. paused for review	Kareem	OK
14	PB164193BL	PBS193	MB	10/18/24 13:27		Kareem	OK
15	PB164193BS	LCS193	LCS	10/18/24 13:31		Kareem	OK
16	P4411-07	MJHJ05	SAM	10/18/24 13:36		Kareem	OK
17	P4411-09	MJHJ05S	MS	10/18/24 13:40		Kareem	OK
18	P4411-08	MJHJ05D	DUP	10/18/24 13:45		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442
ICV Standard	MP82450
CCV Standard	MP82453
ICSA Standard	MP82451,MP82452
CRI Standard	
LCS Standard	
Chk Standard	MP82455,MP82456

19	P4411-01	MJHHZ8	SAM	10/18/24 13:49		Kareem	OK
20	P4411-02	MJHHZ9	SAM	10/18/24 13:53		Kareem	OK
21	P4411-03	MJHJ01	SAM	10/18/24 14:05		Kareem	OK
22	P4411-04	MJHJ02	SAM	10/18/24 14:09		Kareem	OK
23	P4411-05	MJHJ03	SAM	10/18/24 14:13		Kareem	OK
24	P4411-06	MJHJ04	SAM	10/18/24 14:18		Kareem	OK
25	P4411-07L	MJHJ05L	SD	10/18/24 14:30		Kareem	OK
26	CCV013	CCV013	CCV	10/18/24 14:34		Kareem	OK
27	CCB013	CCB013	CCB	10/18/24 14:39		Kareem	OK
28	P4248-03DL	MYD9Q3	SAM	10/18/24 14:59	Straight 2x Dilution for all elements	Kareem	OK
29	P4248-04DL	MYD9Q4	SAM	10/18/24 15:03	Straight 2x Dilution for all elements	Kareem	OK
30	P4248-05DL	MYD9Q5	SAM	10/18/24 15:07	Straight 2x Dilution for all elements	Kareem	OK
31	P4248-06DL	MYD9Q6	SAM	10/18/24 15:12	Straight 2x Dilution for all elements	Kareem	OK
32	P4248-07DL	MYD9Q7	SAM	10/18/24 15:16	Straight 2x Dilution for all elements	Kareem	OK
33	P4248-08DL	MYD9Q8	SAM	10/18/24 15:21	Straight 2x Dilution for all elements	Kareem	OK
34	P4248-09DL	MYD9Q9	SAM	10/18/24 15:25	Straight 2x Dilution for all elements	Kareem	OK
35	P4248-10DL	MYD9R0	SAM	10/18/24 15:29	Straight 2x Dilution for all elements	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442
ICV Standard	MP82450
CCV Standard	MP82453
ICSA Standard	MP82451,MP82452
CRI Standard	
LCS Standard	
Chk Standard	MP82455,MP82456

36	P4248-11DL	MYD9R1	SAM	10/18/24 15:34	Straight 2x Dilution for all elements	Kareem	OK
37	P4248-12DL	MYD9R2	SAM	10/18/24 15:38	Straight 2x Dilution for all elements	Kareem	OK
38	P4248-13DL	MYD9R3	SAM	10/18/24 15:42	Straight 2x Dilution for all elements	Kareem	OK
39	P4248-14DL	MYD9R4	SAM	10/18/24 15:47	Straight 2x Dilution for all elements	Kareem	OK
40	P4248-15DL	MYD9R5	SAM	10/18/24 15:51	Straight 2x Dilution for all elements	Kareem	OK
41	P4248-16DL	MYD9R6	SAM	10/18/24 15:56	Straight 2x Dilution for all elements	Kareem	OK
42	P4248-17DL	MYD9R7	SAM	10/18/24 16:00	Straight 2x Dilution for all elements	Kareem	OK
43	P4248-19DL	MYD9R9	SAM	10/18/24 16:09	Straight 2x Dilution for all elements	Kareem	OK
44	P4248-20DL	MYD9S0	SAM	10/18/24 16:13	Straight 2x Dilution for all elements	Kareem	OK
45	P4248-18DL	MYD9R8	SAM	10/18/24 16:20	Straight 2x Dilution for all elements	Kareem	OK
46	P4248-20LDL	MYD9S0L	SD	10/18/24 16:24	Straight 10x Dilution for all elements	Kareem	OK
47	P4248-21DL	MYD9S0D	DUP	10/18/24 16:29	Straight 2x Dilution for all elements	Kareem	OK
48	CCV014	CCV014	CCV	10/18/24 16:33		Kareem	OK
49	CCB014	CCB014	CCB	10/18/24 16:37		Kareem	OK
50	P4248-22DL	MYD9S0S	MS	10/18/24 16:42	MS fail for As (139.20 ppb), Pb (124.83 ppb), Se,Tl (Below RL), Straight 2x Dilution for all elements	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard			
Chk Standard	MP82455,MP82456		

51	P4291-01DL	MYD0B5	SAM	10/18/24 16:46	NOT USE	Kareem	Not Ok
52	P4291-02DL	MYD0B6	SAM	10/18/24 16:50	Straight 2x Dilution for all elements	Kareem	OK
53	P4291-03DL	MYD0B7	SAM	10/18/24 16:55	Straight 2x Dilution for all elements	Kareem	OK
54	P4291-04DL	MYD0B8	SAM	10/18/24 16:59	Straight 2x Dilution for all elements	Kareem	OK
55	P4291-05DL	MYD0B8D	DUP	10/18/24 17:04	Straight 2x Dilution for all elements	Kareem	OK
56	P4291-04LDL	MYD0B8L	SD	10/18/24 17:08	Straight 10x Dilution for all elements	Kareem	OK
57	P4291-06DL	MYD0B8S	MS	10/18/24 17:13	Straight 2x Dilution for all elements	Kareem	OK
58	P4291-07DL	MYD0B9	SAM	10/18/24 17:17	Straight 2x Dilution for all elements	Kareem	OK
59	P4291-08DL	MYD0C0	SAM	10/18/24 17:21	Straight 2x Dilution for all elements	Kareem	OK
60	P4291-09DL	MYD0C1	SAM	10/18/24 17:26	Straight 2x Dilution for all elements	Kareem	OK
61	P4291-10DL	MYD0C2	SAM	10/18/24 17:30	Straight 2x Dilution for all elements	Kareem	OK
62	P4291-11DL	MYD0C3	SAM	10/18/24 17:35	Straight 2x Dilution for all elements	Kareem	OK
63	P4291-12DL	MYD0C4	SAM	10/18/24 17:39	Straight 2x Dilution for all elements	Kareem	OK
64	P4291-13DL	MYD0C5	SAM	10/18/24 17:43	Straight 2x Dilution for all elements	Kareem	OK
65	P4291-14DL	MYD0C6	SAM	10/18/24 17:48	Straight 2x Dilution for all elements	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442
ICV Standard	MP82450
CCV Standard	MP82453
ICSA Standard	MP82451,MP82452
CRI Standard	
LCS Standard	
Chk Standard	MP82455,MP82456

66	P4291-15DL	MYD0C7	SAM	10/18/24 17:52	Straight 2x Dilution for all elements	Kareem	OK
67	P4291-16DL	MYD0C8	SAM	10/18/24 17:57	Straight 2x Dilution for all elements	Kareem	OK
68	P4291-17DL	MYD0C9	SAM	10/18/24 18:01	Straight 2x Dilution for all elements	Kareem	OK
69	P4291-18DL	MYD0D0	SAM	10/18/24 18:05	Straight 2x Dilution for all elements	Kareem	OK
70	CCV015	CCV015	CCV	10/18/24 18:10		Kareem	OK
71	CCB015	CCB015	CCB	10/18/24 18:14		Kareem	OK
72	P4291-19DL	MYD0D1	SAM	10/18/24 18:19	Straight 2x Dilution for all elements	Kareem	OK
73	P4291-20DL	MYD0D2	SAM	10/18/24 18:23	Straight 2x Dilution for all elements	Kareem	OK
74	P4291-21DL	MYD0D3	SAM	10/18/24 18:27	Straight 2x Dilution for all elements	Kareem	OK
75	P4291-22DL	MYD0D4	SAM	10/18/24 18:32	Straight 2x Dilution for all elements	Kareem	OK
76	P4292-01DL	MYD0D5	SAM	10/18/24 18:36	Straight 2x Dilution for all elements	Kareem	OK
77	P4292-02DL	MYD0S9	SAM	10/18/24 18:41	Straight 2x Dilution for all elements	Kareem	OK
78	P4292-03DL	MYD0T0	SAM	10/18/24 18:45	Straight 2x Dilution for all elements	Kareem	OK
79	P4292-04DL	MYD0T1	SAM	10/18/24 18:49	Straight 2x Dilution for all elements	Kareem	OK
80	P4292-05DL	MYD0T2	SAM	10/18/24 18:54	Straight 2x Dilution for all elements	Kareem	OK
81	P4292-06DL	MYD0T3	SAM	10/18/24 18:58	Straight 2x Dilution for all elements	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442
ICV Standard	MP82450
CCV Standard	MP82453
ICSA Standard	MP82451,MP82452
CRI Standard	
LCS Standard	
Chk Standard	MP82455,MP82456

82	P4292-07DL	MYD0T4	SAM	10/18/24 19:03	Straight 2x Dilution for all elements	Kareem	OK
83	P4292-08DL	MYD0T5	SAM	10/18/24 19:07	Straight 2x Dilution for all elements	Kareem	OK
84	P4292-09DL	MYD0T6	SAM	10/18/24 19:11	Straight 2x Dilution for all elements	Kareem	OK
85	P4292-10DL	MYD0T7	SAM	10/18/24 19:16	Straight 2x Dilution for all elements	Kareem	OK
86	P4292-11DL	MYD0T8	SAM	10/18/24 19:20	Straight 2x Dilution for all elements	Kareem	OK
87	P4292-12DL	MYD0T9	SAM	10/18/24 19:25	Straight 2x Dilution for all elements	Kareem	OK
88	P4292-13DL	MYD0W0	SAM	10/18/24 19:29	Straight 2x Dilution for all elements	Kareem	OK
89	P4292-14DL	MYD0W1	SAM	10/18/24 19:33	Straight 2x Dilution for all elements	Kareem	OK
90	P4292-15DL	MYD0W2	SAM	10/18/24 19:38	Straight 2x Dilution for all elements	Kareem	OK
91	P4292-16DL	MYD0W2D	DUP	10/18/24 19:42	Straight 2x Dilution for all elements	Kareem	OK
92	CCV016	CCV016	CCV	10/18/24 19:47		Kareem	OK
93	CCB016	CCB016	CCB	10/18/24 19:51		Kareem	OK
94	P4292-15LDL	MYD0W2L	SD	10/18/24 19:55	Straight 10x Dilution for all elements	Kareem	OK
95	P4292-17DL	MYD0W2S	MS	10/18/24 20:00	MS fail for As (287.18 ppb) , Straight 2x Dilution for all elements	Kareem	OK
96	P4292-18DL	MYD0W3	SAM	10/18/24 20:04	Straight 2x Dilution for all elements	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442
ICV Standard	MP82450
CCV Standard	MP82453
ICSA Standard	MP82451,MP82452
CRI Standard	
LCS Standard	
Chk Standard	MP82455,MP82456

97	P4292-19DL	MYD0W4	SAM	10/18/24 20:09	Mn high, Straight 2x Dilution for all elements	Kareem	Dilution
98	P4292-20DL	MYD0W5	SAM	10/18/24 20:13	Straight 2x Dilution for all elements	Kareem	OK
99	P4292-21DL	MYD0W6	SAM	10/18/24 20:17	Straight 2x Dilution for all elements	Kareem	OK
100	P4292-22DL	MYD0W7	SAM	10/18/24 20:22	Straight 2x Dilution for all elements	Kareem	OK
101	P4251-01DL	MYD9W1	SAM	10/18/24 20:26	Straight 2x Dilution for all elements	Kareem	OK
102	P4251-02DL	MYD9W2	SAM	10/18/24 20:30	Straight 2x Dilution for all elements	Kareem	OK
103	P4251-03DL	MYD9W3	SAM	10/18/24 20:35	Straight 2x Dilution for all elements	Kareem	OK
104	P4251-04DL	MYD9W4	SAM	10/18/24 20:39	Straight 2x Dilution for all elements	Kareem	OK
105	P4251-05DL	MYD9W5	SAM	10/18/24 20:44	Straight 2x Dilution for all elements	Kareem	OK
106	P4251-06DL	MYD9W6	SAM	10/18/24 20:48	Straight 2x Dilution for all elements	Kareem	OK
107	P4251-07DL	MYD9W7	SAM	10/18/24 20:52	Straight 2x Dilution for all elements	Kareem	OK
108	P4251-08DL	MYD9W8	SAM	10/18/24 20:57	Straight 2x Dilution for all elements	Kareem	OK
109	P4251-09DL	MYD9W9	SAM	10/18/24 21:01	Straight 2x Dilution for all elements	Kareem	OK
110	P4251-10DL	MYD9W9D	DUP	10/18/24 21:05	Straight 2x Dilution for all elements	Kareem	OK
111	P4251-09LDL	MYD9W9L	SD	10/18/24 21:10	Straight 10x Dilution for all elements	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard			
Chk Standard	MP82455,MP82456		

112	P4251-11DL	MYD9W9S	MS	10/18/24 21:14	MS fail for Zn (434.28 ppb), Se (Below RL), Straight 2x Dilution for all elements	Kareem	OK
113	P4251-12DL	MYD9Y8	SAM	10/18/24 21:19	Straight 2x Dilution for all elements	Kareem	OK
114	CCV017	CCV017	CCV	10/18/24 21:23		Kareem	OK
115	CCB017	CCB017	CCB	10/18/24 21:27		Kareem	OK
116	P4251-13DL	MYD9Y9	SAM	10/18/24 21:32	NOT USE	Kareem	Not Ok
117	P4251-14DL	MYD9Z0	SAM	10/18/24 21:36	NOT USE	Kareem	Not Ok
118	P4251-15DL	MYD9Z1	SAM	10/18/24 21:41	NOT USE	Kareem	Not Ok
119	P4251-16DL	MYD9Z2	SAM	10/18/24 21:45	NOT USE	Kareem	Not Ok
120	P4251-17DL	MYD9Z3	SAM	10/18/24 21:49	NOT USE	Kareem	Not Ok
121	P4251-18DL	MYD9Z4	SAM	10/18/24 21:54	NOT USE	Kareem	Not Ok
122	P4251-19DL	MYD9Z5	SAM	10/18/24 21:58	NOT USE	Kareem	Not Ok
123	P4251-20DL	MYD9Z6	SAM	10/18/24 22:02	NOT USE	Kareem	Not Ok
124	P4251-21DL	MYD9Z7	SAM	10/18/24 22:07	NOT USE	Kareem	Not Ok
125	P4251-22DL	MYD9Z8	SAM	10/18/24 22:11	NOT USE	Kareem	Not Ok
126	P4250-01DL	MYD9S1	SAM	10/18/24 22:16	NOT USE	Kareem	Not Ok
127	P4250-02DL	MYD9S2	SAM	10/18/24 22:20	NOT USE	Kareem	Not Ok
128	P4250-03DL	MYD9S3	SAM	10/18/24 22:24	NOT USE	Kareem	Not Ok
129	P4250-04DL	MYD9S4	SAM	10/18/24 22:29	NOT USE	Kareem	Not Ok
130	P4250-05DL	MYD9S5	SAM	10/18/24 22:33	NOT USE	Kareem	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard			
Chk Standard	MP82455,MP82456		

131	P4250-06DL	MYD9S6	SAM	10/18/24 22:38	NOT USE	Kareem	Not Ok
132	P4250-07DL	MYD9S7	SAM	10/18/24 22:42	NOT USE	Kareem	Not Ok
133	P4250-08DL	MYD9S8	SAM	10/18/24 22:46	NOT USE	Kareem	Not Ok
134	P4250-09DL	MYD9S9	SAM	10/18/24 22:51	NOT USE	Kareem	Not Ok
135	P4250-10DL	MYD9T0	SAM	10/18/24 22:55	NOT USE	Kareem	Not Ok
136	CCV018	CCV018	CCV	10/18/24 23:00	%RSD fail for Al,Ba,be,Ca,Cu,Fe,Mg,Mn, K,Na,V	Kareem	OK
137	CCB018	CCB018	CCB	10/18/24 23:04		Kareem	OK
138	P4250-11DL	MYD9T1	SAM	10/18/24 23:08	NOT USE	Kareem	Not Ok
139	P4250-12DL	MYD9T2	SAM	10/18/24 23:13	NOT USE	Kareem	Not Ok
140	P4250-13DL	MYD9T3	SAM	10/18/24 23:17	NOT USE	Kareem	Not Ok
141	P4250-14DL	MYD9T4	SAM	10/18/24 23:21	NOT USE	Kareem	Not Ok
142	P4250-15DL	MYD9T5	SAM	10/18/24 23:26	NOT USE	Kareem	Not Ok
143	P4250-16DL	MYD9T6	SAM	10/18/24 23:30	NOT USE	Kareem	Not Ok
144	P4250-17DL	MYD9T7	SAM	10/18/24 23:35	NOT USE	Kareem	Not Ok
145	P4250-18DL	MYD9T7D	DUP	10/18/24 23:39	NOT USE	Kareem	Not Ok
146	P4250-17LDL	MYD9T7L	SD	10/18/24 23:43	NOT USE	Kareem	Not Ok
147	P4250-19DL	MYD9T7S	MS	10/18/24 23:48	NOT USE	Kareem	Not Ok
148	P4250-20DL	MYD9T8	SAM	10/18/24 23:52	NOT USE	Kareem	Not Ok
149	P4250-21DL	MYD9T9	SAM	10/18/24 23:57	NOT USE	Kareem	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133008

Review By	Mohan Bera	Review On	10/21/2024 10:52:55 PM
Supervise By	Sarabjit Jaswal	Supervise On	10/22/2024 1:54:03 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard			
Chk Standard	MP82455,MP82456		

150	P4250-22DL	MYD9W0	SAM	10/19/24 00:01	NOT USE	Kareem	Not Ok
151	P4252-01DL	MYD9X0	SAM	10/19/24 00:05	NOT USE	Kareem	Not Ok
152	P4252-02DL	MYD9X1	SAM	10/19/24 00:10	NOT USE	Kareem	Not Ok
153	P4252-03DL	MYD9X2	SAM	10/19/24 00:14	NOT USE	Kareem	Not Ok
154	P4252-04DL	MYD9X3	SAM	10/19/24 00:19	NOT USE	Kareem	Not Ok
155	P4252-05DL	MYD9X3D	DUP	10/19/24 00:23	NOT USE	Kareem	Not Ok
156	P4252-04LDL	MYD9X3L	SD	10/19/24 00:27	NOT USE	Kareem	Not Ok
157	P4252-06DL	MYD9X3S	MS	10/19/24 00:32	NOT USE	Kareem	Not Ok
158	CCV019	CCV019	CCV	10/19/24 00:36		Kareem	OK
159	CCB019	CCB019	CCB	10/19/24 00:40	CCB fail for AI	Kareem	OK