

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133087

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:02:42 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:12 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE005	TUNE	10/23/24 03:34		Jaswal	OK
2	S00	S0	CAL1	10/23/24 04:10		Jaswal	OK
3	S02	S02	CAL3	10/23/24 04:13		Jaswal	OK
4	S03	S03	CAL4	10/23/24 04:17		Jaswal	OK
5	S04	S04	CAL5	10/23/24 04:20		Jaswal	OK
6	S05	S05	CAL6	10/23/24 04:23		Jaswal	OK
7	S06	S06	CAL7	10/23/24 04:26		Jaswal	OK
8	S07	S07	CAL8	10/23/24 04:30		Jaswal	OK
9	S08	S08	CAL9	10/23/24 04:32		Jaswal	OK
10	ICV005	ICV005	ICV	10/23/24 05:18	Inst. paused for calibration review	Jaswal	OK
11	ICB005	ICB005	ICB	10/23/24 05:24	Inst. paused for review	Jaswal	OK
12	ICSA005	ICSA005	ICSA	10/23/24 05:27		Jaswal	OK
13	ICSAB005	ICSAB005	ICSAB	10/23/24 05:30		Jaswal	OK
14	CCV020	CCV020	CCV	10/23/24 05:42	Inst. paused for review	Jaswal	OK
15	CCB020	CCB020	CCB	10/23/24 05:46		Jaswal	OK
16	PB163990BL	PBS990	MB	10/23/24 06:25		Jaswal	OK
17	PB163990BS	LCS990	LCS	10/23/24 06:29		Jaswal	OK
18	PB163990BSD	LCSD990	LCSD	10/23/24 06:32		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133087

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:02:42 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:12 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

19	PB163992BL	PBS992	MB	10/23/24 06:36		Jaswal	OK
20	PB163992BS	LCS992	LCS	10/23/24 06:39		Jaswal	OK
21	PB163992BSD	LCSD992	LCSD	10/23/24 06:43		Jaswal	OK
22	PB164027BL	PBS027	MB	10/23/24 06:46	Not Use	Jaswal	Not Ok
23	PB164027BS	LCS027	LCS	10/23/24 06:53	Not Use	Jaswal	Not Ok
24	PB164027BSD	LCSD027	LCSD	10/23/24 07:07	Not Use	Jaswal	Not Ok
25	PB164032BL	PBS032	MB	10/23/24 07:10	Not Use	Jaswal	Not Ok
26	PB164032BS	LCS032	LCS	10/23/24 07:14	Not Use	Jaswal	Not Ok
27	PB164032BSD	LCSD032	LCSD	10/23/24 07:17	Not Use	Jaswal	Not Ok
28	PB164083BL	PBS083	MB	10/23/24 07:21	Not Use	Jaswal	Not Ok
29	PB164083BS	LCS083	LCS	10/23/24 07:24	Not Use	Jaswal	Not Ok
30	PB164083BSD	LCSD083	LCSD	10/23/24 07:27	Not Use	Jaswal	Not Ok
31	P4291-01	MYD0B5	SAM	10/23/24 07:31		Jaswal	OK
32	CCV021	CCV021	CCV	10/23/24 07:34		Jaswal	OK
33	CCB021	CCB021	CCB	10/23/24 07:39		Jaswal	OK
34	P4291-02	MYD0B6	SAM	10/23/24 07:42		Jaswal	OK
35	P4291-03	MYD0B7	SAM	10/23/24 07:45		Jaswal	OK
36	P4291-04	MYD0B8	SAM	10/23/24 07:49	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
37	P4291-05	MYD0B8D	DUP	10/23/24 07:52	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
38	P4291-04L	MYD0B8L	SD	10/23/24 07:56	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133087

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:02:42 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:12 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

39	P4291-06	MYD0B8S	MS	10/23/24 07:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
40	P4291-04MS	MYD0B8SRE	MS	10/23/24 08:02	Ba,Co,Ni,V High	Jaswal	Dilution
41	P4291-08	MYD0C0	SAM	10/23/24 08:06		Jaswal	OK
42	P4291-09	MYD0C1	SAM	10/23/24 08:09		Jaswal	OK
43	P4291-10	MYD0C2	SAM	10/23/24 08:12		Jaswal	OK
44	P4291-11	MYD0C3	SAM	10/23/24 08:16		Jaswal	OK
45	P4291-12	MYD0C4	SAM	10/23/24 08:19		Jaswal	OK
46	P4291-13	MYD0C5	SAM	10/23/24 08:22		Jaswal	OK
47	P4291-14	MYD0C6	SAM	10/23/24 08:26		Jaswal	OK
48	P4291-15	MYD0C7	SAM	10/23/24 08:29		Jaswal	OK
49	P4291-16	MYD0C8	SAM	10/23/24 08:33		Jaswal	OK
50	P4291-17	MYD0C9	SAM	10/23/24 08:36		Jaswal	OK
51	P4291-18	MYD0D0	SAM	10/23/24 08:39		Jaswal	OK
52	P4291-19	MYD0D1	SAM	10/23/24 08:43		Jaswal	OK
53	P4291-20	MYD0D2	SAM	10/23/24 08:46		Jaswal	OK
54	P4291-21	MYD0D3	SAM	10/23/24 08:50		Jaswal	OK
55	P4291-22	MYD0D4	SAM	10/23/24 08:53		Jaswal	OK
56	P4292-01	MYD0D5	SAM	10/23/24 08:56		Jaswal	OK
57	P4292-02	MYD0S9	SAM	10/23/24 09:00		Jaswal	OK
58	P4292-03	MYD0T0	SAM	10/23/24 09:03		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133087

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:02:42 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:12 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

59	P4292-04	MYD0T1	SAM	10/23/24 09:07		Jaswal	OK
60	CCV022	CCV022	CCV	10/23/24 09:10		Jaswal	OK
61	CCB022	CCB022	CCB	10/23/24 09:13		Jaswal	OK