

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	10/23/24 15:36		Jaswal	OK
2	S00	S0	CAL1	10/23/24 16:08		Jaswal	OK
3	S02	S02	CAL3	10/23/24 16:11		Jaswal	OK
4	S03	S03	CAL4	10/23/24 16:15		Jaswal	OK
5	S04	S04	CAL5	10/23/24 16:18		Jaswal	OK
6	S05	S05	CAL6	10/23/24 16:21		Jaswal	OK
7	S06	S06	CAL7	10/23/24 16:23		Jaswal	OK
8	S07	S07	CAL8	10/23/24 16:26		Jaswal	OK
9	S08	S08	CAL9	10/23/24 16:29		Jaswal	OK
10	ICV001	ICV001	ICV	10/23/24 16:54	Inst. paused for Calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	10/23/24 17:12	Inst. paused for review	Jaswal	OK
12	ICSA001	ICSA001	ICSA	10/23/24 17:15		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	10/23/24 17:18		Jaswal	OK
14	CCV001	CCV001	CCV	10/23/24 17:21		Jaswal	OK
15	CCB001	CCB001	CCB	10/23/24 17:24		Jaswal	OK
16	PB164027BL	PBS027	MB	10/23/24 17:27		Jaswal	OK
17	PB164027BS	LCS027	LCS	10/23/24 17:31		Jaswal	OK
18	PB164027BSD	LCSD027	LCSD	10/23/24 17:34		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

19	PB164032BL	PBS032	MB	10/23/24 17:37		Jaswal	OK
20	PB164032BS	LCS032	LCS	10/23/24 17:41		Jaswal	OK
21	PB164032BSD	LCSD032	LCSD	10/23/24 17:44		Jaswal	OK
22	P4292-05	MYD0T2	SAM	10/23/24 17:51		Jaswal	OK
23	P4292-06	MYD0T3	SAM	10/23/24 17:54		Jaswal	OK
24	P4292-07	MYD0T4	SAM	10/23/24 17:57		Jaswal	OK
25	P4292-08	MYD0T5	SAM	10/23/24 18:00		Jaswal	OK
26	P4292-09	MYD0T6	SAM	10/23/24 18:03		Jaswal	OK
27	P4292-10	MYD0T7	SAM	10/23/24 18:06		Jaswal	OK
28	P4292-11	MYD0T8	SAM	10/23/24 18:09		Jaswal	OK
29	P4292-12	MYD0T9	SAM	10/23/24 18:12		Jaswal	OK
30	P4292-13	MYD0W0	SAM	10/23/24 18:15		Jaswal	OK
31	P4292-14	MYD0W1	SAM	10/23/24 18:18		Jaswal	OK
32	P4292-15	MYD0W2	SAM	10/23/24 18:21	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
33	P4292-16	MYD0W2D	DUP	10/23/24 18:24	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
34	P4292-15L	MYD0W2L	SD	10/23/24 18:28	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
35	P4292-17	MYD0W2S	MS	10/23/24 18:31	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
36	P4292-15MS	MYD0W2SRE	MS	10/23/24 18:34	Ba,Co,Ni,V High	Jaswal	Dilution
37	P4292-18	MYD0W3	SAM	10/23/24 18:37		Jaswal	OK
38	P4292-19	MYD0W4	SAM	10/23/24 18:40		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

39	P4292-20	MYD0W5	SAM	10/23/24 18:43		Jaswal	OK
40	P4292-21	MYD0W6	SAM	10/23/24 18:46		Jaswal	OK
41	CCV002	CCV002	CCV	10/23/24 18:49		Jaswal	OK
42	CCB002	CCB002	CCB	10/23/24 18:52		Jaswal	OK
43	P4292-22	MYD0W7	SAM	10/23/24 18:55		Jaswal	OK
44	P4291-04DL	MYD0B8	SAM	10/23/24 18:58	4X For Ba,Co,Ni,V	Jaswal	Confirms
45	P4291-05DL	MYD0B8D	DUP	10/23/24 19:01	4X For Ba,Co,Ni,V	Jaswal	Confirms
46	P4291-04LDL	MYD0B8L	SD	10/23/24 19:05	20X For Ba,Co,Ni,V	Jaswal	Confirms
47	P4291-06DL	MYD0B8S	MS	10/23/24 19:08	4X For Ba,Co,Ni,V	Jaswal	Confirms
48	P4291-04MSDL	MYD0B8SRE	MS	10/23/24 19:11	4X For Ba,Co,Ni,V	Jaswal	Confirms
49	P4292-15DL	MYD0W2	SAM	10/23/24 19:25	4X For Ba,Co,Ni,V	Jaswal	Confirms
50	P4292-16DL	MYD0W2D	DUP	10/23/24 19:28	4X For Ba,Co,Ni,V	Jaswal	Confirms
51	P4292-15LDL	MYD0W2L	SD	10/23/24 19:31	20X For Ba,Co,Ni,V	Jaswal	Confirms
52	P4292-17DL	MYD0W2S	MS	10/23/24 19:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
53	P4292-15MSDL	MYD0W2SRE	MS	10/23/24 19:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
54	P4293-01	MYD9Z9	SAM	10/23/24 19:41		Jaswal	OK
55	P4293-02	MYDA00	SAM	10/23/24 19:44		Jaswal	OK
56	P4293-03	MYDA01	SAM	10/23/24 19:47		Jaswal	OK
57	P4293-04	MYDA02	SAM	10/23/24 19:50		Jaswal	OK
58	P4293-05	MYDA03	SAM	10/23/24 19:53		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

59	P4293-06	MYDA04	SAM	10/23/24 19:56		Jaswal	OK
60	P4293-07	MYDA05	SAM	10/23/24 19:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
61	P4293-08	MYDA05D	DUP	10/23/24 20:02	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
62	P4293-07L	MYDA05L	SD	10/23/24 20:05	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
63	P4293-09	MYDA05S	MS	10/23/24 20:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
64	P4293-07MS	MYDA05SRE	MS	10/23/24 20:12	Ba,Co,Ni,V High	Jaswal	Dilution
65	CCV003	CCV003	CCV	10/23/24 20:29		Jaswal	OK
66	CCB003	CCB003	CCB	10/23/24 20:32		Jaswal	OK
67	P4293-10	MYDA06	SAM	10/23/24 20:36		Jaswal	OK
68	P4293-11	MYDA07	SAM	10/23/24 20:39		Jaswal	OK
69	P4293-12	MYDA08	SAM	10/23/24 20:42		Jaswal	OK
70	P4293-13	MYDA09	SAM	10/23/24 20:45		Jaswal	OK
71	P4293-14	MYDA10	SAM	10/23/24 20:48		Jaswal	OK
72	P4293-15	MYDA11	SAM	10/23/24 20:51		Jaswal	OK
73	P4293-16	MYDA12	SAM	10/23/24 20:54		Jaswal	OK
74	P4293-17	MYDA13	SAM	10/23/24 20:57		Jaswal	OK
75	P4293-18	MYDA14	SAM	10/23/24 21:00		Jaswal	OK
76	P4293-19	MYDA15	SAM	10/23/24 21:03		Jaswal	OK
77	P4293-20	MYDA16	SAM	10/23/24 21:07		Jaswal	OK
78	P4293-21	MYDA17	SAM	10/23/24 21:10		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

79	P4293-22	MYDA18	SAM	10/23/24 21:13		Jaswal	OK
80	P4294-01	MYDA19	SAM	10/23/24 21:16		Jaswal	OK
81	P4294-02	MYDA20	SAM	10/23/24 21:19		Jaswal	OK
82	P4294-03	MYDA21	SAM	10/23/24 21:22		Jaswal	OK
83	P4294-04	MYDA22	SAM	10/23/24 21:25		Jaswal	OK
84	P4294-05	MYDA23	SAM	10/23/24 21:28		Jaswal	OK
85	ICSA002	ICSA002	ICSA	10/23/24 21:31		Jaswal	OK
86	ICSAB002	ICSAB002	ICSAB	10/23/24 21:35		Jaswal	OK
87	CCV004	CCV004	CCV	10/23/24 21:38		Jaswal	OK
88	CCB004	CCB004	CCB	10/23/24 21:40		Jaswal	OK
89	P4294-06	MYDA24	SAM	10/23/24 21:44		Jaswal	OK
90	P4294-07	MYDA25	SAM	10/23/24 21:47		Jaswal	OK
91	P4294-08	MYDA26	SAM	10/23/24 21:50		Jaswal	OK
92	P4294-09	MYDA27	SAM	10/23/24 21:53		Jaswal	OK
93	P4294-10	MYDA28	SAM	10/23/24 21:56		Jaswal	OK
94	P4294-11	MYDA29	SAM	10/23/24 21:59		Jaswal	OK
95	P4294-12	MYDA30	SAM	10/23/24 22:02		Jaswal	OK
96	P4291-07	MYD0B9	SAM	10/23/24 22:07		Jaswal	OK
97	CCV005	CCV005	CCV	10/23/24 22:10		Jaswal	OK
98	CCB005	CCB005	CCB	10/23/24 22:13		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

99	P4294-13	MYDA31	SAM	10/23/24 22:16		Jaswal	OK
100	P4294-14	MYDA32	SAM	10/23/24 22:19		Jaswal	OK
101	P4294-15	MYDA33	SAM	10/23/24 22:22		Jaswal	OK
102	P4294-16	MYDA34	SAM	10/23/24 22:25		Jaswal	OK
103	P4294-17	MYDA35	SAM	10/23/24 22:29		Jaswal	OK
104	P4294-18	MYDA36	SAM	10/23/24 22:32		Jaswal	OK
105	P4294-19	MYDA37	SAM	10/23/24 22:35		Jaswal	OK
106	P4294-20	MYDA42	SAM	10/23/24 22:38	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
107	P4294-21	MYDA42D	DUP	10/23/24 22:41	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
108	P4294-20L	MYDA42L	SD	10/23/24 22:44	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
109	P4294-22	MYDA42S	MS	10/23/24 22:47	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
110	P4294-20MS	MYDA42SRE	MS	10/23/24 22:51	Ba,Co,Ni,V High	Jaswal	Dilution
111	P4295-01	MYD0Z2	SAM	10/23/24 22:53		Jaswal	OK
112	P4295-02	MYD0Z3	SAM	10/23/24 22:57		Jaswal	OK
113	P4295-03	MYD0Z4	SAM	10/23/24 23:00		Jaswal	OK
114	P4295-04	MYD0Z5	SAM	10/23/24 23:03		Jaswal	OK
115	P4295-05	MYD0Z6	SAM	10/23/24 23:06		Jaswal	OK
116	P4295-06	MYD0Z7	SAM	10/23/24 23:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
117	P4295-07	MYD0Z7D	DUP	10/23/24 23:12	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
118	P4295-06L	MYD0Z7L	SD	10/23/24 23:15	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

119	P4295-08	MYD0Z7S	MS	10/23/24 23:18	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
120	P4295-06MS	MYD0Z7SRE	MS	10/23/24 23:21	MS Fail For Pb, Ba,Co,Ni,V High	Jaswal	Dilution
121	P4295-09	MYD0Z8	SAM	10/23/24 23:24		Jaswal	OK
122	P4295-10	MYD0Z9	SAM	10/23/24 23:27		Jaswal	OK
123	P4295-11	MYD100	SAM	10/23/24 23:30		Jaswal	OK
124	P4295-12	MYD101	SAM	10/23/24 23:33		Jaswal	OK
125	CCV006	CCV006	CCV	10/23/24 23:36		Jaswal	OK
126	CCB006	CCB006	CCB	10/23/24 23:39		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE003	TUNE	10/24/24 11:46		Jaswal	OK
2	S00	S0	CAL1	10/24/24 12:20		Jaswal	OK
3	S02	S02	CAL3	10/24/24 12:24		Jaswal	OK
4	S03	S03	CAL4	10/24/24 12:27		Jaswal	OK
5	S04	S04	CAL5	10/24/24 12:30		Jaswal	OK
6	S05	S05	CAL6	10/24/24 12:33		Jaswal	OK
7	S06	S06	CAL7	10/24/24 12:36		Jaswal	OK
8	S07	S07	CAL8	10/24/24 12:38		Jaswal	OK
9	S08	S08	CAL9	10/24/24 12:41		Jaswal	OK
10	ICV003	ICV003	ICV	10/24/24 12:56		Jaswal	OK
11	ICB003	ICB003	ICB	10/24/24 12:59		Jaswal	OK
12	ICSA003	ICSA003	ICSA	10/24/24 13:03		Jaswal	OK
13	ICSAB003	ICSAB003	ICSAB	10/24/24 13:06		Jaswal	OK
14	CCV007	CCV007	CCV	10/24/24 13:12		Jaswal	OK
15	CCB007	CCB007	CCB	10/24/24 13:15		Jaswal	OK
16	P4293-07DL	MYDA05	SAM	10/24/24 13:18	4X For Ba,Co,Ni,V	Jaswal	Confirms
17	P4293-08DL	MYDA05D	DUP	10/24/24 13:22	4X For Ba,Co,Ni,V	Jaswal	Confirms
18	P4293-07LDL	MYDA05L	SD	10/24/24 13:25	20X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

19	P4293-09DL	MYDA05S	MS	10/24/24 13:28	4X For Ba,Co,Ni,V	Jaswal	Confirms
20	P4293-07MSDL	MYDA05SRE	MS	10/24/24 13:31	4X For Ba,Co,Ni,V	Jaswal	Confirms
21	P4294-20DL	MYDA42	SAM	10/24/24 13:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
22	P4294-21DL	MYDA42D	DUP	10/24/24 13:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
23	P4294-20LDL	MYDA42L	SD	10/24/24 13:41	20X For Ba,Co,Ni,V	Jaswal	Confirms
24	P4294-22DL	MYDA42S	MS	10/24/24 13:44	4X For Ba,Co,Ni,V	Jaswal	Confirms
25	P4294-20MSDL	MYDA42SRE	MS	10/24/24 13:47	4X For Ba,Co,Ni,V	Jaswal	Confirms
26	P4295-06DL	MYD0Z7	SAM	10/24/24 13:50	4X For Ba,Co,Ni,V	Jaswal	Confirms
27	P4295-07DL	MYD0Z7D	DUP	10/24/24 13:53	4X For Ba,Co,Ni,V	Jaswal	Confirms
28	P4295-06LDL	MYD0Z7L	SD	10/24/24 13:56	20X For Ba,Co,Ni,V	Jaswal	Confirms
29	P4295-08DL	MYD0Z7S	MS	10/24/24 14:00	4X For Ba,Co,Ni,V	Jaswal	Confirms
30	P4295-06MSDL	MYD0Z7SRE	MS	10/24/24 14:03	4X For Ba,Co,Ni,V	Jaswal	Confirms
31	P4295-06A	MYD0Z7A	PS	10/24/24 14:06	PS For Pb	Jaswal	OK
32	CCV008	CCV008	CCV	10/24/24 14:09		Jaswal	OK
33	CCB008	CCB008	CCB	10/24/24 14:12		Jaswal	OK
34	P4295-13	MYD102	SAM	10/24/24 14:15		Jaswal	OK
35	P4295-14	MYD103	SAM	10/24/24 14:18		Jaswal	OK
36	P4295-15	MYD104	SAM	10/24/24 14:21		Jaswal	OK
37	P4295-16	MYD105	SAM	10/24/24 14:24		Jaswal	OK
38	P4295-17	MYDA38	SAM	10/24/24 14:27		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	
Chk Standard	MP82587,MP82586

39	P4295-18	MYDA39	SAM	10/24/24 14:31		Jaswal	OK
40	P4295-19	MYDA40	SAM	10/24/24 14:34		Jaswal	OK
41	P4295-20	MYDA41	SAM	10/24/24 14:37		Jaswal	OK
42	P4295-21	MYDBB3	SAM	10/24/24 14:40		Jaswal	OK
43	PB164085BL	PBS085	MB	10/24/24 14:43		Jaswal	OK
44	PB164085BS	LCS085	LCS	10/24/24 14:46		Jaswal	OK
45	PB164085BSD	LCSD085	LCSD	10/24/24 14:50		Jaswal	OK
46	PB164090BL	PBS090	MB	10/24/24 14:53		Jaswal	OK
47	PB164090BS	LCS090	LCS	10/24/24 14:56		Jaswal	OK
48	PB164090BSD	LCSD090	LCSD	10/24/24 14:59		Jaswal	OK
49	P4296-01	MYD0W8	SAM	10/24/24 15:03		Jaswal	OK
50	P4296-02	MYD0W9	SAM	10/24/24 15:06		Jaswal	OK
51	P4296-03	MYD0X0	SAM	10/24/24 15:09		Jaswal	OK
52	P4296-04	MYD0Y2	SAM	10/24/24 15:12		Jaswal	OK
53	P4296-05	MYD0Y3	SAM	10/24/24 15:15	Ni High	Jaswal	Dilution
54	P4296-06	MYD0Y4	SAM	10/24/24 15:18		Jaswal	OK
55	P4296-07	MYD0Y5	SAM	10/24/24 15:22	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	
Chk Standard	MP82587,MP82586

56	P4296-08	MYD0Y5D	DUP	10/24/24 15:25	Ba,Co,Ni,V High in MSRE(ICB),Bismuth(Dilution),Holmium(Dilution),Lithium(Dilution),Rhodium(Dilution),Scandium(Dilution),Terbium(Dilution),Yttrium(Dilution),	Jaswal	Dilution
57	P4296-07L	MYD0Y5L	SD	10/24/24 15:28	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
58	P4296-09	MYD0Y5S	MS	10/24/24 15:31	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
59	P4296-07MS	MYD0Y5SRE	MS	10/24/24 15:34	MS Fail For As(220.75ppb),Cu(3015.15ppb),Zn(187.38ppb),Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
60	CCV009	CCV009	CCV	10/24/24 15:41		Jaswal	OK
61	CCB009	CCB009	CCB	10/24/24 15:44		Jaswal	OK
62	P4296-10	MYD0Y9	SAM	10/24/24 15:47		Jaswal	OK
63	P4296-11	MYD0Z0	SAM	10/24/24 15:50		Jaswal	OK
64	P4296-12	MYD0Z1	SAM	10/24/24 15:53		Jaswal	OK
65	P4296-13	MYDA43	SAM	10/24/24 15:56		Jaswal	OK
66	P4296-14	MYDA44	SAM	10/24/24 15:59		Jaswal	OK
67	P4296-15	MYDA45	SAM	10/24/24 16:02		Jaswal	OK
68	P4296-16	MYDA46	SAM	10/24/24 16:06		Jaswal	OK
69	P4296-17	MYDA47	SAM	10/24/24 16:09		Jaswal	OK
70	P4296-18	MYDA48	SAM	10/24/24 16:12		Jaswal	OK
71	P4296-19	MYDA49	SAM	10/24/24 16:15		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

72	P4296-20	MYDA50	SAM	10/24/24 16:18		Jaswal	OK
73	P4296-21	MYDA51	SAM	10/24/24 16:21		Jaswal	OK
74	P4296-22	MYDA52	SAM	10/24/24 16:24		Jaswal	OK
75	PB164128BL	PBS128	MB	10/24/24 16:27		Jaswal	OK
76	PB164128BS	LCS128	LCS	10/24/24 16:31		Jaswal	OK
77	PB164128BSD	LCSD128	LCSD	10/24/24 16:34		Jaswal	OK
78	P4297-01	MYDA53	SAM	10/24/24 16:37		Jaswal	OK
79	P4297-02	MYDA54	SAM	10/24/24 16:40		Jaswal	OK
80	P4297-03	MYDA55	SAM	10/24/24 16:43		Jaswal	OK
81	P4297-04	MYDA56	SAM	10/24/24 16:47		Jaswal	OK
82	P4297-05	MYDA57	SAM	10/24/24 16:50		Jaswal	OK
83	P4297-06	MYDA58	SAM	10/24/24 16:53		Jaswal	OK
84	P4297-07	MYDA59	SAM	10/24/24 16:56		Jaswal	OK
85	P4297-08	MYDA60	SAM	10/24/24 16:59		Jaswal	OK
86	P4297-09	MYDA61	SAM	10/24/24 17:02		Jaswal	OK
87	P4297-10	MYDA62	SAM	10/24/24 17:05		Jaswal	OK
88	CCV010	CCV010	CCV	10/24/24 17:14		Jaswal	OK
89	CCB010	CCB010	CCB	10/24/24 17:17		Jaswal	OK
90	P4297-11	MYDA63	SAM	10/24/24 17:20		Jaswal	OK
91	P4297-12	MYDA64	SAM	10/24/24 17:24		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

92	P4297-13	MYDA65	SAM	10/24/24 17:27		Jaswal	OK
93	P4297-14	MYDA66	SAM	10/24/24 17:30	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
94	P4297-15	MYDA66D	DUP	10/24/24 17:33	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
95	P4297-14L	MYDA66L	SD	10/24/24 17:36	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
96	P4297-16	MYDA66S	MS	10/24/24 17:39	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
97	P4297-14MS	MYDA66SRE	MS	10/24/24 17:42	MS Fail For Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
98	P4297-17	MYDA67	SAM	10/24/24 17:45		Jaswal	OK
99	P4297-18	MYDA68	SAM	10/24/24 17:48		Jaswal	OK
100	P4297-19	MYDA69	SAM	10/24/24 17:52		Jaswal	OK
101	P4297-20	MYDA70	SAM	10/24/24 17:55		Jaswal	OK
102	P4297-21	MYDA71	SAM	10/24/24 17:58		Jaswal	OK
103	P4297-22	MYDA72	SAM	10/24/24 18:01		Jaswal	OK
104	P4296-07DL	MYD0Y5	SAM	10/24/24 18:04	4X For Ba,Co,Ni,V	Jaswal	Confirms
105	P4296-08DL	MYD0Y5D	DUP	10/24/24 18:07	4X For Ba,Co,Ni,V	Jaswal	Confirms
106	P4296-07LDL	MYD0Y5L	SD	10/24/24 18:10	20X For Ba,Co,Ni,V	Jaswal	Confirms
107	P4296-09DL	MYD0Y5S	MS	10/24/24 18:14	4X For Ba,Co,Ni,V	Jaswal	Confirms
108	P4296-07MSDL	MYD0Y5SRE	MS	10/24/24 18:17	4X For Ba,Co,Ni,V	Jaswal	Confirms
109	P4297-14DL	MYDA66	SAM	10/24/24 18:20	4X For Ba,Co,Ni,V	Jaswal	Confirms
110	P4297-15DL	MYDA66D	DUP	10/24/24 18:23	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

111	P4297-14LDL	MYDA66L	SD	10/24/24 18:26	20X For Ba,Co,Ni,V	Jaswal	Confirms
112	P4297-16DL	MYDA66S	MS	10/24/24 18:30	4X For Ba,Co,Ni,V	Jaswal	Confirms
113	CCV011	CCV011	CCV	10/24/24 18:49		Jaswal	OK
114	CCB011	CCB011	CCB	10/24/24 18:52		Jaswal	OK
115	P4297-14MSDL	MYDA66SRE	MS	10/24/24 18:55	4X For Ba,Co,Ni,V	Jaswal	Confirms
116	PB164126BL	PBS126	MB	10/24/24 19:02		Jaswal	OK
117	PB164126BS	LCS126	LCS	10/24/24 19:06		Jaswal	OK
118	PB164126BSD	LCSD126	LCSD	10/24/24 19:09		Jaswal	OK
119	P4298-01	MYDA73	SAM	10/24/24 19:12		Jaswal	OK
120	P4298-02	MYDA74	SAM	10/24/24 19:16		Jaswal	OK
121	P4298-03	MYDA75	SAM	10/24/24 19:19		Jaswal	OK
122	P4298-04	MYDA76	SAM	10/24/24 19:22		Jaswal	OK
123	P4298-05	MYDA77	SAM	10/24/24 19:25		Jaswal	OK
124	P4298-06	MYDA78	SAM	10/24/24 19:28		Jaswal	OK
125	P4298-07	MYDA79	SAM	10/24/24 19:31		Jaswal	OK
126	P4298-08	MYDA80	SAM	10/24/24 19:34		Jaswal	OK
127	P4298-09	MYDA81	SAM	10/24/24 19:37		Jaswal	OK
128	P4298-10	MYDA82	SAM	10/24/24 19:40		Jaswal	OK
129	P4298-11	MYDA83	SAM	10/24/24 19:43		Jaswal	OK
130	P4298-12	MYDA84	SAM	10/24/24 19:47		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

131	P4298-13	MYDA85	SAM	10/24/24 19:50		Jaswal	OK
132	P4298-14	MYDA86	SAM	10/24/24 19:53		Jaswal	OK
133	P4298-15	MYDA87	SAM	10/24/24 19:56		Jaswal	OK
134	P4298-16	MYDA88	SAM	10/24/24 19:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
135	P4298-17	MYDA88D	DUP	10/24/24 20:02	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
136	P4298-16L	MYDA88L	SD	10/24/24 20:05	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
137	P4298-18	MYDA88S	MS	10/24/24 20:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
138	P4298-16MS	MYDA88SRE	MS	10/24/24 20:12	MS Fail For As(356.29ppb),Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
139	P4298-19	MYDA89	SAM	10/24/24 20:15		Jaswal	OK
140	CCV012	CCV012	CCV	10/24/24 20:18		Jaswal	OK
141	CCB012	CCB012	CCB	10/24/24 20:20		Jaswal	OK
142	P4298-20	MYDA90	SAM	10/24/24 20:24		Jaswal	OK
143	P4298-21	MYDA91	SAM	10/24/24 20:27		Jaswal	OK
144	P4298-22	MYDA92	SAM	10/24/24 20:30		Jaswal	OK
145	P4299-01	MYDA93	SAM	10/24/24 20:33		Jaswal	OK
146	P4299-02	MYDA94	SAM	10/24/24 20:36		Jaswal	OK
147	P4299-03	MYDA95	SAM	10/24/24 20:39		Jaswal	OK
148	P4299-04	MYDA96	SAM	10/24/24 20:42		Jaswal	OK
149	P4299-05	MYDA97	SAM	10/24/24 20:45		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

150	P4299-06	MYDA98	SAM	10/24/24 20:48		Jaswal	OK
151	P4299-07	MYDA99	SAM	10/24/24 20:52		Jaswal	OK
152	P4299-08	MYDAA0	SAM	10/24/24 20:55		Jaswal	OK
153	P4299-09	MYDAA1	SAM	10/24/24 20:58		Jaswal	OK
154	P4299-10	MYDAA2	SAM	10/24/24 21:01		Jaswal	OK
155	P4299-11	MYDAA3	SAM	10/24/24 21:04		Jaswal	OK
156	P4299-12	MYDAA4	SAM	10/24/24 21:07		Jaswal	OK
157	P4299-13	MYDAA5	SAM	10/24/24 21:10		Jaswal	OK
158	P4299-14	MYDAA6	SAM	10/24/24 21:13		Jaswal	OK
159	P4299-15	MYDAA7	SAM	10/24/24 21:16		Jaswal	OK
160	P4299-16	MYDAA8	SAM	10/24/24 21:19		Jaswal	OK
161	P4299-17	MYDAA9	SAM	10/24/24 21:22		Jaswal	OK
162	P4299-18	MYDAB0	SAM	10/24/24 21:26		Jaswal	OK
163	P4299-19	MYDAB1	SAM	10/24/24 21:29	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
164	P4299-20	MYDAB1D	DUP	10/24/24 21:32	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
165	P4299-19L	MYDAB1L	SD	10/24/24 21:35	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
166	P4299-21	MYDAB1S	MS	10/24/24 21:38	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
167	P4299-19MS	MYDAB1SRE	MS	10/24/24 21:41	MS Fail For As(146.42ppb),Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
168	P4299-22	MYDAB2	SAM	10/24/24 21:44		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QCBatch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	
Chk Standard	MP82587,MP82586

169	CCV013	CCV013	CCV	10/24/24 21:47		Jaswal	OK
170	CCB013	CCB013	CCB	10/24/24 21:51		Jaswal	OK