

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	10/23/24 15:36		Jaswal	OK
2	S00	S0	CAL1	10/23/24 16:08		Jaswal	OK
3	S02	S02	CAL3	10/23/24 16:11		Jaswal	OK
4	S03	S03	CAL4	10/23/24 16:15		Jaswal	OK
5	S04	S04	CAL5	10/23/24 16:18		Jaswal	OK
6	S05	S05	CAL6	10/23/24 16:21		Jaswal	OK
7	S06	S06	CAL7	10/23/24 16:23		Jaswal	OK
8	S07	S07	CAL8	10/23/24 16:26		Jaswal	OK
9	S08	S08	CAL9	10/23/24 16:29		Jaswal	OK
10	ICV001	ICV001	ICV	10/23/24 16:54	Inst. paused for Calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	10/23/24 17:12	Inst. paused for review	Jaswal	OK
12	ICSA001	ICSA001	ICSA	10/23/24 17:15		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	10/23/24 17:18		Jaswal	OK
14	CCV001	CCV001	CCV	10/23/24 17:21		Jaswal	OK
15	CCB001	CCB001	CCB	10/23/24 17:24		Jaswal	OK
16	PB164027BL	PBS027	MB	10/23/24 17:27		Jaswal	OK
17	PB164027BS	LCS027	LCS	10/23/24 17:31		Jaswal	OK
18	PB164027BSD	LCSD027	LCSD	10/23/24 17:34		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

19	PB164032BL	PBS032	MB	10/23/24 17:37		Jaswal	OK
20	PB164032BS	LCS032	LCS	10/23/24 17:41		Jaswal	OK
21	PB164032BSD	LCSD032	LCSD	10/23/24 17:44		Jaswal	OK
22	P4292-05	MYD0T2	SAM	10/23/24 17:51		Jaswal	OK
23	P4292-06	MYD0T3	SAM	10/23/24 17:54		Jaswal	OK
24	P4292-07	MYD0T4	SAM	10/23/24 17:57		Jaswal	OK
25	P4292-08	MYD0T5	SAM	10/23/24 18:00		Jaswal	OK
26	P4292-09	MYD0T6	SAM	10/23/24 18:03		Jaswal	OK
27	P4292-10	MYD0T7	SAM	10/23/24 18:06		Jaswal	OK
28	P4292-11	MYD0T8	SAM	10/23/24 18:09		Jaswal	OK
29	P4292-12	MYD0T9	SAM	10/23/24 18:12		Jaswal	OK
30	P4292-13	MYD0W0	SAM	10/23/24 18:15		Jaswal	OK
31	P4292-14	MYD0W1	SAM	10/23/24 18:18		Jaswal	OK
32	P4292-15	MYD0W2	SAM	10/23/24 18:21	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
33	P4292-16	MYD0W2D	DUP	10/23/24 18:24	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
34	P4292-15L	MYD0W2L	SD	10/23/24 18:28	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
35	P4292-17	MYD0W2S	MS	10/23/24 18:31	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
36	P4292-15MS	MYD0W2SRE	MS	10/23/24 18:34	Ba,Co,Ni,V High	Jaswal	Dilution
37	P4292-18	MYD0W3	SAM	10/23/24 18:37		Jaswal	OK
38	P4292-19	MYD0W4	SAM	10/23/24 18:40		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

39	P4292-20	MYD0W5	SAM	10/23/24 18:43		Jaswal	OK
40	P4292-21	MYD0W6	SAM	10/23/24 18:46		Jaswal	OK
41	CCV002	CCV002	CCV	10/23/24 18:49		Jaswal	OK
42	CCB002	CCB002	CCB	10/23/24 18:52		Jaswal	OK
43	P4292-22	MYD0W7	SAM	10/23/24 18:55		Jaswal	OK
44	P4291-04DL	MYD0B8	SAM	10/23/24 18:58	4X For Ba,Co,Ni,V	Jaswal	Confirms
45	P4291-05DL	MYD0B8D	DUP	10/23/24 19:01	4X For Ba,Co,Ni,V	Jaswal	Confirms
46	P4291-04LDL	MYD0B8L	SD	10/23/24 19:05	20X For Ba,Co,Ni,V	Jaswal	Confirms
47	P4291-06DL	MYD0B8S	MS	10/23/24 19:08	4X For Ba,Co,Ni,V	Jaswal	Confirms
48	P4291-04MSDL	MYD0B8SRE	MS	10/23/24 19:11	4X For Ba,Co,Ni,V	Jaswal	Confirms
49	P4292-15DL	MYD0W2	SAM	10/23/24 19:25	4X For Ba,Co,Ni,V	Jaswal	Confirms
50	P4292-16DL	MYD0W2D	DUP	10/23/24 19:28	4X For Ba,Co,Ni,V	Jaswal	Confirms
51	P4292-15LDL	MYD0W2L	SD	10/23/24 19:31	20X For Ba,Co,Ni,V	Jaswal	Confirms
52	P4292-17DL	MYD0W2S	MS	10/23/24 19:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
53	P4292-15MSDL	MYD0W2SRE	MS	10/23/24 19:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
54	P4293-01	MYD9Z9	SAM	10/23/24 19:41		Jaswal	OK
55	P4293-02	MYDA00	SAM	10/23/24 19:44		Jaswal	OK
56	P4293-03	MYDA01	SAM	10/23/24 19:47		Jaswal	OK
57	P4293-04	MYDA02	SAM	10/23/24 19:50		Jaswal	OK
58	P4293-05	MYDA03	SAM	10/23/24 19:53		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

59	P4293-06	MYDA04	SAM	10/23/24 19:56		Jaswal	OK
60	P4293-07	MYDA05	SAM	10/23/24 19:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
61	P4293-08	MYDA05D	DUP	10/23/24 20:02	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
62	P4293-07L	MYDA05L	SD	10/23/24 20:05	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
63	P4293-09	MYDA05S	MS	10/23/24 20:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
64	P4293-07MS	MYDA05SRE	MS	10/23/24 20:12	Ba,Co,Ni,V High	Jaswal	Dilution
65	CCV003	CCV003	CCV	10/23/24 20:29		Jaswal	OK
66	CCB003	CCB003	CCB	10/23/24 20:32		Jaswal	OK
67	P4293-10	MYDA06	SAM	10/23/24 20:36		Jaswal	OK
68	P4293-11	MYDA07	SAM	10/23/24 20:39		Jaswal	OK
69	P4293-12	MYDA08	SAM	10/23/24 20:42		Jaswal	OK
70	P4293-13	MYDA09	SAM	10/23/24 20:45		Jaswal	OK
71	P4293-14	MYDA10	SAM	10/23/24 20:48		Jaswal	OK
72	P4293-15	MYDA11	SAM	10/23/24 20:51		Jaswal	OK
73	P4293-16	MYDA12	SAM	10/23/24 20:54		Jaswal	OK
74	P4293-17	MYDA13	SAM	10/23/24 20:57		Jaswal	OK
75	P4293-18	MYDA14	SAM	10/23/24 21:00		Jaswal	OK
76	P4293-19	MYDA15	SAM	10/23/24 21:03		Jaswal	OK
77	P4293-20	MYDA16	SAM	10/23/24 21:07		Jaswal	OK
78	P4293-21	MYDA17	SAM	10/23/24 21:10		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

79	P4293-22	MYDA18	SAM	10/23/24 21:13		Jaswal	OK
80	P4294-01	MYDA19	SAM	10/23/24 21:16		Jaswal	OK
81	P4294-02	MYDA20	SAM	10/23/24 21:19		Jaswal	OK
82	P4294-03	MYDA21	SAM	10/23/24 21:22		Jaswal	OK
83	P4294-04	MYDA22	SAM	10/23/24 21:25		Jaswal	OK
84	P4294-05	MYDA23	SAM	10/23/24 21:28		Jaswal	OK
85	ICSA002	ICSA002	ICSA	10/23/24 21:31		Jaswal	OK
86	ICSAB002	ICSAB002	ICSAB	10/23/24 21:35		Jaswal	OK
87	CCV004	CCV004	CCV	10/23/24 21:38		Jaswal	OK
88	CCB004	CCB004	CCB	10/23/24 21:40		Jaswal	OK
89	P4294-06	MYDA24	SAM	10/23/24 21:44		Jaswal	OK
90	P4294-07	MYDA25	SAM	10/23/24 21:47		Jaswal	OK
91	P4294-08	MYDA26	SAM	10/23/24 21:50		Jaswal	OK
92	P4294-09	MYDA27	SAM	10/23/24 21:53		Jaswal	OK
93	P4294-10	MYDA28	SAM	10/23/24 21:56		Jaswal	OK
94	P4294-11	MYDA29	SAM	10/23/24 21:59		Jaswal	OK
95	P4294-12	MYDA30	SAM	10/23/24 22:02		Jaswal	OK
96	P4291-07	MYD0B9	SAM	10/23/24 22:07		Jaswal	OK
97	CCV005	CCV005	CCV	10/23/24 22:10		Jaswal	OK
98	CCB005	CCB005	CCB	10/23/24 22:13		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

99	P4294-13	MYDA31	SAM	10/23/24 22:16		Jaswal	OK
100	P4294-14	MYDA32	SAM	10/23/24 22:19		Jaswal	OK
101	P4294-15	MYDA33	SAM	10/23/24 22:22		Jaswal	OK
102	P4294-16	MYDA34	SAM	10/23/24 22:25		Jaswal	OK
103	P4294-17	MYDA35	SAM	10/23/24 22:29		Jaswal	OK
104	P4294-18	MYDA36	SAM	10/23/24 22:32		Jaswal	OK
105	P4294-19	MYDA37	SAM	10/23/24 22:35		Jaswal	OK
106	P4294-20	MYDA42	SAM	10/23/24 22:38	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
107	P4294-21	MYDA42D	DUP	10/23/24 22:41	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
108	P4294-20L	MYDA42L	SD	10/23/24 22:44	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
109	P4294-22	MYDA42S	MS	10/23/24 22:47	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
110	P4294-20MS	MYDA42SRE	MS	10/23/24 22:51	Ba,Co,Ni,V High	Jaswal	Dilution
111	P4295-01	MYD0Z2	SAM	10/23/24 22:53		Jaswal	OK
112	P4295-02	MYD0Z3	SAM	10/23/24 22:57		Jaswal	OK
113	P4295-03	MYD0Z4	SAM	10/23/24 23:00		Jaswal	OK
114	P4295-04	MYD0Z5	SAM	10/23/24 23:03		Jaswal	OK
115	P4295-05	MYD0Z6	SAM	10/23/24 23:06		Jaswal	OK
116	P4295-06	MYD0Z7	SAM	10/23/24 23:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
117	P4295-07	MYD0Z7D	DUP	10/23/24 23:12	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
118	P4295-06L	MYD0Z7L	SD	10/23/24 23:15	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133092

Review By	Sarabjit Jaswal	Review On	10/24/2024 12:10:24 PM
Supervise By	Mohan Bera	Supervise On	10/24/2024 10:33:28 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

119	P4295-08	MYD0Z7S	MS	10/23/24 23:18	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
120	P4295-06MS	MYD0Z7SRE	MS	10/23/24 23:21	MS Fail For Pb, Ba,Co,Ni,V High	Jaswal	Dilution
121	P4295-09	MYD0Z8	SAM	10/23/24 23:24		Jaswal	OK
122	P4295-10	MYD0Z9	SAM	10/23/24 23:27		Jaswal	OK
123	P4295-11	MYD100	SAM	10/23/24 23:30		Jaswal	OK
124	P4295-12	MYD101	SAM	10/23/24 23:33		Jaswal	OK
125	CCV006	CCV006	CCV	10/23/24 23:36		Jaswal	OK
126	CCB006	CCB006	CCB	10/23/24 23:39		Jaswal	OK