

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE003	TUNE	10/24/24 11:46		Jaswal	OK
2	S00	S0	CAL1	10/24/24 12:20		Jaswal	OK
3	S02	S02	CAL3	10/24/24 12:24		Jaswal	OK
4	S03	S03	CAL4	10/24/24 12:27		Jaswal	OK
5	S04	S04	CAL5	10/24/24 12:30		Jaswal	OK
6	S05	S05	CAL6	10/24/24 12:33		Jaswal	OK
7	S06	S06	CAL7	10/24/24 12:36		Jaswal	OK
8	S07	S07	CAL8	10/24/24 12:38		Jaswal	OK
9	S08	S08	CAL9	10/24/24 12:41		Jaswal	OK
10	ICV003	ICV003	ICV	10/24/24 12:56	Inst. paused for calibration review	Jaswal	OK
11	ICB003	ICB003	ICB	10/24/24 12:59		Jaswal	OK
12	ICSA003	ICSA003	ICSA	10/24/24 13:03		Jaswal	OK
13	ICSAB003	ICSAB003	ICSAB	10/24/24 13:06		Jaswal	OK
14	CCV007	CCV007	CCV	10/24/24 13:12	Inst. paused for review	Jaswal	OK
15	CCB007	CCB007	CCB	10/24/24 13:15		Jaswal	OK
16	P4293-07DL	MYDA05	SAM	10/24/24 13:18	4X For Ba,Co,Ni,V	Jaswal	Confirms
17	P4293-08DL	MYDA05D	DUP	10/24/24 13:22	4X For Ba,Co,Ni,V	Jaswal	Confirms
18	P4293-07LDL	MYDA05L	SD	10/24/24 13:25	20X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

19	P4293-09DL	MYDA05S	MS	10/24/24 13:28	4X For Ba,Co,Ni,V	Jaswal	Confirms
20	P4293-07MSDL	MYDA05SRE	MS	10/24/24 13:31	4X For Ba,Co,Ni,V	Jaswal	Confirms
21	P4294-20DL	MYDA42	SAM	10/24/24 13:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
22	P4294-21DL	MYDA42D	DUP	10/24/24 13:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
23	P4294-20LDL	MYDA42L	SD	10/24/24 13:41	20X For Ba,Co,Ni,V	Jaswal	Confirms
24	P4294-22DL	MYDA42S	MS	10/24/24 13:44	4X For Ba,Co,Ni,V	Jaswal	Confirms
25	P4294-20MSDL	MYDA42SRE	MS	10/24/24 13:47	4X For Ba,Co,Ni,V	Jaswal	Confirms
26	P4295-06DL	MYD0Z7	SAM	10/24/24 13:50	4X For Ba,Co,Ni,V	Jaswal	Confirms
27	P4295-07DL	MYD0Z7D	DUP	10/24/24 13:53	4X For Ba,Co,Ni,V	Jaswal	Confirms
28	P4295-06LDL	MYD0Z7L	SD	10/24/24 13:56	20X For Ba,Co,Ni,V	Jaswal	Confirms
29	P4295-08DL	MYD0Z7S	MS	10/24/24 14:00	4X For Ba,Co,Ni,V	Jaswal	Confirms
30	P4295-06MSDL	MYD0Z7SRE	MS	10/24/24 14:03	4X For Ba,Co,Ni,V	Jaswal	Confirms
31	P4295-06A	MYD0Z7A	PS	10/24/24 14:06	PS for Pb - 0.0142 ml - MP82593- -10 ml Sample	Jaswal	OK
32	CCV008	CCV008	CCV	10/24/24 14:09		Jaswal	OK
33	CCB008	CCB008	CCB	10/24/24 14:12		Jaswal	OK
34	P4295-13	MYD102	SAM	10/24/24 14:15		Jaswal	OK
35	P4295-14	MYD103	SAM	10/24/24 14:18		Jaswal	OK
36	P4295-15	MYD104	SAM	10/24/24 14:21		Jaswal	OK
37	P4295-16	MYD105	SAM	10/24/24 14:24		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

38	P4295-17	MYDA38	SAM	10/24/24 14:27		Jaswal	OK
39	P4295-18	MYDA39	SAM	10/24/24 14:31		Jaswal	OK
40	P4295-19	MYDA40	SAM	10/24/24 14:34		Jaswal	OK
41	P4295-20	MYDA41	SAM	10/24/24 14:37		Jaswal	OK
42	P4295-21	MYDBB3	SAM	10/24/24 14:40		Jaswal	OK
43	PB164085BL	PBS085	MB	10/24/24 14:43		Jaswal	OK
44	PB164085BS	LCS085	LCS	10/24/24 14:46		Jaswal	OK
45	PB164085BSD	LCSD085	LCSD	10/24/24 14:50		Jaswal	OK
46	PB164090BL	PBS090	MB	10/24/24 14:53		Jaswal	OK
47	PB164090BS	LCS090	LCS	10/24/24 14:56		Jaswal	OK
48	PB164090BSD	LCSD090	LCSD	10/24/24 14:59		Jaswal	OK
49	P4296-01	MYD0W8	SAM	10/24/24 15:03		Jaswal	OK
50	P4296-02	MYD0W9	SAM	10/24/24 15:06		Jaswal	OK
51	P4296-03	MYD0X0	SAM	10/24/24 15:09		Jaswal	OK
52	P4296-04	MYD0Y2	SAM	10/24/24 15:12		Jaswal	OK
53	P4296-05	MYD0Y3	SAM	10/24/24 15:15	Wrong Sample	Jaswal	Not Ok
54	P4296-06	MYD0Y4	SAM	10/24/24 15:18		Jaswal	OK
55	P4296-07	MYD0Y5	SAM	10/24/24 15:22	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

56	P4296-08	MYD0Y5D	DUP	10/24/24 15:25	Ba,Co,Ni,V High in MSRE(ICB),Bismuth(Dilution),Holmium(Dilution),Lithium(Dilution),Rhodium(Dilution),Scandium(Dilution),Terbium(Dilution),Yttrium(Dilution),	Jaswal	Dilution
57	P4296-07L	MYD0Y5L	SD	10/24/24 15:28	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
58	P4296-09	MYD0Y5S	MS	10/24/24 15:31	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
59	P4296-07MS	MYD0Y5SRE	MS	10/24/24 15:34	MS Fail For As(220.75ppb),Cu(3015.15ppb),Zn(187.38ppb),Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
60	CCV009	CCV009	CCV	10/24/24 15:41		Jaswal	OK
61	CCB009	CCB009	CCB	10/24/24 15:44		Jaswal	OK
62	P4296-10	MYD0Y9	SAM	10/24/24 15:47		Jaswal	OK
63	P4296-11	MYD0Z0	SAM	10/24/24 15:50		Jaswal	OK
64	P4296-12	MYD0Z1	SAM	10/24/24 15:53		Jaswal	OK
65	P4296-13	MYDA43	SAM	10/24/24 15:56		Jaswal	OK
66	P4296-14	MYDA44	SAM	10/24/24 15:59		Jaswal	OK
67	P4296-15	MYDA45	SAM	10/24/24 16:02		Jaswal	OK
68	P4296-16	MYDA46	SAM	10/24/24 16:06		Jaswal	OK
69	P4296-17	MYDA47	SAM	10/24/24 16:09		Jaswal	OK
70	P4296-18	MYDA48	SAM	10/24/24 16:12		Jaswal	OK
71	P4296-19	MYDA49	SAM	10/24/24 16:15		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

72	P4296-20	MYDA50	SAM	10/24/24 16:18		Jaswal	OK
73	P4296-21	MYDA51	SAM	10/24/24 16:21		Jaswal	OK
74	P4296-22	MYDA52	SAM	10/24/24 16:24		Jaswal	OK
75	PB164128BL	PBS128	MB	10/24/24 16:27		Jaswal	OK
76	PB164128BS	LCS128	LCS	10/24/24 16:31		Jaswal	OK
77	PB164128BSD	LCSD128	LCSD	10/24/24 16:34		Jaswal	OK
78	P4297-01	MYDA53	SAM	10/24/24 16:37		Jaswal	OK
79	P4297-02	MYDA54	SAM	10/24/24 16:40		Jaswal	OK
80	P4297-03	MYDA55	SAM	10/24/24 16:43		Jaswal	OK
81	P4297-04	MYDA56	SAM	10/24/24 16:47		Jaswal	OK
82	P4297-05	MYDA57	SAM	10/24/24 16:50		Jaswal	OK
83	P4297-06	MYDA58	SAM	10/24/24 16:53		Jaswal	OK
84	P4297-07	MYDA59	SAM	10/24/24 16:56		Jaswal	OK
85	P4297-08	MYDA60	SAM	10/24/24 16:59		Jaswal	OK
86	P4297-09	MYDA61	SAM	10/24/24 17:02		Jaswal	OK
87	P4297-10	MYDA62	SAM	10/24/24 17:05		Jaswal	OK
88	CCV010	CCV010	CCV	10/24/24 17:14		Jaswal	OK
89	CCB010	CCB010	CCB	10/24/24 17:17		Jaswal	OK
90	P4297-11	MYDA63	SAM	10/24/24 17:20		Jaswal	OK
91	P4297-12	MYDA64	SAM	10/24/24 17:24		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

92	P4297-13	MYDA65	SAM	10/24/24 17:27		Jaswal	OK
93	P4297-14	MYDA66	SAM	10/24/24 17:30	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
94	P4297-15	MYDA66D	DUP	10/24/24 17:33	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
95	P4297-14L	MYDA66L	SD	10/24/24 17:36	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
96	P4297-16	MYDA66S	MS	10/24/24 17:39	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
97	P4297-14MS	MYDA66SRE	MS	10/24/24 17:42	MS Fail For Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
98	P4297-17	MYDA67	SAM	10/24/24 17:45		Jaswal	OK
99	P4297-18	MYDA68	SAM	10/24/24 17:48		Jaswal	OK
100	P4297-19	MYDA69	SAM	10/24/24 17:52		Jaswal	OK
101	P4297-20	MYDA70	SAM	10/24/24 17:55		Jaswal	OK
102	P4297-21	MYDA71	SAM	10/24/24 17:58		Jaswal	OK
103	P4297-22	MYDA72	SAM	10/24/24 18:01		Jaswal	OK
104	P4296-07DL	MYD0Y5	SAM	10/24/24 18:04	4X For Ba,Co,Ni,V	Jaswal	Confirms
105	P4296-08DL	MYD0Y5D	DUP	10/24/24 18:07	4X For Ba,Co,Ni,V	Jaswal	Confirms
106	P4296-07LDL	MYD0Y5L	SD	10/24/24 18:10	20X For Ba,Co,Ni,V	Jaswal	Confirms
107	P4296-09DL	MYD0Y5S	MS	10/24/24 18:14	4X For Ba,Co,Ni,V	Jaswal	Confirms
108	P4296-07MSDL	MYD0Y5SRE	MS	10/24/24 18:17	4X For Ba,Co,Ni,V	Jaswal	Confirms
109	P4297-14DL	MYDA66	SAM	10/24/24 18:20	4X For Ba,Co,Ni,V	Jaswal	Confirms
110	P4297-15DL	MYDA66D	DUP	10/24/24 18:23	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

111	P4297-14LDL	MYDA66L	SD	10/24/24 18:26	20X For Ba,Co,Ni,V	Jaswal	Confirms
112	P4297-16DL	MYDA66S	MS	10/24/24 18:30	4X For Ba,Co,Ni,V	Jaswal	Confirms
113	CCV011	CCV011	CCV	10/24/24 18:49		Jaswal	OK
114	CCB011	CCB011	CCB	10/24/24 18:52		Jaswal	OK
115	P4297-14MSDL	MYDA66SRE	MS	10/24/24 18:55	4X For Ba,Co,Ni,V	Jaswal	Confirms
116	PB164126BL	PBS126	MB	10/24/24 19:02		Jaswal	OK
117	PB164126BS	LCS126	LCS	10/24/24 19:06		Jaswal	OK
118	PB164126BSD	LCSD126	LCSD	10/24/24 19:09		Jaswal	OK
119	P4298-01	MYDA73	SAM	10/24/24 19:12		Jaswal	OK
120	P4298-02	MYDA74	SAM	10/24/24 19:16		Jaswal	OK
121	P4298-03	MYDA75	SAM	10/24/24 19:19		Jaswal	OK
122	P4298-04	MYDA76	SAM	10/24/24 19:22		Jaswal	OK
123	P4298-05	MYDA77	SAM	10/24/24 19:25		Jaswal	OK
124	P4298-06	MYDA78	SAM	10/24/24 19:28		Jaswal	OK
125	P4298-07	MYDA79	SAM	10/24/24 19:31		Jaswal	OK
126	P4298-08	MYDA80	SAM	10/24/24 19:34		Jaswal	OK
127	P4298-09	MYDA81	SAM	10/24/24 19:37		Jaswal	OK
128	P4298-10	MYDA82	SAM	10/24/24 19:40		Jaswal	OK
129	P4298-11	MYDA83	SAM	10/24/24 19:43		Jaswal	OK
130	P4298-12	MYDA84	SAM	10/24/24 19:47		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

131	P4298-13	MYDA85	SAM	10/24/24 19:50		Jaswal	OK
132	P4298-14	MYDA86	SAM	10/24/24 19:53		Jaswal	OK
133	P4298-15	MYDA87	SAM	10/24/24 19:56		Jaswal	OK
134	P4298-16	MYDA88	SAM	10/24/24 19:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
135	P4298-17	MYDA88D	DUP	10/24/24 20:02	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
136	P4298-16L	MYDA88L	SD	10/24/24 20:05	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
137	P4298-18	MYDA88S	MS	10/24/24 20:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
138	P4298-16MS	MYDA88SRE	MS	10/24/24 20:12	MS Fail For As(356.29ppb),Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
139	P4298-19	MYDA89	SAM	10/24/24 20:15		Jaswal	OK
140	CCV012	CCV012	CCV	10/24/24 20:18		Jaswal	OK
141	CCB012	CCB012	CCB	10/24/24 20:20		Jaswal	OK
142	P4298-20	MYDA90	SAM	10/24/24 20:24		Jaswal	OK
143	P4298-21	MYDA91	SAM	10/24/24 20:27		Jaswal	OK
144	P4298-22	MYDA92	SAM	10/24/24 20:30		Jaswal	OK
145	P4299-01	MYDA93	SAM	10/24/24 20:33		Jaswal	OK
146	P4299-02	MYDA94	SAM	10/24/24 20:36		Jaswal	OK
147	P4299-03	MYDA95	SAM	10/24/24 20:39		Jaswal	OK
148	P4299-04	MYDA96	SAM	10/24/24 20:42		Jaswal	OK
149	P4299-05	MYDA97	SAM	10/24/24 20:45		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

150	P4299-06	MYDA98	SAM	10/24/24 20:48		Jaswal	OK
151	P4299-07	MYDA99	SAM	10/24/24 20:52		Jaswal	OK
152	P4299-08	MYDAA0	SAM	10/24/24 20:55		Jaswal	OK
153	P4299-09	MYDAA1	SAM	10/24/24 20:58		Jaswal	OK
154	P4299-10	MYDAA2	SAM	10/24/24 21:01		Jaswal	OK
155	P4299-11	MYDAA3	SAM	10/24/24 21:04		Jaswal	OK
156	P4299-12	MYDAA4	SAM	10/24/24 21:07		Jaswal	OK
157	P4299-13	MYDAA5	SAM	10/24/24 21:10		Jaswal	OK
158	P4299-14	MYDAA6	SAM	10/24/24 21:13		Jaswal	OK
159	P4299-15	MYDAA7	SAM	10/24/24 21:16		Jaswal	OK
160	P4299-16	MYDAA8	SAM	10/24/24 21:19		Jaswal	OK
161	P4299-17	MYDAA9	SAM	10/24/24 21:22		Jaswal	OK
162	P4299-18	MYDAB0	SAM	10/24/24 21:26		Jaswal	OK
163	P4299-19	MYDAB1	SAM	10/24/24 21:29	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
164	P4299-20	MYDAB1D	DUP	10/24/24 21:32	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
165	P4299-19L	MYDAB1L	SD	10/24/24 21:35	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
166	P4299-21	MYDAB1S	MS	10/24/24 21:38	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
167	P4299-19MS	MYDAB1SRE	MS	10/24/24 21:41	MS Fail For As(146.42ppb),Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
168	P4299-22	MYDAB2	SAM	10/24/24 21:44		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QCBatch ID # LB133117

Review By	Sarabjit Jaswal	Review On	10/25/2024 6:26:24 AM
Supervise By	Mohan Bera	Supervise On	10/25/2024 6:28:50 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

169	CCV013	CCV013	CCV	10/24/24 21:47		Jaswal	OK
170	CCB013	CCB013	CCB	10/24/24 21:51		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133137

Review By	Sarabjit Jaswal	Review On	10/25/2024 5:57:56 PM
Supervise By	Mohan Bera	Supervise On	10/25/2024 5:58:38 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE005	TUNE	10/25/24 13:54		Jaswal	OK
2	S00	S0	CAL1	10/25/24 14:34		Jaswal	OK
3	S02	S02	CAL3	10/25/24 14:37		Jaswal	OK
4	S03	S03	CAL4	10/25/24 14:40		Jaswal	OK
5	S04	S04	CAL5	10/25/24 14:43		Jaswal	OK
6	S05	S05	CAL6	10/25/24 14:46		Jaswal	OK
7	S06	S06	CAL7	10/25/24 14:49		Jaswal	OK
8	S07	S07	CAL8	10/25/24 14:52		Jaswal	OK
9	S08	S08	CAL9	10/25/24 14:54		Jaswal	OK
10	ICV005	ICV005	ICV	10/25/24 15:00	Inst. paused for Calibration review	Jaswal	OK
11	ICB005	ICB005	ICB	10/25/24 15:03		Jaswal	OK
12	ICSA005	ICSA005	ICSA	10/25/24 15:14	Inst. paused for review	Jaswal	OK
13	ICSAB005	ICSAB005	ICSAB	10/25/24 15:17		Jaswal	OK
14	CCV019	CCV019	CCV	10/25/24 15:27	Inst. paused for review	Jaswal	OK
15	CCB019	CCB019	CCB	10/25/24 15:33	Inst. paused for review	Jaswal	OK
16	PB164088BL	PBS088	MB	10/25/24 16:22		Jaswal	OK
17	PB164088BS	LCS088	LCS	10/25/24 16:26		Jaswal	OK
18	PB164088BSD	LCSD088	LCSD	10/25/24 16:29		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133137

Review By	Sarabjit Jaswal	Review On	10/25/2024 5:57:56 PM
Supervise By	Mohan Bera	Supervise On	10/25/2024 5:58:38 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

19	P4296-05	MYD0Y3	SAM	10/25/24 16:42		Jaswal	OK
20	P4296-07A	MYD0Y5ARE	PS	10/25/24 16:45	PS For As,Cu,Ag,Zn	Jaswal	OK
21	P4301-09A	MYDAC1ARE	PS	10/25/24 16:48	PS For Ag - 0.002 ml - MP82908-10 ml Sample	Jaswal	OK
22	P4302-09A	MYDAE1ARE	PS	10/25/24 16:53	PS For Ag - 0.002 ml - MP82908-10 ml Sample	Jaswal	OK
23	CCV020	CCV020	CCV	10/25/24 17:03		Jaswal	OK
24	CCB020	CCB020	CCB	10/25/24 17:08		Jaswal	OK