

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDA73
 Instrument ID: P7 Start Date: 10/24/2024
 Analytical Method: ICP-MS Run Batch: LB133117
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	1.9	-5	50	52	4
Arsenic	0	0.0	0	1	1.0	0	50	53	6
Barium	0	0.0	0	10	9.2	-8	250	250	0
Beryllium	0	0.0	0	1	1.1	10	50	56	12
Cadmium	0	0.0	0	1	1.1	10	50	55	10
Chromium	0	0.0	0	2	2.0	0	50	53	6
Cobalt	0	0.0	0	1	1.0	0	50	52	4
Copper	0	0.0	0	2	2.1	5	500	530	6
Lead	0	0.0	0	1	0.85	-15	250	250	0
Nickel	0	0.0	0	1	1.0	0	50	53	6
Selenium	0	0.0	0	5	5.2	4	50	53	6
Silver	0	0.0	0	1	0.98	-2	50	55	10
Thallium	0	0.0	0	1	0.86	-14	50	45	-10
Vanadium	0	0.0	0	5	4.9	-2	50	53	6
Zinc	0	0.0	0	5	5.0	0	500	530	6

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDA73
 Instrument ID: P7 Start Date: 10/24/2024
 Analytical Method: ICP-MS Run Batch: LB133117
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	130	4	250	260	4	500	500	0
Arsenic	125	130	4	250	250	0	500	490	-2
Barium	625	650	4	1250	1300	4	2500	2500	0
Beryllium	125	140	12	250	260	4	500	510	2
Cadmium	125	130	4	250	260	4	500	490	-2
Chromium	125	130	4	250	260	4	500	500	0
Cobalt	125	130	4	250	250	0	500	500	0
Copper	1250	1300	4	2500	2600	4	5000	5000	0
Lead	625	630	1	1250	1200	-4	2500	2500	0
Nickel	125	130	4	250	260	4	500	500	0
Selenium	125	130	4	250	260	4	500	500	0
Silver	125	140	12	250	270	8	500	510	2
Thallium	125	130	4	250	250	0	500	500	0
Vanadium	125	130	4	250	250	0	500	500	0
Zinc	1250	1300	4	2500	2600	4	5000	5000	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDA73
Instrument ID: P7 Start Date: 10/24/2024
Analytical Method: ICP-MS Run Batch: LB133117
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	1000	0						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	990	-1						
Cadmium	1000	1000	0						
Chromium	1000	1000	0						
Cobalt	1000	1000	0						
Copper	10000	9900	-1						
Lead	5000	5000	0						
Nickel	1000	1000	0						
Selenium	1000	1000	0						
Silver	1000	990	-1						
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	10000	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDA73
 Instrument ID: P7 Start Date: 10/25/2024
 Analytical Method: ICP-MS Run Batch: LB133124
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	2.0	0	50	50	0
Arsenic	0	0.0	0	1	0.96	-4	50	50	0
Barium	0	0.0	0	10	9.4	-6	250	250	0
Beryllium	0	0.0	0	1	1.1	10	50	55	10
Cadmium	0	0.0	0	1	1.1	10	50	52	4
Chromium	0	0.0	0	2	2.1	5	50	52	4
Cobalt	0	0.0	0	1	1.1	10	50	52	4
Copper	0	0.0	0	2	2.2	10	500	520	4
Lead	0	0.0	0	1	0.86	-14	250	240	-4
Nickel	0	0.0	0	1	1.1	10	50	52	4
Selenium	0	0.0	0	5	5.2	4	50	52	4
Silver	0	0.0	0	1	0.96	-4	50	53	6
Thallium	0	0.0	0	1	0.86	-14	50	46	-8
Vanadium	0	0.0	0	5	5.2	4	50	52	4
Zinc	0	0.0	0	5	6.4	28	500	520	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDA73
 Instrument ID: P7 Start Date: 10/25/2024
 Analytical Method: ICP-MS Run Batch: LB133124
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	130	4	250	260	4	500	500	0
Arsenic	125	130	4	250	250	0	500	490	-2
Barium	625	640	2	1250	1300	4	2500	2500	0
Beryllium	125	140	12	250	270	8	500	530	6
Cadmium	125	130	4	250	260	4	500	500	0
Chromium	125	130	4	250	260	4	500	500	0
Cobalt	125	130	4	250	250	0	500	500	0
Copper	1250	1300	4	2500	2600	4	5000	5100	2
Lead	625	620	-1	1250	1200	-4	2500	2500	0
Nickel	125	130	4	250	260	4	500	500	0
Selenium	125	130	4	250	250	0	500	500	0
Silver	125	130	4	250	260	4	500	500	0
Thallium	125	120	-4	250	240	-4	500	500	0
Vanadium	125	130	4	250	260	4	500	500	0
Zinc	1250	1300	4	2500	2600	4	5000	5100	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDA73
Instrument ID: P7 Start Date: 10/25/2024
Analytical Method: ICP-MS Run Batch: LB133124
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	1000	0						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	980	-2						
Cadmium	1000	1000	0						
Chromium	1000	1000	0						
Cobalt	1000	1000	0						
Copper	10000	9900	-1						
Lead	5000	5000	0						
Nickel	1000	1000	0						
Selenium	1000	1000	0						
Silver	1000	990	-1						
Thallium	1000	1000	0						
Vanadium	1000	990	-1						
Zinc	10000	9900	-1						