

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDAD3Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB133124 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV004				ID: CCV014				ID: CCV015		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	210	103	2	500	500	101	1	490	97	0
Arsenic	200	200	99	1	500	480	96	1	480	96	1
Barium	100	98	98	1	2500	2500	100	0	2500	98	1
Beryllium	100	110	109	1	500	540	107	1	550	110	1
Cadmium	100	110	105	2	500	520	104	2	510	102	1
Chromium	100	100	105	0	500	490	98	1	500	100	1
Cobalt	100	100	104	1	500	520	104	1	540	108	0
Copper	100	98	98	0	5000	5000	100	0	5200	105	0
Lead	200	200	100	1	2500	2500	101	2	2500	101	1
Nickel	110	100	94	0	500	500	101	0	530	105	0
Selenium	200	210	104	1	500	470	94	1	470	95	1
Silver	50.0	52	103	2	500	490	98	2	470	95	0
Thallium	210	210	102	2	500	510	102	1	510	102	1
Vanadium	100	100	102	1	500	500	99	0	510	102	0
Zinc	200	200	100	1	5000	4900	98	0	5100	103	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDAD3
 Initial Calibration Verification Source : MP82578
 Continuing Calibration Verification Source : MP82579
 Run Batch: LB133124 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV016				ID: CCV017		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	480	96	1	480	95	1
Arsenic					500	470	93	1	470	94	1
Barium					2500	2400	98	1	2400	97	2
Beryllium					500	530	106	2	520	104	1
Cadmium					500	490	99	1	490	99	1
Chromium					500	500	100	1	490	99	1
Cobalt					500	540	108	1	540	107	1
Copper					5000	5200	105	1	5200	105	0
Lead					2500	2500	101	0	2500	100	0
Nickel					500	520	104	0	520	104	0
Selenium					500	460	93	1	470	93	0
Silver					500	460	93	1	460	92	1
Thallium					500	510	101	2	500	101	2
Vanadium					500	500	101	1	500	101	1
Zinc					5000	5200	103	1	5200	103	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDAD3
 Initial Calibration Verification Source : MP82578
 Continuing Calibration Verification Source : MP82579
 Run Batch: LB133137 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV005				ID: CCV019				ID: CCV020		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	210	103	1	500	490	97	1	520	105	3
Arsenic	200	200	100	1	500	480	96	1	490	97	1
Barium	100	96	96	1	2500	2400	98	0	2500	100	2
Beryllium	100	99	99	2	500	530	105	1	550	110	1
Cadmium	100	100	104	1	500	500	99	1	520	103	2
Chromium	100	100	105	1	500	490	98	1	490	98	0
Cobalt	100	100	104	1	500	500	101	1	510	102	1
Copper	100	97	97	0	5000	5000	99	1	4900	99	0
Lead	200	200	99	1	2500	2500	101	2	2500	101	1
Nickel	110	100	95	1	500	490	98	1	500	100	1
Selenium	200	210	104	1	500	470	94	1	480	96	2
Silver	50.0	51	101	1	500	480	97	2	520	103	2
Thallium	210	210	101	2	500	510	102	1	520	103	1
Vanadium	100	100	101	0	500	490	99	1	500	99	1
Zinc	200	200	99	1	5000	4900	98	1	5000	99	2