

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACECase No.: 51772

MA No.: _____

SDG No.: MYDAE7Instrument ID: P5Start Date : 10/21/2024Analytical Method: ICP-AESRun Batch : LB133036

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999981	0.112078	-0.003156	WLR Blank	Inverse Conc
Antimony	0.997706	0.203567	0.000269	WLR Blank	Inverse Conc
Arsenic	0.999742	0.026566	0.000007	WLR Blank	Inverse Conc
Barium	0.999942	5.350314	0.044257	WLR Blank	Inverse Conc
Beryllium	0.999866	6.605957	-0.002206	WLR Blank	Inverse Conc
Boron	0.999255	0.986751	0.008926	WLR Blank	Inverse Conc
Cadmium	0.999889	2.189967	0.000328	WLR Blank	Inverse Conc
Calcium	0.999877	0.075714	0.003695	WLR Blank	Inverse Conc
Chromium	0.999729	0.322153	0.000042	WLR Blank	Inverse Conc
Cobalt	0.999969	0.654368	-0.000307	WLR Blank	Inverse Conc
Copper	0.999983	0.198942	0.001697	WLR Blank	Inverse Conc
Iron	0.999587	0.015766	0.000648	WLR Blank	Inverse Conc
Lead	0.999994	0.151655	-0.000516	WLR Blank	Inverse Conc
Lithium	0.999677	1.660774	0.007312	WLR Blank	Inverse Conc
Magnesium	0.999607	0.006851	-0.000402	WLR Blank	Inverse Conc
Manganese	0.999738	0.346976	-0.000331	WLR Blank	Inverse Conc
Molybdenum	0.998939	0.992161	-0.000193	WLR Blank	Inverse Conc
Nickel	0.999914	0.618772	-0.000449	WLR Blank	Inverse Conc
Phosphorus	0.999800	0.055904	-0.001058	WLR Blank	Inverse Conc
Potassium	0.999901	0.032423	0.046538	WLR Blank	Inverse Conc
Selenium	0.998588	0.080470	0.000582	WLR Blank	Inverse Conc
Silicon	0.994714	1.465690	0.014349	WLR Blank	Inverse Conc
Silver	0.999957	0.337429	0.028929	WLR Blank	Inverse Conc
Sodium	0.999963	0.011378	-0.005033	WLR Blank	Inverse Conc
Strontium	0.999913	11.674492	0.001763	WLR Blank	Inverse Conc
Sulfur	0.999655	0.028646	0.000427	WLR Blank	Inverse Conc
Thallium	0.999622	0.077442	-0.000336	WLR Blank	Inverse Conc
Tin	0.999993	0.107204	0.000438	WLR Blank	Inverse Conc
Vanadium	0.999816	0.086834	-0.000900	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDAE7
Instrument ID: P5 Start Date : 10/21/2024
Analytical Method: ICP-AES Run Batch : LB133036

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.998653	2.928286	0.002282	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACECase No.: 51772

MA No.: _____

SDG No.: MYDAE7Instrument ID: P5Start Date : 10/23/2024Analytical Method: ICP-AESRun Batch : LB133105

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999984	0.117102	-0.004383	WLR Blank	Inverse Conc
Antimony	0.997995	0.216420	-0.000109	WLR Blank	Inverse Conc
Arsenic	0.999796	0.035542	-0.000211	WLR Blank	Inverse Conc
Barium	0.999915	5.853687	0.031335	WLR Blank	Inverse Conc
Beryllium	0.999823	6.166224	-0.002784	WLR Blank	Inverse Conc
Boron	0.999346	1.021182	0.005887	WLR Blank	Inverse Conc
Cadmium	0.999825	2.190312	0.000381	WLR Blank	Inverse Conc
Calcium	0.999678	0.076093	0.004191	WLR Blank	Inverse Conc
Chromium	0.999569	0.210716	-0.000236	WLR Blank	Inverse Conc
Cobalt	0.999945	0.668366	-0.000412	WLR Blank	Inverse Conc
Copper	0.999945	0.206040	0.001411	WLR Blank	Inverse Conc
Iron	0.999193	0.015325	0.000295	WLR Blank	Inverse Conc
Lead	0.999964	0.154371	-0.000731	WLR Blank	Inverse Conc
Lithium	0.999860	1.819000	0.018018	WLR Blank	Inverse Conc
Magnesium	0.999961	0.006828	-0.000175	WLR Blank	Inverse Conc
Manganese	0.999704	0.322849	0.000055	WLR Blank	Inverse Conc
Molybdenum	0.999108	1.045786	-0.000039	WLR Blank	Inverse Conc
Nickel	0.999864	0.631384	-0.000442	WLR Blank	Inverse Conc
Phosphorus	0.999840	0.059235	-0.000999	WLR Blank	Inverse Conc
Potassium	0.999946	0.034280	0.046025	WLR Blank	Inverse Conc
Selenium	0.998646	0.086639	0.000458	WLR Blank	Inverse Conc
Silicon	0.994627	1.479083	0.011298	WLR Blank	Inverse Conc
Silver	0.999945	0.344548	0.026407	WLR Blank	Inverse Conc
Sodium	0.999989	0.011931	-0.008731	WLR Blank	Inverse Conc
Strontium	0.999854	12.366566	-0.000095	WLR Blank	Inverse Conc
Sulfur	0.999712	0.030427	0.000275	WLR Blank	Inverse Conc
Thallium	0.999479	0.078570	-0.000432	WLR Blank	Inverse Conc
Tin	0.999962	0.109578	0.000265	WLR Blank	Inverse Conc
Vanadium	0.999910	0.087259	0.000076	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDAE7
Instrument ID: P5 Start Date : 10/23/2024
Analytical Method: ICP-AES Run Batch : LB133105

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.999168	3.058256	0.001681	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDAE7
Instrument ID: P5 Start Date : 10/23/2024
Analytical Method: ICP-AES Run Batch : LB133105

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------