

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1555	100		100		100		100		100	
S02	1558	104		102		103		102		103	
S03	1601	104		105		106		104		106	
S04	1605	107		106		108		106		109	
S05	1608	111		107		109		107		110	
S06	1610	109		107		111		107		114	
S07	1616	107		107		112		107		116	
S08	1619	110		108		109		100		113	
ICV002	1629	101		112		112		109		109	
ICB002	1637	98		104		106		106		106	
ICSA002	1643	97		105		108		103		108	
ICSAB002	1647	100		113		117		112		118	
CCV007	1658	97		112		115		105		115	
CCB007	1707	93		107		110		107		110	
ZZZZZZ	1711										
ZZZZZZ	1714										
ZZZZZZ	1718										
ZZZZZZ	1721										
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1731										
ZZZZZZ	1737										
ZZZZZZ	1741										
ZZZZZZ	1744										
ZZZZZZ	1747										
ZZZZZZ	1750										
ZZZZZZ	1754										
ZZZZZZ	1757										
ZZZZZZ	1800										
ZZZZZZ	1803										
ZZZZZZ	1807										
ZZZZZZ	1810										
ZZZZZZ	1813										
ZZZZZZ	1817										
CCV008	1820	98		113		116		107		117	
CCB008	1823	89		106		108		106		107	
ZZZZZZ	1826										
ZZZZZZ	1829										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		6Li	Q	45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q
ZZZZZZ	1833										
ZZZZZZ	1836										
ZZZZZZ	1839										
ZZZZZZ	1842										
ZZZZZZ	1846										
ZZZZZZ	1849										
ZZZZZZ	1852										
ZZZZZZ	1855										
ZZZZZZ	1858										
ZZZZZZ	1901										
ZZZZZZ	1905										
ZZZZZZ	1908										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1921										
ZZZZZZ	1924										
ZZZZZZ	1927										
ZZZZZZ	1930										
ZZZZZZ	1933										
CCV009	1946	94		115		118		109		118	
CCB009	1949	90		108		110		108		110	
ZZZZZZ	1952										
ZZZZZZ	1955										
ZZZZZZ	1959										
ZZZZZZ	2002										
ZZZZZZ	2005										
ZZZZZZ	2008										
ZZZZZZ	2012										
ZZZZZZ	2015										
ZZZZZZ	2018										
ZZZZZZ	2021										
PBS189	2025	87		110		114		113		112	
LCS189	2028	87		112		114		112		114	
LCSD189	2031	88		112		115		113		113	
MYDAQ6	2035	91		119		124		112		116	
MYDAQ6D	2039	92		120		124		113		116	
MYDAQ6L	2042	90		113		116		113		115	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		6Li	Q	45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q
MYDAQ6S	2045	93		120		124		113		116	
MYDAQ6SRE	2048	94		119		124		112		116	
MYDAQ6	2101	93		116		118		111		114	
MYDAQ6D	2104	91		114		116		113		114	
MYDAQ6L	2108	92		111		116		112		114	
MYDAQ6S	2111	93		114		116		111		114	
MYDAQ6SRE	2114	94		114		116		112		114	
MYDAR4	2118	95		119		124		111		115	
CCV010	2121	93		111		114		105		114	
CCB010	2124	86		105		108		107		107	
MYDAR5	2127	92		121		125		114		118	
MYDAR6	2131	91		121		126	*	113		116	
MYDAR7	2134	92		121		126	*	112		116	
MYDAR5	2137	94		117		119		112		115	
MYDAR6	2140	93		116		120		114		114	
MYDAR7	2143	96		116		120		111		114	
MYDAR8	2147	95		120		126	*	112		117	
MYDAR9	2150	96		119		125		111		115	
MYDAS0	2153	93		116		122		111		116	
MYDAS1	2156	90		118		125		110		114	
MYDAS2	2200	90		119		124		111		114	
MYDAS3	2203	87		117		124		107		112	
MYDAS4	2206	88		121		126	*	111		115	
MYDAS5	2209	86		121		125		109		114	
MYDAS6	2212	88		121		125		110		116	
MYDAS7	2216	91		119		125		112		116	
MYDAS8	2219	91		126	*	133	*	113		118	
MYDAS9	2222	93		124		130	*	115		119	
MYDAT0	2225	91		120		126	*	111		114	
MYDAT1	2228	89		118		124		111		116	
MYDAT2	2232	85		117		120		107		113	
MYDAR8	2235	88		118		122		113		116	
MYDAS4	2238	86		115		121		110		116	
MYDAS6	2241	86		117		122		112		115	
MYDAS7	2247	87		116		120		111		114	
MYDAS8	2250	89		120		124		111		114	
CCV011	2257	89		111		114		104		113	
CCB011	2307	84		107		110		108		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		6Li	Q	45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q
MYDAS9	2317	85		114		117		110		111	
ZZZZZZ	2322										
ZZZZZZ	2325										
ZZZZZZ	2333										
MYDAT1	2337	90		115		118		108		110	
ZZZZZZ	2343										
ZZZZZZ	2346										
ZZZZZZ	2350										
ZZZZZZ	2353										
ZZZZZZ	2359										
ZZZZZZ	0004										
ZZZZZZ	0008										
ZZZZZZ	0011										
ZZZZZZ	0014										
ZZZZZZ	0017										
ZZZZZZ	0021										
ZZZZZZ	0024										
ZZZZZZ	0027										
ZZZZZZ	0030										
ZZZZZZ	0034										
ZZZZZZ	0037										
ZZZZZZ	0040										
ZZZZZZ	0043										
CCV012	0047	84		111		114		105		114	
CCB012	0050	86		109		113		111		113	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		45Sc		89Y		103Rh		159Tb		165Ho	
S0	1555	100		100		100		100		100	
S02	1558	101		98		100		100		101	
S03	1601	103		101		104		103		103	
S04	1605	103		101		103		106		106	
S05	1608	102		102		103		105		106	
S06	1610	101		102		102		107		108	
S07	1616	101		102		102		109		109	
S08	1619	106		105		98		108		109	
ICV002	1629	110		108		108		106		107	
ICB002	1637	104		103		104		101		102	
ICSA002	1643	105		106		103		107		106	
ICSAB002	1647	108		107		104		108		108	
CCV007	1658	110		106		102		108		108	
CCB007	1707	106		105		108		106		105	
ZZZZZZ	1711										
ZZZZZZ	1714										
ZZZZZZ	1718										
ZZZZZZ	1721										
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1731										
ZZZZZZ	1737										
ZZZZZZ	1741										
ZZZZZZ	1744										
ZZZZZZ	1747										
ZZZZZZ	1750										
ZZZZZZ	1754										
ZZZZZZ	1757										
ZZZZZZ	1800										
ZZZZZZ	1803										
ZZZZZZ	1807										
ZZZZZZ	1810										
ZZZZZZ	1813										
ZZZZZZ	1817										
CCV008	1820	111		107		103		108		109	
CCB008	1823	105		105		105		103		103	
ZZZZZZ	1826										
ZZZZZZ	1829										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		45Sc		89Y		103Rh		159Tb		165Ho	
ZZZZZZ	1833										
ZZZZZZ	1836										
ZZZZZZ	1839										
ZZZZZZ	1842										
ZZZZZZ	1846										
ZZZZZZ	1849										
ZZZZZZ	1852										
ZZZZZZ	1855										
ZZZZZZ	1858										
ZZZZZZ	1901										
ZZZZZZ	1905										
ZZZZZZ	1908										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1921										
ZZZZZZ	1924										
ZZZZZZ	1927										
ZZZZZZ	1930										
ZZZZZZ	1933										
CCV009	1946	112		111		105		109		110	
CCB009	1949	106		106		108		104		106	
ZZZZZZ	1952										
ZZZZZZ	1955										
ZZZZZZ	1959										
ZZZZZZ	2002										
ZZZZZZ	2005										
ZZZZZZ	2008										
ZZZZZZ	2012										
ZZZZZZ	2015										
ZZZZZZ	2018										
ZZZZZZ	2021										
PBS189	2025	109		108		109		105		106	
LCS189	2028	109		109		109		106		107	
LCSD189	2031	110		110		109		107		107	
MYDAQ6	2035	117		118		110		109		109	
MYDAQ6D	2039	118		119		111		109		110	
MYDAQ6L	2042	107		106		106		103		104	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
MYDAQ6S	2045	116		117		108		109		109	
MYDAQ6SRE	2048	114		115		108		108		109	
MYDAQ6	2101	113		114		112		108		108	
MYDAQ6D	2104	111		111		110		108		107	
MYDAQ6L	2108	109		108		109		107		108	
MYDAQ6S	2111	110		111		109		107		108	
MYDAQ6SRE	2114	109		110		109		108		109	
MYDAR4	2118	116		118		109		108		110	
CCV010	2121	109		107		103		107		109	
CCB010	2124	103		104		105		103		104	
MYDAR5	2127	120		119		112		110		109	
MYDAR6	2131	118		119		111		109		110	
MYDAR7	2134	120		121		110		110		111	
MYDAR5	2137	113		114		110		107		110	
MYDAR6	2140	113		113		109		108		109	
MYDAR7	2143	114		113		109		108		108	
MYDAR8	2147	117		120		109		110		109	
MYDAR9	2150	116		117		109		107		109	
MYDAS0	2153	114		115		106		107		107	
MYDAS1	2156	117		119		107		107		108	
MYDAS2	2200	117		119		110		107		109	
MYDAS3	2203	119		120		108		110		109	
MYDAS4	2206	120		120		108		108		109	
MYDAS5	2209	120		120		106		107		107	
MYDAS6	2212	119		119		108		109		108	
MYDAS7	2216	116		117		108		109		108	
MYDAS8	2219	123		126	*	110		110		110	
MYDAS9	2222	120		121		112		111		112	
MYDAT0	2225	117		118		107		107		108	
MYDAT1	2228	114		115		106		106		107	
MYDAT2	2232	116		115		106		104		104	
MYDAR8	2235	113		114		109		106		107	
MYDAS4	2238	116		114		110		107		107	
MYDAS6	2241	115		116		110		108		108	
MYDAS7	2247	114		114		110		107		108	
MYDAS8	2250	115		116		108		107		106	
CCV011	2257	109		106		101		106		108	
CCB011	2307	101		101		101		96		98	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/27/2024
 Run Batch: LB133170 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
MYDAS9	2317	115		115		109		106		105	
ZZZZZZ	2322										
ZZZZZZ	2325										
ZZZZZZ	2333										
MYDAT1	2337	114		113		107		105		105	
ZZZZZZ	2343										
ZZZZZZ	2346										
ZZZZZZ	2350										
ZZZZZZ	2353										
ZZZZZZ	2359										
ZZZZZZ	0004										
ZZZZZZ	0008										
ZZZZZZ	0011										
ZZZZZZ	0014										
ZZZZZZ	0017										
ZZZZZZ	0021										
ZZZZZZ	0024										
ZZZZZZ	0027										
ZZZZZZ	0030										
ZZZZZZ	0034										
ZZZZZZ	0037										
ZZZZZZ	0040										
ZZZZZZ	0043										
CCV012	0047	108		106		102		107		107	
CCB012	0050	106		107		109		106		108	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1322	100		100		100		100		100	
S02	1326	98		99		99		99		99	
S03	1329	96		97		97		98		98	
S04	1332	102		99		100		99		101	
S05	1335	109		100		101		99		103	
S06	1338	113		100		103		100		104	
S07	1341	116		99		103		99		105	
S08	1344	112		99		99		91		103	
ICV003	1403	104		100		100		99		101	
ICB003	1406	101		98		98		97		98	
ICSA004	1410	107		99		102		98		105	
ICSAB004	1413	108		104		105		100		107	
CCV018	1416	106		102		105		97		107	
CCB018	1419	109		102		104		103		107	
ZZZZZZ	1422										
ZZZZZZ	1425										
ZZZZZZ	1429										
ZZZZZZ	1432										
ZZZZZZ	1435										
ZZZZZZ	1438										
ZZZZZZ	1442										
ZZZZZZ	1446										
ZZZZZZ	1449										
ZZZZZZ	1452										
ZZZZZZ	1455										
ZZZZZZ	1458										
ZZZZZZ	1502										
ZZZZZZ	1505										
ZZZZZZ	1508										
ZZZZZZ	1511										
ZZZZZZ	1515										
ZZZZZZ	1518										
ZZZZZZ	1521										
ZZZZZZ	1525										
CCV019	1528	98		98		102		95		105	
CCB019	1535	97		101		102		102		104	
ZZZZZZ	1539										
ZZZZZZ	1542										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		6Li	Q	45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q
ZZZZZZ	1545										
ZZZZZZ	1549										
ZZZZZZ	1552										
ZZZZZZ	1555										
ZZZZZZ	1559										
ZZZZZZ	1602										
ZZZZZZ	1605										
ZZZZZZ	1609										
ZZZZZZ	1612										
ZZZZZZ	1615										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1638										
ZZZZZZ	1641										
ZZZZZZ	1645										
ZZZZZZ	1648										
ZZZZZZ	1651										
CCV020	1655	99		99		104		96		106	
CCB020	1658	101		98		101		102		106	
ZZZZZZ	1701										
ZZZZZZ	1705										
ZZZZZZ	1708										
ZZZZZZ	1711										
ZZZZZZ	1714										
MYDAQ6ARE	1717	97		108		114		103		110	
MYDAT0	1720	99		119		127	*	101		110	
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1730										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
CCV021	1743	92		97		103		94		104	
CCB021	1745	96		99		103		103		106	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		6Li	Q	45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q
ZZZZZZ	1749										
ZZZZZZ	1753										
ZZZZZZ	1756										
ZZZZZZ	1759										
ZZZZZZ	1802										
ZZZZZZ	1805										
ZZZZZZ	1809										
ZZZZZZ	1812										
ZZZZZZ	1815										
ZZZZZZ	1818										
ZZZZZZ	1821										
ZZZZZZ	1824										
ZZZZZZ	1828										
MYDAT0	1831	91		105		106		102		106	
CCV022	1834	99		98		102		96		106	
CCB022	1837	105		102		105		106		110	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		45Sc		89Y		103Rh		159Tb		165Ho	
S0	1322	100		100		100		100		100	
S02	1326	100		100		99		101		102	
S03	1329	99		99		99		101		101	
S04	1332	100		100		99		102		105	
S05	1335	99		100		100		104		107	
S06	1338	100		100		100		107		109	
S07	1341	97		98		97		107		108	
S08	1344	99		100		92		106		107	
ICV003	1403	100		100		101		102		103	
ICB003	1406	97		98		98		100		100	
ICSA004	1410	102		104		100		109		110	
ICSAB004	1413	104		104		102		110		111	
CCV018	1416	102		104		98		107		109	
CCB018	1419	99		102		104		107		108	
ZZZZZZ	1422										
ZZZZZZ	1425										
ZZZZZZ	1429										
ZZZZZZ	1432										
ZZZZZZ	1435										
ZZZZZZ	1438										
ZZZZZZ	1442										
ZZZZZZ	1446										
ZZZZZZ	1449										
ZZZZZZ	1452										
ZZZZZZ	1455										
ZZZZZZ	1458										
ZZZZZZ	1502										
ZZZZZZ	1505										
ZZZZZZ	1508										
ZZZZZZ	1511										
ZZZZZZ	1515										
ZZZZZZ	1518										
ZZZZZZ	1521										
ZZZZZZ	1525										
CCV019	1528	97		100		96		106		106	
CCB019	1535	98		102		102		104		105	
ZZZZZZ	1539										
ZZZZZZ	1542										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1545										
ZZZZZZ	1549										
ZZZZZZ	1552										
ZZZZZZ	1555										
ZZZZZZ	1559										
ZZZZZZ	1602										
ZZZZZZ	1605										
ZZZZZZ	1609										
ZZZZZZ	1612										
ZZZZZZ	1615										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1638										
ZZZZZZ	1641										
ZZZZZZ	1645										
ZZZZZZ	1648										
ZZZZZZ	1651										
CCV020	1655	98		101		97		106		108	
CCB020	1658	95		99		101		103		105	
ZZZZZZ	1701										
ZZZZZZ	1705										
ZZZZZZ	1708										
ZZZZZZ	1711										
ZZZZZZ	1714										
MYDAQ6ARE	1717	106		111		104		108		109	
MYDAT0	1720	117		124		101		110		110	
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1730										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
CCV021	1743	95		98		95		104		106	
CCB021	1745	95		99		102		105		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAQ6
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1749										
ZZZZZZ	1753										
ZZZZZZ	1756										
ZZZZZZ	1759										
ZZZZZZ	1802										
ZZZZZZ	1805										
ZZZZZZ	1809										
ZZZZZZ	1812										
ZZZZZZ	1815										
ZZZZZZ	1818										
ZZZZZZ	1821										
ZZZZZZ	1824										
ZZZZZZ	1828										
MYDAT0	1831	98		102		99		102		104	
CCV022	1834	95		97		96		107		108	
CCB022	1837	94		99		102		106		106	