

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51772 MA No.: \_\_\_\_\_ SDG No.: MYDAT3  
 Instrument ID: P5 Start Date: 10/23/2024  
 Analytical Method: ICP-AES Run Batch: LB133105  
 Concentration Units: ug/L

| Analyte   | True | Found  | %D | True | Found | %D  | True  | Found | %D  |
|-----------|------|--------|----|------|-------|-----|-------|-------|-----|
| Aluminum  | 0    | -0.010 | 0  | 200  | 210   | 5   | 38750 | 39000 | 1   |
| Antimony  | 0    | 0.020  | 0  | 60   | 55    | -8  | 1250  | 1100  | -12 |
| Arsenic   | 0    | -0     | 0  | 10   | 13    | 30  | 1250  | 1200  | -4  |
| Barium    | 0    | -0.020 | 0  | 200  | 210   | 5   | 2500  | 2500  | 0   |
| Beryllium | 0    | -0     | 0  | 5    | 5.7   | 14  | 125   | 130   | 4   |
| Cadmium   | 0    | -0     | 0  | 5    | 5.2   | 4   | 625   | 650   | 4   |
| Calcium   | 0    | -1.1   | 0  | 5000 | 5600  | 12  | 62500 | 66000 | 6   |
| Chromium  | 0    | -0     | 0  | 10   | 12    | 20  | 3875  | 4100  | 6   |
| Cobalt    | 0    | -0     | 0  | 50   | 50    | 0   | 625   | 630   | 1   |
| Copper    | 0    | -0     | 0  | 25   | 24    | -4  | 3750  | 3800  | 1   |
| Iron      | 0    | -0.040 | 0  | 100  | 120   | 20  | 37500 | 41000 | 9   |
| Lead      | 0    | -0     | 0  | 10   | 11    | 10  | 6250  | 6400  | 2   |
| Magnesium | 0    | -0.28  | 0  | 5000 | 5300  | 6   | 62500 | 62000 | -1  |
| Manganese | 0    | -0     | 0  | 15   | 15    | 0   | 3750  | 4000  | 7   |
| Nickel    | 0    | -0     | 0  | 40   | 43    | 8   | 625   | 640   | 2   |
| Potassium | 0    | 0.20   | 0  | 5000 | 5000  | 0   | 37500 | 37000 | -1  |
| Selenium  | 0    | 0.010  | 0  | 35   | 26    | -26 | 1250  | 1100  | -12 |
| Silver    | 0    | -0     | 0  | 10   | 12    | 20  | 312.5 | 310   | -1  |
| Sodium    | 0    | -0.29  | 0  | 5000 | 5200  | 4   | 37500 | 38000 | 1   |
| Thallium  | 0    | -0     | 0  | 25   | 26    | 4   | 1250  | 1300  | 4   |
| Vanadium  | 0    | -0     | 0  | 50   | 53    | 6   | 625   | 640   | 2   |
| Zinc      | 0    | 0.010  | 0  | 60   | 57    | -5  | 3750  | 3600  | -4  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51772 MA No.: \_\_\_\_\_ SDG No.: MYDAT3  
 Instrument ID: P5 Start Date: 10/23/2024  
 Analytical Method: ICP-AES Run Batch: LB133105  
 Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True   | Found  | %D | True   | Found  | %D |
|-----------|--------|--------|----|--------|--------|----|--------|--------|----|
| Aluminum  | 77500  | 77000  | -1 | 155000 | 160000 | 3  | 310000 | 310000 | 0  |
| Antimony  | 2500   | 2300   | -8 | 5000   | 4800   | -4 | 10000  | 11000  | 10 |
| Arsenic   | 2500   | 2400   | -4 | 5000   | 5000   | 0  | 10000  | 10000  | 0  |
| Barium    | 5000   | 5100   | 2  | 10000  | 10000  | 0  | 20000  | 20000  | 0  |
| Beryllium | 250    | 250    | 0  | 500    | 500    | 0  | 1000   | 990    | -1 |
| Cadmium   | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 4900   | -2 |
| Calcium   | 125000 | 130000 | 4  | 250000 | 260000 | 4  | 500000 | 510000 | 2  |
| Chromium  | 7750   | 8100   | 5  | 15500  | 16000  | 3  | 31000  | 30000  | -3 |
| Cobalt    | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 5000   | 0  |
| Copper    | 7500   | 7600   | 1  | 15000  | 15000  | 0  | 30000  | 30000  | 0  |
| Iron      | 75000  | 81000  | 8  | 150000 | 160000 | 7  | 300000 | 310000 | 3  |
| Lead      | 12500  | 13000  | 4  | 25000  | 25000  | 0  | 50000  | 50000  | 0  |
| Magnesium | 125000 | 120000 | -4 | 250000 | 250000 | 0  | 500000 | 500000 | 0  |
| Manganese | 7500   | 7700   | 3  | 15000  | 15000  | 0  | 30000  | 29000  | -3 |
| Nickel    | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 4900   | -2 |
| Potassium | 75000  | 74000  | -1 | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Selenium  | 2500   | 2300   | -8 | 5000   | 4800   | -4 | 10000  | 10000  | 0  |
| Silver    | 625    | 620    | -1 | 1250   | 1200   | -4 | 2500   | 2500   | 0  |
| Sodium    | 75000  | 75000  | 0  | 150000 | 150000 | 0  | 300000 | 300000 | 0  |
| Thallium  | 2500   | 2600   | 4  | 5000   | 5100   | 2  | 10000  | 9700   | -3 |
| Vanadium  | 1250   | 1300   | 4  | 2500   | 2500   | 0  | 5000   | 4900   | -2 |
| Zinc      | 7500   | 7100   | -5 | 15000  | 14000  | -7 | 30000  | 31000  | 3  |

FORM 15-IN  
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51772 MA No.:                      SDG No.: MYDAT3  
Instrument ID: P5 Start Date: 10/23/2024  
Analytical Method: ICP-AES Run Batch: LB133105  
Concentration Units: ug/L

| Analyte   | True   | Found  | %D | True | Found | %D | True | Found | %D |
|-----------|--------|--------|----|------|-------|----|------|-------|----|
| Aluminum  | 800000 | 800000 | 0  |      |       |    |      |       |    |
| Calcium   | 800000 | 780000 | -3 |      |       |    |      |       |    |
| Iron      | 800000 | 780000 | -3 |      |       |    |      |       |    |
| Magnesium | 800000 | 790000 | -1 |      |       |    |      |       |    |
| Sodium    | 800000 | 800000 | 0  |      |       |    |      |       |    |