

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAX3
 Instrument ID: P5 Start Date: 10/24/2024
 Analytical Method: ICP-AES Run Batch: LB133118
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.040	0	200	160	-20	38750	39000	1
Antimony	0	0.010	0	60	55	-8	1250	1100	-12
Arsenic	0	0.0	0	10	7.8	-22	1250	1200	-4
Barium	0	-0.020	0	200	210	5	2500	2600	4
Beryllium	0	-0	0	5	5.4	8	125	130	4
Cadmium	0	-0	0	5	5.1	2	625	650	4
Calcium	0	-1.0	0	5000	5500	10	62500	66000	6
Chromium	0	-0	0	10	11	10	3875	4100	6
Cobalt	0	-0	0	50	51	2	625	640	2
Copper	0	-0	0	25	25	0	3750	3900	4
Iron	0	-0.040	0	100	120	20	37500	41000	9
Lead	0	-0	0	10	13	30	6250	6400	2
Magnesium	0	-0.32	0	5000	5300	6	62500	62000	-1
Manganese	0	-0	0	15	19	27	3750	4000	7
Nickel	0	-0.010	0	40	44	10	625	650	4
Potassium	0	0.070	0	5000	5100	2	37500	37000	-1
Selenium	0	0.010	0	35	32	-9	1250	1200	-4
Silver	0	-0	0	10	10	0	312.5	310	-1
Sodium	0	0.10	0	5000	4800	-4	37500	38000	1
Thallium	0	-0	0	25	26	4	1250	1300	4
Vanadium	0	-0.010	0	50	59	18	625	640	2
Zinc	0	0.010	0	60	58	-3	3750	3600	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDAX3
 Instrument ID: P5 Start Date: 10/24/2024
 Analytical Method: ICP-AES Run Batch: LB133118
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	77000	-1	155000	150000	-3	310000	310000	0
Antimony	2500	2300	-8	5000	4800	-4	10000	11000	10
Arsenic	2500	2500	0	5000	5000	0	10000	10000	0
Barium	5000	5100	2	10000	10000	0	20000	20000	0
Beryllium	250	260	4	500	500	0	1000	980	-2
Cadmium	1250	1300	4	2500	2500	0	5000	4900	-2
Calcium	125000	130000	4	250000	260000	4	500000	500000	0
Chromium	7750	8100	5	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2500	0	5000	4900	-2
Copper	7500	7600	1	15000	15000	0	30000	30000	0
Iron	75000	82000	9	150000	160000	7	300000	310000	3
Lead	12500	13000	4	25000	25000	0	50000	49000	-2
Magnesium	125000	130000	4	250000	250000	0	500000	500000	0
Manganese	7500	7800	4	15000	15000	0	30000	29000	-3
Nickel	1250	1300	4	2500	2500	0	5000	4900	-2
Potassium	75000	74000	-1	150000	150000	0	300000	300000	0
Selenium	2500	2400	-4	5000	4800	-4	10000	10000	0
Silver	625	620	-1	1250	1200	-4	2500	2500	0
Sodium	75000	76000	1	150000	150000	0	300000	300000	0
Thallium	2500	2600	4	5000	5100	2	10000	9700	-3
Vanadium	1250	1300	4	2500	2500	0	5000	4900	-2
Zinc	7500	7100	-5	15000	15000	0	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDAX3
Instrument ID: P5 Start Date: 10/24/2024
Analytical Method: ICP-AES Run Batch: LB133118
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	800000	0						
Calcium	800000	780000	-3						
Iron	800000	780000	-3						
Magnesium	800000	790000	-1						
Sodium	800000	800000	0						