

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE002	TUNE	10/27/24 15:19		Jaswal	OK
2	S00	S0	CAL1	10/27/24 15:55		Jaswal	OK
3	S02	S02	CAL3	10/27/24 15:58		Jaswal	OK
4	S03	S03	CAL4	10/27/24 16:01		Jaswal	OK
5	S04	S04	CAL5	10/27/24 16:05		Jaswal	OK
6	S05	S05	CAL6	10/27/24 16:08		Jaswal	OK
7	S06	S06	CAL7	10/27/24 16:10		Jaswal	OK
8	S07	S07	CAL8	10/27/24 16:16		Jaswal	OK
9	S08	S08	CAL9	10/27/24 16:19		Jaswal	OK
10	ICV002	ICV002	ICV	10/27/24 16:29	Inst pause for calibration Review	Jaswal	OK
11	ICB002	ICB002	ICB	10/27/24 16:37	Inst pause for review	Jaswal	OK
12	ICSA002	ICSA002	ICSA	10/27/24 16:43	Inst pause for review	Jaswal	OK
13	ICSAB002	ICSAB002	ICSAB	10/27/24 16:47		Jaswal	OK
14	CCV007	CCV007	CCV	10/27/24 16:58	Inst pause for review	Jaswal	OK
15	CCB007	CCB007	CCB	10/27/24 17:07	Inst pause for review	Jaswal	OK
16	P4303-01DL	MYDAE7	SAM	10/27/24 17:11	4X for Ba,Co,Ni,V	Jaswal	Confirms
17	P4303-02DL	MYDAE7D	DUP	10/27/24 17:14	4X for Ba,Co,Ni,V	Jaswal	Confirms
18	P4303-01LDL	MYDAE7L	SD	10/27/24 17:18	4X for Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

19	P4303-03DL	MYDAE7S	MS	10/27/24 17:21	4X for Ba,Co,Ni,V	Jaswal	Confirms
20	P4303-01MSDL	MYDAE7SRE	MS	10/27/24 17:24	4X for Ba,Co,Ni,V	Jaswal	Confirms
21	P4304-08DL	MYDAP2	SAM	10/27/24 17:27	4X for Ba,Co,Ni,V	Jaswal	Confirms
22	P4304-09DL	MYDAP2D	DUP	10/27/24 17:31	4X for Ba,Co,Ni,V	Jaswal	Confirms
23	P4304-10DL	MYDAP2S	MS	10/27/24 17:37	4X for Ba,Co,Ni,V	Jaswal	Confirms
24	P4304-08MSDL	MYDAP2SRE	MS	10/27/24 17:41	4X for Ba,Co,Ni,V	Jaswal	Confirms
25	P4305-01DL	MYDAJ5	SAM	10/27/24 17:44	4X for Ba,Co,Ni,V	Jaswal	Confirms
26	P4305-02DL	MYDAJ5D	DUP	10/27/24 17:47	4X for Ba,Co,Ni,V	Jaswal	Confirms
27	P4304-08LDL	MYDAP2L	SD	10/27/24 17:50	4X for Ba,Co,Ni,V	Jaswal	Confirms
28	P4305-01LDL	MYDAJ5L	SD	10/27/24 17:54	4X for Ba,Co,Ni,V	Jaswal	Confirms
29	P4305-03DL	MYDAJ5S	MS	10/27/24 17:57	4X for Ba,Co,Ni,V	Jaswal	Confirms
30	P4305-01MSDL	MYDAJ5SRE	MS	10/27/24 18:00	4X for Ba,Co,Ni,V	Jaswal	Confirms
31	P4306-11DL	MYDAM9	SAM	10/27/24 18:03	4X for Ba,Co,Ni,V	Jaswal	Confirms
32	P4306-12DL	MYDAM9D	DUP	10/27/24 18:07	4X for Ba,Co,Ni,V	Jaswal	Confirms
33	P4306-11LDL	MYDAM9L	SD	10/27/24 18:10	4X for Ba,Co,Ni,V	Jaswal	Confirms
34	P4306-13DL	MYDAM9S	MS	10/27/24 18:13	4X for Ba,Co,Ni,V	Jaswal	Confirms
35	P4306-11MSDL	MYDAM9SRE	MS	10/27/24 18:17	4X for Ba,Co,Ni,V	Jaswal	Confirms
36	CCV008	CCV008	CCV	10/27/24 18:20		Jaswal	OK
37	CCB008	CCB008	CCB	10/27/24 18:23		Jaswal	OK
38	PB164187BL	PBS187	MB	10/27/24 18:26		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

39	PB164187BS	LCS187	LCS	10/27/24 18:29		Jaswal	OK
40	PB164187BSD	LCSD187	LCSD	10/27/24 18:33		Jaswal	OK
41	P4308-01	MYDAN1	SAM	10/27/24 18:36	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
42	P4308-02	MYDAN1D	DUP	10/27/24 18:39	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
43	P4308-01L	MYDAN1L	SD	10/27/24 18:42	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
44	P4308-03	MYDAN1S	MS	10/27/24 18:46	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
45	P4308-01MS	MYDAN1SRE	MS	10/27/24 18:49	Ba,Co,Ni,V high,Fail for PB(45ppb),Ag(BelowRL)	Jaswal	Dilution
46	P4308-04	MYDAP0	SAM	10/27/24 18:52		Jaswal	OK
47	P4308-05	MYDAP1	SAM	10/27/24 18:55		Jaswal	OK
48	P4308-06	MYDAP6	SAM	10/27/24 18:58		Jaswal	OK
49	P4308-07	MYDAP7	SAM	10/27/24 19:01		Jaswal	OK
50	P4308-08	MYDAP8	SAM	10/27/24 19:05		Jaswal	OK
51	P4308-09	MYDAP9	SAM	10/27/24 19:08		Jaswal	OK
52	P4308-10	MYDAQ0	SAM	10/27/24 19:11		Jaswal	OK
53	P4308-11	MYDAQ1	SAM	10/27/24 19:14		Jaswal	OK
54	P4308-12	MYDAQ2	SAM	10/27/24 19:17		Jaswal	OK
55	P4308-13	MYDAQ3	SAM	10/27/24 19:21		Jaswal	OK
56	P4308-14	MYDAQ4	SAM	10/27/24 19:24		Jaswal	OK
57	P4308-15	MYDAQ5	SAM	10/27/24 19:27		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

58	P4308-16	MYDAQ7	SAM	10/27/24 19:30		Jaswal	OK
59	P4308-17	MYDAQ8	SAM	10/27/24 19:33		Jaswal	OK
60	CCV009	CCV009	CCV	10/27/24 19:46		Jaswal	OK
61	CCB009	CCB009	CCB	10/27/24 19:49		Jaswal	OK
62	P4308-18	MYDAQ9	SAM	10/27/24 19:52		Jaswal	OK
63	P4308-19	MYDAR0	SAM	10/27/24 19:55		Jaswal	OK
64	P4308-20	MYDAR1	SAM	10/27/24 19:59		Jaswal	OK
65	P4308-21	MYDAR2	SAM	10/27/24 20:02		Jaswal	OK
66	P4308-22	MYDAR3	SAM	10/27/24 20:05		Jaswal	OK
67	P4308-01DL	MYDAN1	SAM	10/27/24 20:08	4X for Ba,Co,Ni,V	Jaswal	Confirms
68	P4308-02DL	MYDAN1D	DUP	10/27/24 20:12	4X for Ba,Co,Ni,V	Jaswal	Confirms
69	P4308-01LDL	MYDAN1L	SD	10/27/24 20:15	4X for Ba,Co,Ni,V	Jaswal	Confirms
70	P4308-03DL	MYDAN1S	MS	10/27/24 20:18	4X for Ba,Co,Ni,V	Jaswal	Confirms
71	P4308-01MSDL	MYDAN1SRE	MS	10/27/24 20:21	4X for Ba,Co,Ni,V	Jaswal	Confirms
72	PB164189BL	PBS189	MB	10/27/24 20:25		Jaswal	OK
73	PB164189BS	LCS189	LCS	10/27/24 20:28		Jaswal	OK
74	PB164189BSD	LCSD189	LCSD	10/27/24 20:31		Jaswal	OK
75	P4309-01	MYDAQ6	SAM	10/27/24 20:35	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
76	P4309-02	MYDAQ6D	DUP	10/27/24 20:39	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
77	P4309-01L	MYDAQ6L	SD	10/27/24 20:42	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

78	P4309-03	MYDAQ6S	MS	10/27/24 20:45	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
79	P4309-01MS	MYDAQ6SRE	MS	10/27/24 20:48	Ba,Co,Ni,V high,	Jaswal	Dilution
80	P4309-01DL	MYDAQ6	SAM	10/27/24 21:01	4X for Ba,Co,Ni,V	Jaswal	Confirms
81	P4309-02DL	MYDAQ6D	DUP	10/27/24 21:04	4X for Ba,Co,Ni,V	Jaswal	Confirms
82	P4309-01LDL	MYDAQ6L	SD	10/27/24 21:08	4X for Ba,Co,Ni,V	Jaswal	Confirms
83	P4309-03DL	MYDAQ6S	MS	10/27/24 21:11	4X for Ba,Co,Ni,V	Jaswal	Confirms
84	P4309-01MSDL	MYDAQ6SRE	MS	10/27/24 21:14	4X for Ba,Co,Ni,V	Jaswal	Confirms
85	P4309-04	MYDAR4	SAM	10/27/24 21:18		Jaswal	OK
86	CCV010	CCV010	CCV	10/27/24 21:21		Jaswal	OK
87	CCB010	CCB010	CCB	10/27/24 21:24		Jaswal	OK
88	P4309-05	MYDAR5	SAM	10/27/24 21:27		Jaswal	OK
89	P4309-06	MYDAR6	SAM	10/27/24 21:31	Int_Std 89Y(1) fail	Jaswal	Dilution
90	P4309-07	MYDAR7	SAM	10/27/24 21:34	Int_Std 89Y(1) fail	Jaswal	Dilution
91	P4309-05DL	MYDAR5	SAM	10/27/24 21:37	Not use	Jaswal	Not Ok
92	P4309-06DL	MYDAR6	SAM	10/27/24 21:40	2X for Int_Std 89Y(1)	Jaswal	Confirms
93	P4309-07DL	MYDAR7	SAM	10/27/24 21:43	2X for Int_Std 89Y(1)	Jaswal	Confirms
94	P4309-08	MYDAR8	SAM	10/27/24 21:47	Int_Std 89Y(1) fail	Jaswal	Dilution
95	P4309-09	MYDAR9	SAM	10/27/24 21:50		Jaswal	OK
96	P4309-10	MYDAS0	SAM	10/27/24 21:53		Jaswal	OK
97	P4309-11	MYDAS1	SAM	10/27/24 21:56		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

98	P4309-12	MYDAS2	SAM	10/27/24 22:00		Jaswal	OK
99	P4309-13	MYDAS3	SAM	10/27/24 22:03		Jaswal	OK
100	P4309-14	MYDAS4	SAM	10/27/24 22:06	Int_Std 89Y(1) fail	Jaswal	Dilution
101	P4309-15	MYDAS5	SAM	10/27/24 22:09		Jaswal	OK
102	P4309-16	MYDAS6	SAM	10/27/24 22:12		Jaswal	OK
103	P4309-17	MYDAS7	SAM	10/27/24 22:16		Jaswal	OK
104	P4309-18	MYDAS8	SAM	10/27/24 22:19	Int_Std 89Y(1),(2) fail	Jaswal	Dilution
105	P4309-19	MYDAS9	SAM	10/27/24 22:22	Int_Std 89Y(1) fail	Jaswal	Dilution
106	P4309-20	MYDAT0	SAM	10/27/24 22:25	Int_Std 89Y(1) fail	Jaswal	Dilution
107	P4309-21	MYDAT1	SAM	10/27/24 22:28		Jaswal	OK
108	P4309-22	MYDAT2	SAM	10/27/24 22:32		Jaswal	OK
109	P4309-08DL	MYDAR8	SAM	10/27/24 22:35	2X for Int_Std 89Y(1)	Jaswal	Confirms
110	P4309-14DL	MYDAS4	SAM	10/27/24 22:38	2X for Int_Std 89Y(1)	Jaswal	Confirms
111	P4309-16DL	MYDAS6	SAM	10/27/24 22:41	Not use	Jaswal	Not Ok
112	P4309-17DL	MYDAS7	SAM	10/27/24 22:47	Not use	Jaswal	Not Ok
113	P4309-18DL	MYDAS8	SAM	10/27/24 22:50	2X for Int_Std 89Y(1)(2)	Jaswal	Confirms
114	CCV011	CCV011	CCV	10/27/24 22:57		Jaswal	OK
115	CCB011	CCB011	CCB	10/27/24 23:07	Inst pause for review	Jaswal	OK
116	P4309-19DL	MYDAS9	SAM	10/27/24 23:17	2X for Int_Std 89Y(1)	Jaswal	Confirms
117	PB164201BL	PBS201	MB	10/27/24 23:22		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

118	PB164201BSD	LCSD201	LCSD	10/27/24 23:25		Jaswal	OK
119	PB164201BS	LCS201	LCS	10/27/24 23:33		Jaswal	OK
120	P4309-21DL	MYDAT1	SAM	10/27/24 23:37	Not use	Jaswal	Not Ok
121	P4310-01	MYDAT3	SAM	10/27/24 23:43		Jaswal	OK
122	P4310-02	MYDAT4	SAM	10/27/24 23:46		Jaswal	OK
123	P4310-03	MYDAT5	SAM	10/27/24 23:50	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
124	P4310-04	MYDAT5D	DUP	10/27/24 23:53	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
125	P4310-05	MYDAT5S	MS	10/27/24 23:59	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
126	P4310-03MS	MYDAT5SRE	MS	10/28/24 00:04	Ba,Co,Ni,V high,Fail for PB(48ppb)	Jaswal	Dilution
127	P4310-03L	MYDAT5L	SD	10/28/24 00:08	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
128	P4310-03DL	MYDAT5	SAM	10/28/24 00:11	4X for Ba,Co,Ni,V	Jaswal	Confirms
129	P4310-04DL	MYDAT5D	DUP	10/28/24 00:14	4X for Ba,Co,Ni,V	Jaswal	Confirms
130	P4310-03LDL	MYDAT5L	SD	10/28/24 00:17	4X for Ba,Co,Ni,V	Jaswal	Confirms
131	P4310-05DL	MYDAT5S	MS	10/28/24 00:21	4X for Ba,Co,Ni,V	Jaswal	Confirms
132	P4310-03MSDL	MYDAT5SRE	MS	10/28/24 00:24	4X for Ba,Co,Ni,V	Jaswal	Confirms
133	P4310-06	MYDAT6	SAM	10/28/24 00:27		Jaswal	OK
134	P4310-07	MYDAT7	SAM	10/28/24 00:30		Jaswal	OK
135	P4310-08	MYDAT8	SAM	10/28/24 00:34		Jaswal	OK
136	P4310-09	MYDAT9	SAM	10/28/24 00:37		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

137	P4310-10	MYDAW0	SAM	10/28/24 00:40		Jaswal	OK
138	P4310-11	MYDAW1	SAM	10/28/24 00:43		Jaswal	OK
139	CCV012	CCV012	CCV	10/28/24 00:47		Jaswal	OK
140	CCB012	CCB012	CCB	10/28/24 00:50		Jaswal	OK
141	P4310-12	MYDAW2	SAM	10/28/24 00:53		Jaswal	OK
142	P4310-13	MYDAW3	SAM	10/28/24 00:56		Jaswal	OK
143	P4310-14	MYDAW4	SAM	10/28/24 00:59		Jaswal	OK
144	P4310-15	MYDAW5	SAM	10/28/24 01:03		Jaswal	OK
145	P4310-16	MYDAW6	SAM	10/28/24 01:06		Jaswal	OK
146	P4310-17	MYDAW7	SAM	10/28/24 01:09		Jaswal	OK
147	P4310-18	MYDAW8	SAM	10/28/24 01:12		Jaswal	OK
148	P4310-19	MYDAW9	SAM	10/28/24 01:15		Jaswal	OK
149	P4310-20	MYDAX0	SAM	10/28/24 01:19		Jaswal	OK
150	P4310-21	MYDAX1	SAM	10/28/24 01:22		Jaswal	OK
151	P4310-22	MYDAX2	SAM	10/28/24 01:25		Jaswal	OK
152	PB164224BL	PBS224	MB	10/28/24 01:28		Jaswal	OK
153	PB164224BS	LCS224	LCS	10/28/24 01:32		Jaswal	OK
154	PB164224BSD	LCSD224	LCSD	10/28/24 01:35		Jaswal	OK
155	ICSA003	ICSA003	ICSA	10/28/24 01:38		Jaswal	OK
156	ICSAB003	ICSAB003	ICSAB	10/28/24 01:41		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

157	CCV013	CCV013	CCV	10/28/24 01:45		Jaswal	OK
158	CCB013	CCB013	CCB	10/28/24 01:47		Jaswal	OK
159	P4311-01	MYDAX3	SAM	10/28/24 01:51		Jaswal	OK
160	P4311-20	MYDB06	SAM	10/28/24 01:54	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
161	P4311-20L	MYDB06L	SD	10/28/24 02:00	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
162	P4311-22	MYDB06S	MS	10/28/24 02:04	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
163	P4311-21	MYDB06D	DUP	10/28/24 02:10	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
164	P4311-20MS	MYDB06SRE	MS	10/28/24 02:13	Ba,Co,Ni,V high,ail for Ag(BelowRL)	Jaswal	Dilution
165	P4311-20DL	MYDB06	SAM	10/28/24 02:16	4X for Ba,Co,Ni,V	Jaswal	Confirms
166	P4311-21DL	MYDB06D	DUP	10/28/24 02:20	4X for Ba,Co,Ni,V	Jaswal	Confirms
167	P4311-20LDL	MYDB06L	SD	10/28/24 02:23	4X for Ba,Co,Ni,V	Jaswal	Confirms
168	P4311-22DL	MYDB06S	MS	10/28/24 02:26	4X for Ba,Co,Ni,V	Jaswal	Confirms
169	P4311-02	MYDAX4	SAM	10/28/24 02:33		Jaswal	OK
170	P4311-03	MYDAX5	SAM	10/28/24 02:36		Jaswal	OK
171	P4311-20MSDL	MYDB06SRE	MS	10/28/24 02:39	4X for Ba,Co,Ni,V	Jaswal	Confirms
172	P4311-04	MYDAX6	SAM	10/28/24 02:42		Jaswal	OK
173	P4311-05	MYDAX7	SAM	10/28/24 02:46	Int_Std 89Y(1) fail	Jaswal	Dilution
174	P4311-06	MYDAX8	SAM	10/28/24 02:49		Jaswal	OK
175	P4311-07	MYDAX9	SAM	10/28/24 02:52		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

176	P4311-08	MYDAY0	SAM	10/28/24 02:55		Jaswal	OK
177	P4311-09	MYDAY1	SAM	10/28/24 02:59		Jaswal	OK
178	P4311-10	MYDAY2	SAM	10/28/24 03:02		Jaswal	OK
179	P4311-11	MYDAY3	SAM	10/28/24 03:05	Int_Std 89Y(1) fail	Jaswal	Dilution
180	P4311-12	MYDAY4	SAM	10/28/24 03:08		Jaswal	OK
181	P4311-13	MYDAY5	SAM	10/28/24 03:11		Jaswal	OK
182	CCV014	CCV014	CCV	10/28/24 03:15		Jaswal	OK
183	CCB014	CCB014	CCB	10/28/24 03:17		Jaswal	OK
184	P4311-14	MYDAY6	SAM	10/28/24 03:21	Int_Std 89Y(1) fail	Jaswal	Dilution
185	P4311-15	MYDAY7	SAM	10/28/24 03:24		Jaswal	OK
186	P4311-16	MYDAY8	SAM	10/28/24 03:27		Jaswal	OK
187	P4311-17	MYDAY9	SAM	10/28/24 03:30		Jaswal	OK
188	P4311-18	MYDAZ0	SAM	10/28/24 03:33		Jaswal	OK
189	P4311-19	MYDAZ1	SAM	10/28/24 03:37		Jaswal	OK
190	PB164253BL	PBS253	MB	10/28/24 03:40		Jaswal	OK
191	PB164253BS	LCS253	LCS	10/28/24 03:43		Jaswal	OK
192	PB164253BSD	LCSD253	LCSD	10/28/24 03:46		Jaswal	OK
193	P4312-01	MYDAZ2	SAM	10/28/24 03:50		Jaswal	OK
194	P4312-02	MYDAZ3	SAM	10/28/24 03:53		Jaswal	OK
195	P4312-03	MYDAZ4	SAM	10/28/24 03:56		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

196	P4312-04	MYDAZ5	SAM	10/28/24 03:59		Jaswal	OK
197	P4312-05	MYDAZ6	SAM	10/28/24 04:03		Jaswal	OK
198	P4312-06	MYDAZ7	SAM	10/28/24 04:06		Jaswal	OK
199	P4312-07	MYDAZ8	SAM	10/28/24 04:09		Jaswal	OK
200	P4312-08	MYDAZ9	SAM	10/28/24 04:12		Jaswal	OK
201	P4312-09	MYDB00	SAM	10/28/24 04:15	Int_Std 89Y(1) fail	Jaswal	Dilution
202	P4312-10	MYDB01	SAM	10/28/24 04:19		Jaswal	OK
203	P4312-11	MYDB02	SAM	10/28/24 04:22		Jaswal	OK
204	P4312-12	MYDB03	SAM	10/28/24 04:25		Jaswal	OK
205	P4312-13	MYDB04	SAM	10/28/24 04:28		Jaswal	OK
206	P4312-14	MYDB05	SAM	10/28/24 04:31		Jaswal	OK
207	P4312-15	MYDB07	SAM	10/28/24 04:35		Jaswal	OK
208	P4312-16	MYDB08	SAM	10/28/24 04:38		Jaswal	OK
209	P4312-17	MYDB09	SAM	10/28/24 04:41		Jaswal	OK
210	CCV015	CCV015	CCV	10/28/24 04:44		Jaswal	OK
211	CCB015	CCB015	CCB	10/28/24 04:47		Jaswal	OK
212	P4312-18	MYDB10	SAM	10/28/24 04:54		Jaswal	OK
213	P4312-19	MYDB11	SAM	10/28/24 04:57		Jaswal	OK
214	P4312-20	MYDB17	SAM	10/28/24 05:00	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
215	P4312-21	MYDB17D	DUP	10/28/24 05:03	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

216	P4312-20L	MYDB17L	SD	10/28/24 05:06	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
217	P4312-22	MYDB17S	MS	10/28/24 05:10	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
218	P4312-20MS	MYDB17SRE	MS	10/28/24 05:13	Ba,Co,Ni,V high,Fail for PB(103ppb),Ag(BelowRL)	Jaswal	Dilution
219	CCV016	CCV016	CCV	10/28/24 05:16		Jaswal	OK
220	CCB016	CCB016	CCB	10/28/24 05:19		Jaswal	OK
221	PB164260BL	PBS260	MB	10/28/24 05:25		Jaswal	OK
222	PB164260BS	LCS260	LCS	10/28/24 05:29		Jaswal	OK
223	PB164260BSD	LCSD260	LCSD	10/28/24 05:32		Jaswal	OK
224	P4313-01RE	MYDB23	SAM	10/28/24 05:35	Not use	Jaswal	Not Ok
225	P4313-02RE	MYDB24	SAM	10/28/24 05:39	Not use	Jaswal	Not Ok
226	P4313-03RE	MYDB25	SAM	10/28/24 05:42	Not use	Jaswal	Not Ok
227	P4313-04RE	MYDB26	SAM	10/28/24 05:45	Not use	Jaswal	Not Ok
228	P4313-05RE	MYDB26D	DUP	10/28/24 05:48	Not use	Jaswal	Not Ok
229	P4313-04LRE	MYDB26L	SD	10/28/24 05:52	Not use	Jaswal	Not Ok
230	P4313-06RE	MYDB26S	MS	10/28/24 05:55	Not use	Jaswal	Not Ok
231	P4313-04MSRE	MYDB26SRE	MS	10/28/24 05:58	Not use	Jaswal	Not Ok
232	P4313-07RE	MYDB27	SAM	10/28/24 06:02	Not use	Jaswal	Not Ok
233	P4313-08RE	MYDB28	SAM	10/28/24 06:05	Not use	Jaswal	Not Ok
234	P4313-01	MYDB23	SAM	10/28/24 06:18		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133170

Review By	Sarabjit Jaswal	Review On	10/28/2024 4:53:22 PM
Supervise By	Mohan Bera	Supervise On	10/28/2024 4:54:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

235	P4313-02	MYDB24	SAM	10/28/24 06:21		Jaswal	OK
236	P4313-03	MYDB25	SAM	10/28/24 06:24		Jaswal	OK
237	P4313-04	MYDB26	SAM	10/28/24 06:27	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
238	P4313-05	MYDB26D	DUP	10/28/24 06:30	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
239	P4313-04L	MYDB26L	SD	10/28/24 06:34	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
240	P4313-06	MYDB26S	MS	10/28/24 06:37	Ba,Co,Ni,V high in MSRE	Jaswal	Dilution
241	P4313-04MS	MYDB26SRE	MS	10/28/24 06:40	Ba,Co,Ni,V high	Jaswal	Dilution
242	P4313-07	MYDB27	SAM	10/28/24 06:43		Jaswal	OK
243	P4313-08	MYDB28	SAM	10/28/24 06:46		Jaswal	OK
244	P4313-09	MYDB29	SAM	10/28/24 06:50		Jaswal	OK
245	CCV017	CCV017	CCV	10/28/24 06:53		Jaswal	OK
246	CCB017	CCB017	CCB	10/28/24 06:59	Inst pause for review	Jaswal	OK