

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDB23
 Initial Calibration Verification Source : MP82578
 Continuing Calibration Verification Source : MP82579
 Run Batch: LB133170 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV007				ID: CCV008		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	200	100	1	500	530	107	2	520	104	2
Arsenic	200	200	102	1	500	490	99	1	490	98	2
Barium	100	95	95	1	2500	2400	98	1	2400	97	2
Beryllium	100	110	105	1	500	530	106	1	520	104	1
Cadmium	100	110	107	2	500	500	100	1	490	99	2
Chromium	100	100	102	0	500	480	95	1	480	96	0
Cobalt	100	100	100	0	500	490	98	1	490	98	1
Copper	100	95	95	0	5000	4800	96	2	4800	95	0
Lead	200	190	96	5	2500	2400	97	0	2400	96	1
Nickel	110	100	94	0	500	480	96	1	480	97	0
Selenium	200	210	106	2	500	490	99	1	490	98	2
Silver	50.0	51	102	2	500	500	101	1	500	99	1
Thallium	210	210	101	1	500	480	97	1	490	98	1
Vanadium	100	98	98	0	500	490	97	1	480	97	1
Zinc	200	200	98	1	5000	4900	98	2	4900	98	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDB23Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB133170 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV009				ID: CCV010		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	520	104	3	490	98	3
Arsenic					500	490	98	0	480	97	0
Barium					2500	2400	97	3	2400	98	2
Beryllium					500	530	106	2	530	106	1
Cadmium					500	530	106	1	500	100	2
Chromium					500	480	97	1	480	96	1
Cobalt					500	530	105	1	490	98	0
Copper					5000	5200	103	0	4800	97	1
Lead					2500	2400	98	2	2400	96	2
Nickel					500	520	104	1	480	97	1
Selenium					500	490	99	2	500	99	2
Silver					500	540	108	1	510	102	2
Thallium					500	490	98	1	490	97	3
Vanadium					500	490	98	1	480	96	1
Zinc					5000	5300	107	1	4900	98	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDB23Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB133170 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV011				ID: CCV012		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	490	98	2	480	97	1
Arsenic					500	490	97	0	480	95	1
Barium					2500	2400	97	1	2400	95	1
Beryllium					500	550	109	1	540	109	0
Cadmium					500	500	99	2	480	97	2
Chromium					500	490	97	2	470	95	1
Cobalt					500	500	100	1	500	99	2
Copper					5000	4900	98	1	4800	96	1
Lead					2500	2400	97	1	2400	96	1
Nickel					500	490	98	0	480	97	1
Selenium					500	500	100	1	480	96	1
Silver					500	510	101	1	490	98	1
Thallium					500	490	98	2	490	98	1
Vanadium					500	490	98	1	480	95	1
Zinc					5000	5000	99	0	4800	96	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDB23
 Initial Calibration Verification Source : MP82578
 Continuing Calibration Verification Source : MP82579
 Run Batch: LB133170 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV013				ID: CCV014		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	490	99	1	480	96	1
Arsenic					500	520	104	3	500	100	1
Barium					2500	2500	98	0	2400	96	1
Beryllium					500	550	110	1	550	109	1
Cadmium					500	500	101	1	490	98	2
Chromium					500	490	97	1	490	97	1
Cobalt					500	520	104	1	510	102	0
Copper					5000	5000	100	1	5000	99	0
Lead					2500	2500	98	1	2400	97	2
Nickel					500	510	101	1	500	99	0
Selenium					500	510	101	1	490	98	0
Silver					500	510	101	1	500	100	1
Thallium					500	500	99	2	490	97	2
Vanadium					500	500	99	0	490	98	1
Zinc					5000	5000	100	2	5000	100	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDB23Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB133170 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV015				ID: CCV016		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	480	96	1	480	96	2
Arsenic					500	500	100	1	490	98	1
Barium					2500	2400	95	2	2400	96	1
Beryllium					500	530	105	1	540	107	1
Cadmium					500	490	98	1	490	98	0
Chromium					500	490	98	2	480	97	1
Cobalt					500	500	101	2	510	101	1
Copper					5000	5000	99	1	5000	100	1
Lead					2500	2400	96	1	2400	96	1
Nickel					500	490	98	1	490	98	1
Selenium					500	490	99	1	490	98	1
Silver					500	500	100	2	510	101	0
Thallium					500	480	96	0	480	97	1
Vanadium					500	490	98	1	490	98	2
Zinc					5000	5100	101	1	5100	102	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDB23Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB133170 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV017				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	480	95	3			
Arsenic					500	490	98	0			
Barium					2500	2400	96	4			
Beryllium					500	520	104	1			
Cadmium					500	490	97	3			
Chromium					500	490	98	3			
Cobalt					500	520	104	1			
Copper					5000	5100	102	1			
Lead					2500	2400	95	2			
Nickel					500	500	101	2			
Selenium					500	480	96	1			
Silver					500	500	100	2			
Thallium					500	480	96	1			
Vanadium					500	490	98	2			
Zinc					5000	5100	101	1			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYDB23
 Initial Calibration Verification Source : MP82578
 Continuing Calibration Verification Source : MP82579
 Run Batch: LB133179 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV003				ID: CCV018				ID: CCV019		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	190	97	1	500	460	92	1	470	93	1
Arsenic	200	200	101	2	500	490	99	0	510	101	0
Barium	100	94	94	1	2500	2300	94	1	2400	96	0
Beryllium	100	110	108	1	500	490	99	1	510	103	1
Cadmium	100	100	104	2	500	490	98	2	490	98	1
Chromium	100	100	101	1	500	470	93	1	480	96	1
Cobalt	100	100	101	1	500	500	101	0	510	103	1
Copper	100	94	94	1	5000	4800	97	1	4900	98	1
Lead	200	180	92	1	2500	2300	94	1	2400	96	2
Nickel	110	100	94	0	500	490	98	1	500	100	1
Selenium	200	210	106	2	500	500	100	0	500	100	1
Silver	50.0	48	97	2	500	480	97	2	490	98	1
Thallium	210	210	102	1	500	470	94	3	480	95	1
Vanadium	100	97	97	1	500	460	92	0	480	97	0
Zinc	200	190	97	1	5000	4800	97	1	5000	100	1