

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB92
 Instrument ID: P5 Start Date: 10/23/2024
 Analytical Method: ICP-AES Run Batch: LB133105
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	-0.010	0	200	210	5	38750	39000	1
Antimony	0	0.020	0	60	55	-8	1250	1100	-12
Arsenic	0	-0	0	10	13	30	1250	1200	-4
Barium	0	-0.020	0	200	210	5	2500	2500	0
Beryllium	0	-0	0	5	5.7	14	125	130	4
Cadmium	0	-0	0	5	5.2	4	625	650	4
Calcium	0	-1.1	0	5000	5600	12	62500	66000	6
Chromium	0	-0	0	10	12	20	3875	4100	6
Cobalt	0	-0	0	50	50	0	625	630	1
Copper	0	-0	0	25	24	-4	3750	3800	1
Iron	0	-0.040	0	100	120	20	37500	41000	9
Lead	0	-0	0	10	11	10	6250	6400	2
Magnesium	0	-0.28	0	5000	5300	6	62500	62000	-1
Manganese	0	-0	0	15	15	0	3750	4000	7
Nickel	0	-0	0	40	43	8	625	640	2
Potassium	0	0.20	0	5000	5000	0	37500	37000	-1
Selenium	0	0.010	0	35	26	-26	1250	1100	-12
Silver	0	-0	0	10	12	20	312.5	310	-1
Sodium	0	-0.29	0	5000	5200	4	37500	38000	1
Thallium	0	-0	0	25	26	4	1250	1300	4
Vanadium	0	-0	0	50	53	6	625	640	2
Zinc	0	0.010	0	60	57	-5	3750	3600	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB92
 Instrument ID: P5 Start Date: 10/23/2024
 Analytical Method: ICP-AES Run Batch: LB133105
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	77000	-1	155000	160000	3	310000	310000	0
Antimony	2500	2300	-8	5000	4800	-4	10000	11000	10
Arsenic	2500	2400	-4	5000	5000	0	10000	10000	0
Barium	5000	5100	2	10000	10000	0	20000	20000	0
Beryllium	250	250	0	500	500	0	1000	990	-1
Cadmium	1250	1300	4	2500	2500	0	5000	4900	-2
Calcium	125000	130000	4	250000	260000	4	500000	510000	2
Chromium	7750	8100	5	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2500	0	5000	5000	0
Copper	7500	7600	1	15000	15000	0	30000	30000	0
Iron	75000	81000	8	150000	160000	7	300000	310000	3
Lead	12500	13000	4	25000	25000	0	50000	50000	0
Magnesium	125000	120000	-4	250000	250000	0	500000	500000	0
Manganese	7500	7700	3	15000	15000	0	30000	29000	-3
Nickel	1250	1300	4	2500	2500	0	5000	4900	-2
Potassium	75000	74000	-1	150000	150000	0	300000	300000	0
Selenium	2500	2300	-8	5000	4800	-4	10000	10000	0
Silver	625	620	-1	1250	1200	-4	2500	2500	0
Sodium	75000	75000	0	150000	150000	0	300000	300000	0
Thallium	2500	2600	4	5000	5100	2	10000	9700	-3
Vanadium	1250	1300	4	2500	2500	0	5000	4900	-2
Zinc	7500	7100	-5	15000	14000	-7	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
Instrument ID: P5 Start Date: 10/23/2024
Analytical Method: ICP-AES Run Batch: LB133105
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	800000	0						
Calcium	800000	780000	-3						
Iron	800000	780000	-3						
Magnesium	800000	790000	-1						
Sodium	800000	800000	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB92
 Instrument ID: P5 Start Date: 01/08/2025
 Analytical Method: ICP-AES Run Batch: LB134200
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	-0.040	0	200	260	30	38750	38000	-2
Antimony	0	0.010	0	60	57	-5	1250	1200	-4
Arsenic	0	-0	0	10	12	20	1250	1300	4
Barium	0	-0.010	0	200	210	5	2500	2500	0
Beryllium	0	-0	0	5	4.8	-4	125	130	4
Cadmium	0	0.0	0	5	4.5	-10	625	600	-4
Calcium	0	-0.070	0	5000	5100	2	62500	63000	1
Chromium	0	-0	0	10	11	10	3875	3900	1
Cobalt	0	0.0	0	50	49	-2	625	620	-1
Copper	0	-0.010	0	25	30	20	3750	4000	7
Iron	0	0.0	0	100	94	-6	37500	38000	1
Lead	0	0.0	0	10	7.4	-26	6250	6000	-4
Magnesium	0	-0.76	0	5000	5400	8	62500	65000	4
Manganese	0	-0	0	15	19	27	3750	3800	1
Nickel	0	0.0	0	40	37	-8	625	610	-2
Potassium	0	0.60	0	5000	4500	-10	37500	37000	-1
Selenium	0	-0	0	35	40	14	1250	1200	-4
Silver	0	0.0	0	10	9.0	-10	312.5	310	-1
Sodium	0	-1.9	0	5000	5800	16	37500	42000	12
Thallium	0	-0	0	25	28	12	1250	1300	4
Vanadium	0	-0.010	0	50	54	8	625	650	4
Zinc	0	0.0	0	60	61	2	3750	3600	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB92
 Instrument ID: P5 Start Date: 01/08/2025
 Analytical Method: ICP-AES Run Batch: LB134200
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	76000	-2	155000	150000	-3	310000	300000	-3
Antimony	2500	2400	-4	5000	4900	-2	10000	10000	0
Arsenic	2500	2600	4	5000	5000	0	10000	9900	-1
Barium	5000	5000	0	10000	9900	-1	20000	20000	0
Beryllium	250	260	4	500	500	0	1000	990	-1
Cadmium	1250	1200	-4	2500	2500	0	5000	5100	2
Calcium	125000	130000	4	250000	240000	-4	500000	490000	-2
Chromium	7750	7900	2	15500	15000	-3	31000	31000	0
Cobalt	1250	1300	4	2500	2500	0	5000	5000	0
Copper	7500	7900	5	15000	15000	0	30000	29000	-3
Iron	75000	76000	1	150000	150000	0	300000	300000	0
Lead	12500	12000	-4	25000	25000	0	50000	51000	2
Magnesium	125000	130000	4	250000	250000	0	500000	500000	0
Manganese	7500	7700	3	15000	15000	0	30000	30000	0
Nickel	1250	1200	-4	2500	2500	0	5000	5100	2
Potassium	75000	75000	0	150000	150000	0	300000	300000	0
Selenium	2500	2400	-4	5000	4900	-2	10000	10000	0
Silver	625	630	1	1250	1200	-4	2500	2500	0
Sodium	75000	83000	11	150000	160000	7	300000	290000	-3
Thallium	2500	2500	0	5000	5000	0	10000	9900	-1
Vanadium	1250	1300	4	2500	2500	0	5000	4900	-2
Zinc	7500	7400	-1	15000	15000	0	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
Instrument ID: P5 Start Date: 01/08/2025
Analytical Method: ICP-AES Run Batch: LB134200
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	820000	3						
Calcium	800000	820000	3						
Iron	800000	800000	0						
Magnesium	800000	790000	-1						
Sodium	800000	790000	-1						