

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1322	100		100		100		100		100	
S02	1326	98		99		99		99		99	
S03	1329	96		97		97		98		98	
S04	1332	102		99		100		99		101	
S05	1335	109		100		101		99		103	
S06	1338	113		100		103		100		104	
S07	1341	116		99		103		99		105	
S08	1344	112		99		99		91		103	
ICV003	1403	104		100		100		99		101	
ICB003	1406	101		98		98		97		98	
ICSA004	1410	107		99		102		98		105	
ICSAB004	1413	108		104		105		100		107	
CCV018	1416	106		102		105		97		107	
CCB018	1419	109		102		104		103		107	
ZZZZZZ	1422										
ZZZZZZ	1425										
ZZZZZZ	1429										
ZZZZZZ	1432										
ZZZZZZ	1435										
ZZZZZZ	1438										
ZZZZZZ	1442										
ZZZZZZ	1446										
ZZZZZZ	1449										
ZZZZZZ	1452										
ZZZZZZ	1455										
ZZZZZZ	1458										
ZZZZZZ	1502										
ZZZZZZ	1505										
ZZZZZZ	1508										
ZZZZZZ	1511										
ZZZZZZ	1515										
ZZZZZZ	1518										
ZZZZZZ	1521										
ZZZZZZ	1525										
CCV019	1528	98		98		102		95		105	
CCB019	1535	97		101		102		102		104	
ZZZZZZ	1539										
ZZZZZZ	1542										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Lab Name:	Alliance Technical Group, LLC	Contract:	68HERH20D0011
Lab Code:	ACE	Case No.:	51772
		MA No.:	SDG No.: MYDB92
Instrument ID:	P7	Start Date :	10/28/2024
Run Batch:	LB133179	End Date :	10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1545										
ZZZZZZ	1549										
ZZZZZZ	1552										
ZZZZZZ	1555										
PBS273	1559	96		99		102		101		103	
LCS273	1602	95		99		100		101		103	
LCSD273	1605	94		99		100		100		101	
ZZZZZZ	1609										
ZZZZZZ	1612										
ZZZZZZ	1615										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1638										
ZZZZZZ	1641										
ZZZZZZ	1645										
ZZZZZZ	1648										
ZZZZZZ	1651										
CCV020	1655	99		99		104		96		106	
CCB020	1658	101		98		101		102		106	
ZZZZZZ	1701										
ZZZZZZ	1705										
ZZZZZZ	1708										
ZZZZZZ	1711										
ZZZZZZ	1714										
ZZZZZZ	1717										
ZZZZZZ	1720										
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1730										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
CCV021	1743	92		97		103		94		104	
CCB021	1745	96		99		103		103		106	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1749										
ZZZZZZ	1753										
ZZZZZZ	1756										
ZZZZZZ	1759										
ZZZZZZ	1802										
ZZZZZZ	1805										
ZZZZZZ	1809										
ZZZZZZ	1812										
ZZZZZZ	1815										
ZZZZZZ	1818										
ZZZZZZ	1821										
ZZZZZZ	1824										
ZZZZZZ	1828										
ZZZZZZ	1831										
CCV022	1834	99		98		102		96		106	
CCB022	1837	105		102		105		106		110	
ZZZZZZ	1841										
ZZZZZZ	1844										
ZZZZZZ	1847										
ZZZZZZ	1851										
ZZZZZZ	1854										
ZZZZZZ	1857										
ZZZZZZ	1900										
ZZZZZZ	1903										
ZZZZZZ	1907										
ZZZZZZ	1910										
ZZZZZZ	1913										
ZZZZZZ	1916										
ZZZZZZ	1919										
ZZZZZZ	1923										
ZZZZZZ	1926										
ZZZZZZ	1929										
ZZZZZZ	1932										
ZZZZZZ	1935										
CCV023	1939	99		99		104		98		108	
CCB023	1941	101		98		103		102		107	
ZZZZZZ	1945										
ZZZZZZ	1948										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1951										
ZZZZZZ	1955										
ZZZZZZ	1958										
ZZZZZZ	2001										
ZZZZZZ	2004										
ZZZZZZ	2008										
ZZZZZZ	2011										
ZZZZZZ	2014										
ZZZZZZ	2017										
ZZZZZZ	2021										
ZZZZZZ	2024										
ZZZZZZ	2027										
ZZZZZZ	2030										
ZZZZZZ	2033										
MYDB92	2036	84		98		105		96		102	
MYDB93	2040	82		96		102		95		101	
MYDB94	2043	77		96		100		95		100	
MYDB95	2046	77		95		103		94		99	
MYDB96	2049	79		97		103		96		102	
MYDB97	2052	81		97		103		96		101	
MYDB98	2056	81		99		103		97		101	
MYDB99	2059	84		99		102		97		100	
MYDBA0	2102	83		97		102		95		100	
MYDBA0D	2105	81		95		102		95		100	
CCV024	2108	96		97		104		98		107	
CCB024	2111	100		98		102		102		108	
MYDBA0L	2115	95		94		100		94		100	
MYDBA0S	2118	91		98		104		96		104	
MYDBA0SRE	2121	87		96		102		96		100	
MYDBA1	2124	81		93		101		91		97	
MYDBA2	2127	80		95		103		95		100	
MYDBA3	2130	81		102		110		98		103	
MYDBA4	2134	82		103		111		99		104	
MYDBA5	2137	80		100		107		95		100	
ICSA005	2141	89		99		102		98		106	
ICSAB005	2145	88		97		102		98		107	
CCV025	2151	94		97		103		97		108	
CCB025	2154	92		97		103		104		108	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1322	100		100		100		100		100	
S02	1326	100		100		99		101		102	
S03	1329	99		99		99		101		101	
S04	1332	100		100		99		102		105	
S05	1335	99		100		100		104		107	
S06	1338	100		100		100		107		109	
S07	1341	97		98		97		107		108	
S08	1344	99		100		92		106		107	
ICV003	1403	100		100		101		102		103	
ICB003	1406	97		98		98		100		100	
ICSA004	1410	102		104		100		109		110	
ICSAB004	1413	104		104		102		110		111	
CCV018	1416	102		104		98		107		109	
CCB018	1419	99		102		104		107		108	
ZZZZZZ	1422										
ZZZZZZ	1425										
ZZZZZZ	1429										
ZZZZZZ	1432										
ZZZZZZ	1435										
ZZZZZZ	1438										
ZZZZZZ	1442										
ZZZZZZ	1446										
ZZZZZZ	1449										
ZZZZZZ	1452										
ZZZZZZ	1455										
ZZZZZZ	1458										
ZZZZZZ	1502										
ZZZZZZ	1505										
ZZZZZZ	1508										
ZZZZZZ	1511										
ZZZZZZ	1515										
ZZZZZZ	1518										
ZZZZZZ	1521										
ZZZZZZ	1525										
CCV019	1528	97		100		96		106		106	
CCB019	1535	98		102		102		104		105	
ZZZZZZ	1539										
ZZZZZZ	1542										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1545										
ZZZZZZ	1549										
ZZZZZZ	1552										
ZZZZZZ	1555										
PBS273	1559	98		100		100		102		103	
LCS273	1602	98		100		101		101		103	
LCSD273	1605	98		101		101		102		102	
ZZZZZZ	1609										
ZZZZZZ	1612										
ZZZZZZ	1615										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1638										
ZZZZZZ	1641										
ZZZZZZ	1645										
ZZZZZZ	1648										
ZZZZZZ	1651										
CCV020	1655	98		101		97		106		108	
CCB020	1658	95		99		101		103		105	
ZZZZZZ	1701										
ZZZZZZ	1705										
ZZZZZZ	1708										
ZZZZZZ	1711										
ZZZZZZ	1714										
ZZZZZZ	1717										
ZZZZZZ	1720										
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1730										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
CCV021	1743	95		98		95		104		106	
CCB021	1745	95		99		102		105		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1749										
ZZZZZZ	1753										
ZZZZZZ	1756										
ZZZZZZ	1759										
ZZZZZZ	1802										
ZZZZZZ	1805										
ZZZZZZ	1809										
ZZZZZZ	1812										
ZZZZZZ	1815										
ZZZZZZ	1818										
ZZZZZZ	1821										
ZZZZZZ	1824										
ZZZZZZ	1828										
ZZZZZZ	1831										
CCV022	1834	95		97		96		107		108	
CCB022	1837	94		99		102		106		106	
ZZZZZZ	1841										
ZZZZZZ	1844										
ZZZZZZ	1847										
ZZZZZZ	1851										
ZZZZZZ	1854										
ZZZZZZ	1857										
ZZZZZZ	1900										
ZZZZZZ	1903										
ZZZZZZ	1907										
ZZZZZZ	1910										
ZZZZZZ	1913										
ZZZZZZ	1916										
ZZZZZZ	1919										
ZZZZZZ	1923										
ZZZZZZ	1926										
ZZZZZZ	1929										
ZZZZZZ	1932										
ZZZZZZ	1935										
CCV023	1939	93		98		95		106		108	
CCB023	1941	92		97		102		106		106	
ZZZZZZ	1945										
ZZZZZZ	1948										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1951										
ZZZZZZ	1955										
ZZZZZZ	1958										
ZZZZZZ	2001										
ZZZZZZ	2004										
ZZZZZZ	2008										
ZZZZZZ	2011										
ZZZZZZ	2014										
ZZZZZZ	2017										
ZZZZZZ	2021										
ZZZZZZ	2024										
ZZZZZZ	2027										
ZZZZZZ	2030										
ZZZZZZ	2033										
MYDB92	2036	97		102		97		100		101	
MYDB93	2040	94		100		96		101		101	
MYDB94	2043	94		100		96		99		100	
MYDB95	2046	95		101		95		99		101	
MYDB96	2049	95		101		97		101		102	
MYDB97	2052	96		103		98		101		101	
MYDB98	2056	97		101		96		100		102	
MYDB99	2059	96		100		97		100		101	
MYDBA0	2102	95		101		96		99		100	
MYDBA0D	2105	94		101		96		100		102	
CCV024	2108	92		98		96		107		109	
CCB024	2111	91		98		102		106		108	
MYDBA0L	2115	89		96		98		102		103	
MYDBA0S	2118	95		102		98		103		103	
MYDBA0SRE	2121	94		100		96		100		101	
MYDBA1	2124	94		101		96		99		101	
MYDBA2	2127	95		102		97		101		101	
MYDBA3	2130	100		107		98		103		103	
MYDBA4	2134	98		105		97		101		102	
MYDBA5	2137	98		105		97		100		102	
ICSA005	2141	94		99		99		106		107	
ICSAB005	2145	93		98		98		104		107	
CCV025	2151	94		98		96		106		108	
CCB025	2154	93		100		103		105		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/28/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
MYDBA6	2157	95		103		99		102		104	
MYDBA7	2200	101		107		98		101		102	
MYDBA9	2203	99		107		94		100		101	
MYDBB0	2207	98		106		99		102		102	
MYDBB1	2210	99		105		95		100		103	
MYDBB2	2213	97		105		96		101		101	
ZZZZZZ	2216										
ZZZZZZ	2219										
ZZZZZZ	2222										
ZZZZZZ	2226										
ZZZZZZ	2229										
ZZZZZZ	2232										
ZZZZZZ	2235										
ZZZZZZ	2239										
ZZZZZZ	2242										
ZZZZZZ	2245										
ZZZZZZ	2248										
ZZZZZZ	2251										
CCV026	2255	93		99		98		107		110	
CCB026	2257	90		98		102		104		105	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1813	100		100		100		100		100	
S02	1819	101		100		99		98		100	
S03	1822	100		100		99		98		100	
S04	1826	102		101		101		99		101	
S05	1829	103		102		102		100		104	
S06	1831	106		99		101		98		104	
S07	1834	114		105		107		102		110	
S08	1837	118		107		106		97		110	
ICV006	1842	105		104		102		101		104	
ICB006	1845	103		101		100		100		101	
ICSA006	1849	111		107		104		99		107	
ICSAB006	1852	107		107		107		103		111	
CCV029	1855	102		109		110		102		112	
CCB029	1858	100		105		105		106		108	
ZZZZZZ	1901										
ZZZZZZ	1904										
ZZZZZZ	1907										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1921										
ZZZZZZ	1924										
ZZZZZZ	1927										
ZZZZZZ	1930										
ZZZZZZ	1934										
ZZZZZZ	1937										
CCV030	1940	118		114		113		105		114	
CCB030	1943	110		106		105		104		107	
ZZZZZZ	1947										
ZZZZZZ	1950										
ZZZZZZ	1953										
ZZZZZZ	1956										
ZZZZZZ	2000										
ZZZZZZ	2006										
ZZZZZZ	2009										
ZZZZZZ	2012										
ZZZZZZ	2016										
ZZZZZZ	2019										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	2022										
ZZZZZZ	2025										
MYDBA0	2028	103		101		102		97		101	
MYDBA0D	2032	99		100		102		102		107	
MYDBA0L	2035	100		98		98		98		100	
MYDBA0S	2038	101		101		102		98		101	
MYDBA0SRE	2041	95		100		101		98		101	
MYDB99	2045	94		108		109		101		105	
MYDBA0ARE	2048	103		103		103		101		106	
ZZZZZZ	2051										
ZZZZZZ	2054										
ZZZZZZ	2057										
ZZZZZZ	2101										
ZZZZZZ	2104										
ZZZZZZ	2107										
MYDB99	2110	107		98		98		99		102	
CCV031	2113	112		108		109		103		112	
CCB031	2116	102		104		104		104		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1813	100		100		100		100		100	
S02	1819	99		98		99		100		98	
S03	1822	101		99		100		102		101	
S04	1826	100		101		100		99		101	
S05	1829	101		99		100		103		102	
S06	1831	102		102		101		106		106	
S07	1834	104		103		101		109		108	
S08	1837	107		104		99		109		108	
ICV006	1842	102		101		102		102		102	
ICB006	1845	101		102		102		103		101	
ICSA006	1849	105		104		101		107		107	
ICSAB006	1852	108		106		104		110		109	
CCV029	1855	110		106		102		108		109	
CCB029	1858	105		104		105		105		105	
ZZZZZZ	1901										
ZZZZZZ	1904										
ZZZZZZ	1907										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1921										
ZZZZZZ	1924										
ZZZZZZ	1927										
ZZZZZZ	1930										
ZZZZZZ	1934										
ZZZZZZ	1937										
CCV030	1940	110		106		104		111		110	
CCB030	1943	101		101		102		104		103	
ZZZZZZ	1947										
ZZZZZZ	1950										
ZZZZZZ	1953										
ZZZZZZ	1956										
ZZZZZZ	2000										
ZZZZZZ	2006										
ZZZZZZ	2009										
ZZZZZZ	2012										
ZZZZZZ	2016										
ZZZZZZ	2019										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	2022										
ZZZZZZ	2025										
MYDBA0	2028	100		101		101		103		102	
MYDBA0D	2032	100		102		101		103		102	
MYDBA0L	2035	99		100		101		102		104	
MYDBA0S	2038	100		101		101		101		101	
MYDBA0SRE	2041	100		100		100		101		102	
MYDB99	2045	108		108		102		104		103	
MYDBA0ARE	2048	100		100		102		103		104	
ZZZZZZ	2051										
ZZZZZZ	2054										
ZZZZZZ	2057										
ZZZZZZ	2101										
ZZZZZZ	2104										
ZZZZZZ	2107										
MYDB99	2110	95		98		101		103		103	
CCV031	2113	107		105		100		109		108	
CCB031	2116	102		102		103		104		104	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 01/08/2025
 Run Batch: LB134279 End Date : 01/08/2025

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1038	100		100		100		100		100	
S02	1045	99		101		101		101		100	
S03	1048	102		102		101		100		102	
S04	1051	104		100		101		100		102	
S05	1054	105		93		97		94		100	
S06	1057	103		85		91		88		98	
S07	1100	101		85		91		87		99	
S08	1102	101		90		92		83		98	
ICV001	1155	110		101		101		102		103	
ICB001	1200	107		100		101		101		103	
ICSA001	1203	106		91		97		91		103	
ICSAB001	1206	107		94		100		95		106	
CCV001	1209	104		93		99		90		105	
CCB001	1212	110		92		100		98		106	
PBS079	1215	110		96		102		103		108	
LCS079	1218	111		97		102		102		106	
LCSD079	1222	111		98		103		103		107	
MYDB99	1234	111		105		110		101		109	
CCV002	1237	107		91		96		89		103	
CCB002	1240	111		93		99		98		104	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB92
 Instrument ID: P7 Start Date : 01/08/2025
 Run Batch: LB134279 End Date : 01/08/2025

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1038	100		100		100		100		100	
S02	1045	100		100		100		100		101	
S03	1048	100		100		100		101		102	
S04	1051	97		97		97		100		101	
S05	1054	84		88		88		97		98	
S06	1057	83		87		87		97		98	
S07	1100	85		89		87		97		99	
S08	1102	92		93		84		96		98	
ICV001	1155	98		100		100		102		105	
ICB001	1200	99		100		102		101		103	
ICSA001	1203	89		94		91		101		105	
ICSAB001	1206	91		96		93		104		106	
CCV001	1209	88		91		87		99		101	
CCB001	1212	90		95		98		103		105	
PBS079	1215	93		97		101		104		107	
LCS079	1218	95		98		101		104		106	
LCSD079	1222	96		99		102		103		106	
MYDB99	1234	101		105		99		105		108	
CCV002	1237	88		91		87		99		103	
CCB002	1240	91		95		98		102		104	