

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE006	TUNE	10/29/24 17:38		Jaswal	OK
2	S00	S0	CAL1	10/29/24 18:13		Jaswal	OK
3	S02	S02	CAL3	10/29/24 18:19		Jaswal	OK
4	S03	S03	CAL4	10/29/24 18:22		Jaswal	OK
5	S04	S04	CAL5	10/29/24 18:26		Jaswal	OK
6	S05	S05	CAL6	10/29/24 18:29		Jaswal	OK
7	S06	S06	CAL7	10/29/24 18:31		Jaswal	OK
8	S07	S07	CAL8	10/29/24 18:34		Jaswal	OK
9	S08	S08	CAL9	10/29/24 18:37		Jaswal	OK
10	ICV006	ICV006	ICV	10/29/24 18:42		Jaswal	OK
11	ICB006	ICB006	ICB	10/29/24 18:45		Jaswal	OK
12	ICSA006	ICSA006	ICSA	10/29/24 18:49		Jaswal	OK
13	ICSAB006	ICSAB006	ICSAB	10/29/24 18:52		Jaswal	OK
14	CCV029	CCV029	CCV	10/29/24 18:55		Jaswal	OK
15	CCB029	CCB029	CCB	10/29/24 18:58		Jaswal	OK
16	PB164339BL	PBW339	MB	10/29/24 19:01		Jaswal	OK
17	PB164339BS	LCS339	LCS	10/29/24 19:04		Jaswal	OK
18	P4322-01	MYD0L6	SAM	10/29/24 19:07		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	
Chk Standard	MP82587,MP82586

19	P4322-02	MYD0Y6	SAM	10/29/24 19:11		Jaswal	OK
20	P4322-03	MYD0Y8	SAM	10/29/24 19:14		Jaswal	OK
21	P4322-04	MYD035	SAM	10/29/24 19:17		Jaswal	OK
22	P4322-05	MYD036	SAM	10/29/24 19:21		Jaswal	OK
23	P4322-06	MYD038	SAM	10/29/24 19:24		Jaswal	OK
24	P4322-07	MYDD00	SAM	10/29/24 19:27		Jaswal	OK
25	P4322-08	MYDD01	SAM	10/29/24 19:30		Jaswal	OK
26	P4322-09	MYDD02	SAM	10/29/24 19:34		Jaswal	OK
27	P4322-10	MYDD03	SAM	10/29/24 19:37		Jaswal	OK
28	CCV030	CCV030	CCV	10/29/24 19:40		Jaswal	OK
29	CCB030	CCB030	CCB	10/29/24 19:43		Jaswal	OK
30	P4314-10DL	MYDB51	SAM	10/29/24 19:47	4X for Ba,Co,Ni,V	Jaswal	Confirms
31	P4314-11DL	MYDB51D	DUP	10/29/24 19:50	4X for Ba,Co,Ni,V	Jaswal	Confirms
32	P4314-10LDL	MYDB51L	SD	10/29/24 19:53	4X for Ba,Co,Ni,V	Jaswal	Confirms
33	P4314-12DL	MYDB51S	MS	10/29/24 19:56	4X for Ba,Co,Ni,V	Jaswal	Confirms
34	P4314-10MSDL	MYDB51SRE	MS	10/29/24 20:00	4X for Ba,Co,Ni,V	Jaswal	Confirms
35	P4314-10A	MYDB51ARE	PS	10/29/24 20:06	PS for As-0.0047ml(M5801),Pb-0.0038ml(M6026),Ag-0.02ml(MP82908),Ti-0.021ml(MP82960)-10ml	Jaswal	OK
36	P4316-16DL	MYDB88	SAM	10/29/24 20:09	4X for Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

37	P4316-17DL	MYDB88D	DUP	10/29/24 20:12	4X for Ba,Co,Ni,V	Jaswal	Confirms
38	P4316-16LDL	MYDB88L	SD	10/29/24 20:16	4X for Ba,Co,Ni,V	Jaswal	Confirms
39	P4316-18DL	MYDB88S	MS	10/29/24 20:19	4X for Ba,Co,Ni,V	Jaswal	Confirms
40	P4316-16MSDL	MYDB88SRE	MS	10/29/24 20:22	4X for Ba,Co,Ni,V	Jaswal	Confirms
41	P4316-16A	MYDB88ARE	PS	10/29/24 20:25	PS for Ag-0.02ml(MP82908)-10ml	Jaswal	OK
42	P4317-09DL	MYDBA0	SAM	10/29/24 20:28	4X for Ba,Co,Ni,V	Jaswal	Confirms
43	P4317-10DL	MYDBA0D	DUP	10/29/24 20:32	4X for Ba,Co,Ni,V	Jaswal	Confirms
44	P4317-09LDL	MYDBA0L	SD	10/29/24 20:35	4X for Ba,Co,Ni,V	Jaswal	Confirms
45	P4317-11DL	MYDBA0S	MS	10/29/24 20:38	4X for Ba,Co,Ni,V	Jaswal	Confirms
46	P4317-09MSDL	MYDBA0SRE	MS	10/29/24 20:41	4X for Ba,Co,Ni,V	Jaswal	Confirms
47	P4317-08DL	MYDB99	SAM	10/29/24 20:45	Not used	Jaswal	Not Ok
48	P4317-09A	MYDBA0ARE	PS	10/29/24 20:48	PS for Pb-0.0034ml(M6026),Ag-0.02ml(MP82908)-10ml	Jaswal	OK
49	P4318-20DL	MYDB72	SAM	10/29/24 20:51	4X for Ba,Co,Ni,V	Jaswal	Confirms
50	P4318-21DL	MYDB72D	DUP	10/29/24 20:54	4X for Ba,Co,Ni,V	Jaswal	Confirms
51	P4318-20LDL	MYDB72L	SD	10/29/24 20:57	4X for Ba,Co,Ni,V	Jaswal	Confirms
52	P4318-22DL	MYDB72S	MS	10/29/24 21:01	4X for Ba,Co,Ni,V	Jaswal	Confirms
53	P4318-20MSDL	MYDB72SRE	MS	10/29/24 21:04	4X for Ba,Co,Ni,V	Jaswal	Confirms
54	P4318-20A	MYDB72ARE	PS	10/29/24 21:07	PS for Ag-0.02ml(MP82908)-10ml	Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

55	P4317-08	MYDB99	SAM	10/29/24 21:10		Jaswal	OK
56	CCV031	CCV031	CCV	10/29/24 21:13		Jaswal	OK
57	CCB031	CCB031	CCB	10/29/24 21:16		Jaswal	OK