

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB12
 Instrument ID: P7 Start Date: 10/28/2024
 Analytical Method: ICP-MS Run Batch: LB133179
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	1.9	-5	50	49	-2
Arsenic	0	0.0	0	1	0.98	-2	50	51	2
Barium	0	0.0	0	10	8.8	-12	250	240	-4
Beryllium	0	0.0	0	1	1.0	0	50	57	14
Cadmium	0	0.0	0	1	0.96	-4	50	53	6
Chromium	0	0.0	0	2	2.0	0	50	51	2
Cobalt	0	0.0	0	1	1.0	0	50	51	2
Copper	0	0.0	0	2	1.8	-10	500	520	4
Lead	0	0.0	0	1	0.82	-18	250	240	-4
Nickel	0	0.0	0	1	1.0	0	50	53	6
Selenium	0	0.0	0	5	4.7	-6	50	54	8
Silver	0	0.0	0	1	0.9	-10	50	51	2
Thallium	0	0.0	0	1	0.83	-17	50	44	-12
Vanadium	0	0.0	0	5	4.8	-4	50	51	2
Zinc	0	0.0	0	5	4.7	-6	500	510	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB12
 Instrument ID: P7 Start Date: 10/28/2024
 Analytical Method: ICP-MS Run Batch: LB133179
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	120	-4	250	250	0	500	500	0
Arsenic	125	130	4	250	250	0	500	500	0
Barium	625	640	2	1250	1300	4	2500	2500	0
Beryllium	125	140	12	250	260	4	500	520	4
Cadmium	125	130	4	250	260	4	500	510	2
Chromium	125	130	4	250	250	0	500	500	0
Cobalt	125	130	4	250	250	0	500	510	2
Copper	1250	1300	4	2500	2600	4	5000	5000	0
Lead	625	620	-1	1250	1300	4	2500	2500	0
Nickel	125	130	4	250	260	4	500	510	2
Selenium	125	130	4	250	260	4	500	500	0
Silver	125	130	4	250	260	4	500	510	2
Thallium	125	110	-12	250	250	0	500	490	-2
Vanadium	125	130	4	250	250	0	500	500	0
Zinc	1250	1300	4	2500	2500	0	5000	5100	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB12
 Instrument ID: P7 Start Date: 10/28/2024
 Analytical Method: ICP-MS Run Batch: LB133179
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	1000	0						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	990	-1						
Cadmium	1000	990	-1						
Chromium	1000	1000	0						
Cobalt	1000	1000	0						
Copper	10000	10000	0						
Lead	5000	5000	0						
Nickel	1000	990	-1						
Selenium	1000	990	-1						
Silver	1000	990	-1						
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	10000	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB12
 Instrument ID: P7 Start Date: 10/29/2024
 Analytical Method: ICP-MS Run Batch: LB133196
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	1.9	-5	50	48	-4
Arsenic	0	0.0	0	1	0.97	-3	50	51	2
Barium	0	0.0	0	10	8.8	-12	250	240	-4
Beryllium	0	0.0	0	1	0.99	-1	50	52	4
Cadmium	0	0.0	0	1	1.0	0	50	53	6
Chromium	0	0.0	0	2	2.0	0	50	51	2
Cobalt	0	0.0	0	1	1.0	0	50	51	2
Copper	0	0.0	0	2	1.9	-5	500	520	4
Lead	0	0.0	0	1	0.81	-19	250	230	-8
Nickel	0	0.0	0	1	1.1	10	50	52	4
Selenium	0	0.0	0	5	5.1	2	50	52	4
Silver	0	0.0	0	1	0.87	-13	50	50	0
Thallium	0	0.0	0	1	0.83	-17	50	44	-12
Vanadium	0	0.0	0	5	4.9	-2	50	51	2
Zinc	0	0.0	0	5	4.5	-10	500	500	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: _____ SDG No.: MYDB12
 Instrument ID: P7 Start Date: 10/29/2024
 Analytical Method: ICP-MS Run Batch: LB133196
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	120	-4	250	240	-4	500	520	4
Arsenic	125	120	-4	250	250	0	500	500	0
Barium	625	590	-6	1250	1300	4	2500	2500	0
Beryllium	125	130	4	250	260	4	500	500	0
Cadmium	125	130	4	250	260	4	500	520	4
Chromium	125	130	4	250	250	0	500	500	0
Cobalt	125	130	4	250	250	0	500	500	0
Copper	1250	1300	4	2500	2500	0	5000	5000	0
Lead	625	620	-1	1250	1200	-4	2500	2500	0
Nickel	125	130	4	250	260	4	500	510	2
Selenium	125	130	4	250	250	0	500	510	2
Silver	125	120	-4	250	260	4	500	510	2
Thallium	125	110	-12	250	250	0	500	510	2
Vanadium	125	130	4	250	250	0	500	500	0
Zinc	1250	1300	4	2500	2500	0	5000	5000	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDB12
Instrument ID: P7 Start Date: 10/29/2024
Analytical Method: ICP-MS Run Batch: LB133196
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	990	-1						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	1000	0						
Cadmium	1000	990	-1						
Chromium	1000	1000	0						
Cobalt	1000	1000	0						
Copper	10000	10000	0						
Lead	5000	5000	0						
Nickel	1000	990	-1						
Selenium	1000	990	-1						
Silver	1000	990	-1						
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	10000	0						