

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51772 MA No.: \_\_\_\_\_ SDG No.: MYD0L6  
 Instrument ID: P7 Start Date : 10/29/2024  
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element		Element		Element		Element		Element	
		6Li	Q	45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q
S0	1813	100		100		100		100		100	
S02	1819	101		100		99		98		100	
S03	1822	100		100		99		98		100	
S04	1826	102		101		101		99		101	
S05	1829	103		102		102		100		104	
S06	1831	106		99		101		98		104	
S07	1834	114		105		107		102		110	
S08	1837	118		107		106		97		110	
ICV006	1842	105		104		102		101		104	
ICB006	1845	103		101		100		100		101	
ICSA006	1849	111		107		104		99		107	
ICSAB006	1852	107		107		107		103		111	
CCV029	1855	102		109		110		102		112	
CCB029	1858	100		105		105		106		108	
PBW339	1901	97		101		101		100		104	
LCS339	1904	98		102		102		101		103	
MYD0L6	1907	94		93		92		92		94	
MYD0Y6	1911	96		92		93		92		95	
MYD0Y8	1914	98		94		94		95		97	
MYD035	1917	99		94		94		95		96	
MYD036	1921	98		93		93		93		95	
MYD038	1924	96		94		95		94		96	
MYDD00	1927	93		93		92		92		95	
MYDD01	1930	94		94		95		94		96	
MYDD02	1934	98		97		95		95		97	
MYDD03	1937	95		92		92		92		95	
CCV030	1940	118		114		113		105		114	
CCB030	1943	110		106		105		104		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51772 MA No.: \_\_\_\_\_ SDG No.: MYD0L6  
 Instrument ID: P7 Start Date : 10/29/2024  
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1813	100		100		100		100		100	
S02	1819	99		98		99		100		98	
S03	1822	101		99		100		102		101	
S04	1826	100		101		100		99		101	
S05	1829	101		99		100		103		102	
S06	1831	102		102		101		106		106	
S07	1834	104		103		101		109		108	
S08	1837	107		104		99		109		108	
ICV006	1842	102		101		102		102		102	
ICB006	1845	101		102		102		103		101	
ICSA006	1849	105		104		101		107		107	
ICSAB006	1852	108		106		104		110		109	
CCV029	1855	110		106		102		108		109	
CCB029	1858	105		104		105		105		105	
PBW339	1901	102		101		103		103		103	
LCS339	1904	101		101		102		102		102	
MYD0L6	1907	94		93		95		97		96	
MYD0Y6	1911	90		90		92		93		94	
MYD0Y8	1914	96		95		98		99		98	
MYD035	1917	96		97		99		99		99	
MYD036	1921	95		95		96		97		96	
MYD038	1924	96		95		97		97		97	
MYDD00	1927	94		93		95		96		95	
MYDD01	1930	95		96		98		97		97	
MYDD02	1934	96		96		97		98		98	
MYDD03	1937	95		95		97		98		96	
CCV030	1940	110		106		104		111		110	
CCB030	1943	101		101		102		104		103	