

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51772 MA No. : SDG No.: MYD0L6Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB133196 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV006				ID: CCV029				ID: CCV030		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	500	480	96	4	50000	50000	100	1	50000	101	1
Antimony	200	190	95	1	500	490	98	3	500	101	0
Arsenic	200	200	102	5	500	540	108	1	540	108	1
Barium	100	91	91	1	2500	2500	100	2	2600	102	1
Beryllium	100	100	102	1	500	520	104	1	500	100	1
Cadmium	100	100	103	1	500	520	103	2	520	104	2
Calcium	2000	2000	101	2	250000	250000	100	2	250000	102	1
Chromium	100	100	103	5	500	500	99	2	500	101	1
Cobalt	100	100	102	5	500	520	104	0	520	104	0
Copper	100	95	95	5	5000	5000	100	1	5100	102	1
Iron	2000	2100	104	6	125000	130000	101	1	130000	102	0
Lead	200	190	94	1	2500	2500	101	2	2600	102	2
Magnesium	1200	1200	103	5	250000	250000	100	1	250000	101	1
Manganese	100	100	101	6	5000	4900	98	1	4900	99	1
Nickel	110	110	96	6	500	510	103	1	510	103	0
Potassium	2000	2100	103	5	125000	130000	100	0	130000	101	1
Selenium	200	210	105	0	500	520	104	2	520	104	2
Silver	50.0	47	94	1	500	510	102	2	520	103	1
Sodium	2000	2100	106	6	250000	250000	99	1	250000	102	0
Thallium	210	220	104	1	500	510	102	2	510	102	2
Vanadium	100	100	100	6	500	510	101	1	500	101	1
Zinc	200	200	98	5	5000	5100	103	1	5200	103	0