

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1322	100		100		100		100		100	
S02	1326	98		99		99		99		99	
S03	1329	96		97		97		98		98	
S04	1332	102		99		100		99		101	
S05	1335	109		100		101		99		103	
S06	1338	113		100		103		100		104	
S07	1341	116		99		103		99		105	
S08	1344	112		99		99		91		103	
ICV003	1403	104		100		100		99		101	
ICB003	1406	101		98		98		97		98	
ICSA004	1410	107		99		102		98		105	
ICSAB004	1413	108		104		105		100		107	
CCV018	1416	106		102		105		97		107	
CCB018	1419	109		102		104		103		107	
ZZZZZZ	1422										
ZZZZZZ	1425										
ZZZZZZ	1429										
ZZZZZZ	1432										
ZZZZZZ	1435										
ZZZZZZ	1438										
ZZZZZZ	1442										
ZZZZZZ	1446										
ZZZZZZ	1449										
ZZZZZZ	1452										
ZZZZZZ	1455										
ZZZZZZ	1458										
ZZZZZZ	1502										
ZZZZZZ	1505										
ZZZZZZ	1508										
ZZZZZZ	1511										
ZZZZZZ	1515										
ZZZZZZ	1518										
ZZZZZZ	1521										
ZZZZZZ	1525										
CCV019	1528	98		98		102		95		105	
CCB019	1535	97		101		102		102		104	
ZZZZZZ	1539										
ZZZZZZ	1542										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Run Batch: LB133179 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1545										
ZZZZZZ	1549										
ZZZZZZ	1552										
ZZZZZZ	1555										
ZZZZZZ	1559										
ZZZZZZ	1602										
ZZZZZZ	1605										
ZZZZZZ	1609										
ZZZZZZ	1612										
ZZZZZZ	1615										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1638										
ZZZZZZ	1641										
ZZZZZZ	1645										
ZZZZZZ	1648										
ZZZZZZ	1651										
CCV020	1655	99		99		104		96		106	
CCB020	1658	101		98		101		102		106	
ZZZZZZ	1701										
ZZZZZZ	1705										
ZZZZZZ	1708										
ZZZZZZ	1711										
ZZZZZZ	1714										
ZZZZZZ	1717										
ZZZZZZ	1720										
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1730										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
CCV021	1743	92		97		103		94		104	
CCB021	1745	96		99		103		103		106	

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Instrument ID: P7	Start Date : 10/28/2024
Run Batch: LB133179	End Date : 10/29/2024

[illegible]

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1951										
ZZZZZZ	1955										
ZZZZZZ	1958										
ZZZZZZ	2001										
ZZZZZZ	2004										
ZZZZZZ	2008										
ZZZZZZ	2011										
ZZZZZZ	2014										
ZZZZZZ	2017										
ZZZZZZ	2021										
ZZZZZZ	2024										
ZZZZZZ	2027										
ZZZZZZ	2030										
ZZZZZZ	2033										
ZZZZZZ	2036										
ZZZZZZ	2040										
ZZZZZZ	2043										
ZZZZZZ	2046										
ZZZZZZ	2049										
ZZZZZZ	2052										
ZZZZZZ	2056										
ZZZZZZ	2059										
ZZZZZZ	2102										
ZZZZZZ	2105										
CCV024	2108	96		97		104		98		107	
CCB024	2111	100		98		102		102		108	
ZZZZZZ	2115										
ZZZZZZ	2118										
ZZZZZZ	2121										
ZZZZZZ	2124										
ZZZZZZ	2127										
ZZZZZZ	2130										
ZZZZZZ	2134										
ZZZZZZ	2137										
ICSA005	2141	89		99		102		98		106	
ICSAB005	2145	88		97		102		98		107	
CCV025	2151	94		97		103		97		108	
CCB025	2154	92		97		103		104		108	

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Lab Name:	Alliance Technical Group, LLC	Contract:	68HERH20D0011
Lab Code:	ACE	Case No.:	51772
		MA No.:	SDG No.: MYDBB4
Instrument ID:	P7	Start Date :	10/28/2024
Run Batch:	LB133179	End Date :	10/29/2024

[illegible]

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	2359										
ZZZZZZ	0002										
ZZZZZZ	0005										
ZZZZZZ	0009										
ZZZZZZ	0012										
ZZZZZZ	0015										
ZZZZZZ	0018										
ZZZZZZ	0021										
CCV027	0027	96		95		100		96		107	
CCB027	0030	97		97		103		103		108	
ZZZZZZ	0033										
ZZZZZZ	0037										
ZZZZZZ	0040										
ZZZZZZ	0043										
ZZZZZZ	0046										
ZZZZZZ	0049										
ZZZZZZ	0053										
ZZZZZZ	0056										
ZZZZZZ	0059										
ZZZZZZ	0103										
ZZZZZZ	0106										
ZZZZZZ	0110										
ZZZZZZ	0113										
MYDBB4	0116	86		104		112		99		105	
MYDBB5	0119	86		106		113		99		105	
MYDBB6	0122	86		108		115		99		106	
MYDBB7	0126	88		104		114		100		106	
MYDBB7D	0129	87		104		112		100		107	
MYDBB7L	0132	85		99		104		102		106	
MYDBB7S	0135	86		103		111		99		106	
MYDBB7SRE	0138	85		101		110		98		105	
MYDBB8	0141	83		105		112		98		105	
MYDBB9	0144	83		105		111		99		105	
MYDBC0	0147	82		100		108		99		106	
CCV028	0151	89		91		97		91		102	
CCB028	0153	94		96		102		102		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1322	100		100		100		100		100	
S02	1326	100		100		99		101		102	
S03	1329	99		99		99		101		101	
S04	1332	100		100		99		102		105	
S05	1335	99		100		100		104		107	
S06	1338	100		100		100		107		109	
S07	1341	97		98		97		107		108	
S08	1344	99		100		92		106		107	
ICV003	1403	100		100		101		102		103	
ICB003	1406	97		98		98		100		100	
ICSA004	1410	102		104		100		109		110	
ICSAB004	1413	104		104		102		110		111	
CCV018	1416	102		104		98		107		109	
CCB018	1419	99		102		104		107		108	
ZZZZZZ	1422										
ZZZZZZ	1425										
ZZZZZZ	1429										
ZZZZZZ	1432										
ZZZZZZ	1435										
ZZZZZZ	1438										
ZZZZZZ	1442										
ZZZZZZ	1446										
ZZZZZZ	1449										
ZZZZZZ	1452										
ZZZZZZ	1455										
ZZZZZZ	1458										
ZZZZZZ	1502										
ZZZZZZ	1505										
ZZZZZZ	1508										
ZZZZZZ	1511										
ZZZZZZ	1515										
ZZZZZZ	1518										
ZZZZZZ	1521										
ZZZZZZ	1525										
CCV019	1528	97		100		96		106		106	
CCB019	1535	98		102		102		104		105	
ZZZZZZ	1539										
ZZZZZZ	1542										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Contract: 68HERH20D0011

MA No.: SDG No.: MYDBB4

Start Date : 10/28/2024

End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
ZZZZZZ	1545										
ZZZZZZ	1549										
ZZZZZZ	1552										
ZZZZZZ	1555										
ZZZZZZ	1559										
ZZZZZZ	1602										
ZZZZZZ	1605										
ZZZZZZ	1609										
ZZZZZZ	1612										
ZZZZZZ	1615										
ZZZZZZ	1619										
ZZZZZZ	1622										
ZZZZZZ	1625										
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1638										
ZZZZZZ	1641										
ZZZZZZ	1645										
ZZZZZZ	1648										
ZZZZZZ	1651										
CCV020	1655	98		101		97		106		108	
CCB020	1658	95		99		101		103		105	
ZZZZZZ	1701										
ZZZZZZ	1705										
ZZZZZZ	1708										
ZZZZZZ	1711										
ZZZZZZ	1714										
ZZZZZZ	1717										
ZZZZZZ	1720										
ZZZZZZ	1724										
ZZZZZZ	1727										
ZZZZZZ	1730										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
CCV021	1743	95		98		95		104		106	
CCB021	1745	95		99		102		105		107	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	1951										
ZZZZZZ	1955										
ZZZZZZ	1958										
ZZZZZZ	2001										
ZZZZZZ	2004										
ZZZZZZ	2008										
ZZZZZZ	2011										
ZZZZZZ	2014										
ZZZZZZ	2017										
ZZZZZZ	2021										
ZZZZZZ	2024										
ZZZZZZ	2027										
ZZZZZZ	2030										
ZZZZZZ	2033										
ZZZZZZ	2036										
ZZZZZZ	2040										
ZZZZZZ	2043										
ZZZZZZ	2046										
ZZZZZZ	2049										
ZZZZZZ	2052										
ZZZZZZ	2056										
ZZZZZZ	2059										
ZZZZZZ	2102										
ZZZZZZ	2105										
CCV024	2108	92		98		96		107		109	
CCB024	2111	91		98		102		106		108	
ZZZZZZ	2115										
ZZZZZZ	2118										
ZZZZZZ	2121										
ZZZZZZ	2124										
ZZZZZZ	2127										
ZZZZZZ	2130										
ZZZZZZ	2134										
ZZZZZZ	2137										
ICSA005	2141	94		99		99		106		107	
ICSAB005	2145	93		98		98		104		107	
CCV025	2151	94		98		96		106		108	
CCB025	2154	93		100		103		105		107	

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Run Batch: LB133179 End Date : 10/29/2024

[illegible]

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/28/2024
 Run Batch: LB133179 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
ZZZZZZ	2359										
ZZZZZZ	0002										
ZZZZZZ	0005										
ZZZZZZ	0009										
ZZZZZZ	0012										
ZZZZZZ	0015										
ZZZZZZ	0018										
ZZZZZZ	0021										
CCV027	0027	89		95		94		105		107	
CCB027	0030	90		98		102		105		106	
ZZZZZZ	0033										
ZZZZZZ	0037										
ZZZZZZ	0040										
ZZZZZZ	0043										
ZZZZZZ	0046										
ZZZZZZ	0049										
ZZZZZZ	0053										
ZZZZZZ	0056										
ZZZZZZ	0059										
ZZZZZZ	0103										
ZZZZZZ	0106										
ZZZZZZ	0110										
ZZZZZZ	0113										
MYDBB4	0116	101		107		99		103		105	
MYDBB5	0119	101		110		99		103		104	
MYDBB6	0122	103		110		99		105		106	
MYDBB7	0126	100		109		101		107		106	
MYDBB7D	0129	100		109		100		105		106	
MYDBB7L	0132	94		101		101		103		106	
MYDBB7S	0135	98		106		98		103		105	
MYDBB7SRE	0138	96		104		97		103		103	
MYDBB8	0141	97		104		94		98		100	
MYDBB9	0144	101		107		98		103		105	
MYDBC0	0147	98		104		100		103		106	
CCV028	0151	89		95		93		104		106	
CCB028	0153	89		97		101		104		105	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1813	100		100		100		100		100	
S02	1819	101		100		99		98		100	
S03	1822	100		100		99		98		100	
S04	1826	102		101		101		99		101	
S05	1829	103		102		102		100		104	
S06	1831	106		99		101		98		104	
S07	1834	114		105		107		102		110	
S08	1837	118		107		106		97		110	
ICV006	1842	105		104		102		101		104	
ICB006	1845	103		101		100		100		101	
ICSA006	1849	111		107		104		99		107	
ICSAB006	1852	107		107		107		103		111	
CCV029	1855	102		109		110		102		112	
CCB029	1858	100		105		105		106		108	
ZZZZZZ	1901										
ZZZZZZ	1904										
ZZZZZZ	1907										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1921										
ZZZZZZ	1924										
ZZZZZZ	1927										
ZZZZZZ	1930										
ZZZZZZ	1934										
ZZZZZZ	1937										
CCV030	1940	118		114		113		105		114	
CCB030	1943	110		106		105		104		107	
ZZZZZZ	1947										
ZZZZZZ	1950										
ZZZZZZ	1953										
ZZZZZZ	1956										
ZZZZZZ	2000										
ZZZZZZ	2006										
ZZZZZZ	2009										
ZZZZZZ	2012										
ZZZZZZ	2016										
ZZZZZZ	2019										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	2022										
ZZZZZZ	2025										
ZZZZZZ	2028										
ZZZZZZ	2032										
ZZZZZZ	2035										
ZZZZZZ	2038										
ZZZZZZ	2041										
ZZZZZZ	2045										
ZZZZZZ	2048										
ZZZZZZ	2051										
ZZZZZZ	2054										
ZZZZZZ	2057										
ZZZZZZ	2101										
ZZZZZZ	2104										
ZZZZZZ	2107										
ZZZZZZ	2110										
CCV031	2113	112		108		109		103		112	
CCB031	2116	102		104		104		104		107	
ZZZZZZ	2120										
ZZZZZZ	2133										
ZZZZZZ	2136										
MYDBC1	2140	95		112		114		102		108	
MYDBC2	2143	96		111		115		105		110	
MYDBC3	2146	98		111		115		102		109	
MYDBC4	2149	97		112		115		104		109	
MYDBC5	2152	99		113		117		102		109	
MYDBC6	2155	92		109		112		102		107	
MYDBC7	2158	95		106		111		102		108	
MYDBC8	2202	95		109		111		102		108	
MYDBC9	2205	97		112		114		104		110	
MYDBD0	2208	94		109		112		101		107	
MYDBD1	2211	97		109		114		102		109	
MYDBD2	2214	90		108		113		102		107	
MYDBD3	2217	88		109		113		102		110	
ZZZZZZ	2220										
ZZZZZZ	2223										
CCV032	2307	92		106		108		100		110	
CCB032	2310	90		102		104		103		105	

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Lab Name:	Alliance Technical Group, LLC	Contract:	68HERH20D0011
Lab Code:	ACE	Case No.:	51772
		MA No.:	SDG No.: MYDBB4
Instrument ID:	P7	Start Date :	10/29/2024
Run Batch:	LB133196	End Date :	10/29/2024

[illegible]

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1813	100		100		100		100		100	
S02	1819	99		98		99		100		98	
S03	1822	101		99		100		102		101	
S04	1826	100		101		100		99		101	
S05	1829	101		99		100		103		102	
S06	1831	102		102		101		106		106	
S07	1834	104		103		101		109		108	
S08	1837	107		104		99		109		108	
ICV006	1842	102		101		102		102		102	
ICB006	1845	101		102		102		103		101	
ICSA006	1849	105		104		101		107		107	
ICSAB006	1852	108		106		104		110		109	
CCV029	1855	110		106		102		108		109	
CCB029	1858	105		104		105		105		105	
ZZZZZZ	1901										
ZZZZZZ	1904										
ZZZZZZ	1907										
ZZZZZZ	1911										
ZZZZZZ	1914										
ZZZZZZ	1917										
ZZZZZZ	1921										
ZZZZZZ	1924										
ZZZZZZ	1927										
ZZZZZZ	1930										
ZZZZZZ	1934										
ZZZZZZ	1937										
CCV030	1940	110		106		104		111		110	
CCB030	1943	101		101		102		104		103	
ZZZZZZ	1947										
ZZZZZZ	1950										
ZZZZZZ	1953										
ZZZZZZ	1956										
ZZZZZZ	2000										
ZZZZZZ	2006										
ZZZZZZ	2009										
ZZZZZZ	2012										
ZZZZZZ	2016										
ZZZZZZ	2019										

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/29/2024
 Run Batch: LB133196 End Date : 10/29/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
ZZZZZZ	2022										
ZZZZZZ	2025										
ZZZZZZ	2028										
ZZZZZZ	2032										
ZZZZZZ	2035										
ZZZZZZ	2038										
ZZZZZZ	2041										
ZZZZZZ	2045										
ZZZZZZ	2048										
ZZZZZZ	2051										
ZZZZZZ	2054										
ZZZZZZ	2057										
ZZZZZZ	2101										
ZZZZZZ	2104										
ZZZZZZ	2107										
ZZZZZZ	2110										
CCV031	2113	107		105		100		109		108	
CCB031	2116	102		102		103		104		104	
ZZZZZZ	2120										
ZZZZZZ	2133										
ZZZZZZ	2136										
MYDBC1	2140	112		111		102		105		104	
MYDBC2	2143	109		111		103		105		104	
MYDBC3	2146	105		105		95		99		98	
MYDBC4	2149	112		112		103		105		105	
MYDBC5	2152	114		115		103		105		104	
MYDBC6	2155	110		111		102		104		103	
MYDBC7	2158	107		108		102		104		104	
MYDBC8	2202	110		110		103		106		105	
MYDBC9	2205	112		111		103		106		106	
MYDBD0	2208	106		108		100		101		99	
MYDBD1	2211	108		109		102		103		103	
MYDBD2	2214	108		109		102		103		103	
MYDBD3	2217	107		106		98		101		99	
ZZZZZZ	2220										
ZZZZZZ	2223										
CCV032	2307	106		104		101		107		107	
CCB032	2310	101		102		104		104		103	

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Lab Name:	Alliance Technical Group, LLC	Contract:	68HERH20D0011
Lab Code:	ACE	Case No.:	51772
		MA No.:	SDG No.: MYDBB4
Instrument ID:	P7	Start Date :	10/29/2024
Run Batch:	LB133196	End Date :	10/29/2024

[illegible]

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/30/2024
 Run Batch: LB133210 End Date : 10/30/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1218	100		100		100		100		100	
S02	1225	92		94		93		95		94	
S03	1228	93		93		94		95		94	
S04	1231	100		94		95		95		97	
S05	1234	107		95		95		95		100	
S06	1237	115		94		96		95		101	
S07	1240	116		93		96		94		104	
S08	1243	116		94		94		88		104	
ICV007	1259	103		94		93		94		96	
ICB007	1304	109		98		98		100		101	
ICSA007	1307	106		95		94		92		100	
ICSAB007	1310	109		96		98		95		104	
CCV035	1313	106		98		99		94		104	
CCB035	1316	104		98		100		101		104	
PBS277	1319	104		99		100		101		104	
LCS277	1323	105		100		99		100		104	
LCSD277	1326	105		99		99		101		105	
MYDBB7	1329	106		99		101		99		103	
MYDBB7D	1333	106		100		101		98		103	
MYDBB7L	1336	106		97		97		99		102	
MYDBB7S	1339	103		96		96		98		101	
MYDBB7SRE	1342	104		97		96		98		101	
MYDBB7S	1348	109		102		102		100		104	
MYDBB7SRE	1351	105		96		97		96		101	
MYDBB7ARE	1354	107		106		107		98		104	
CCV036	1357	108		98		100		96		105	
CCB036	1400	106		99		100		101		105	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51772 MA No.: SDG No.: MYDBB4
 Instrument ID: P7 Start Date : 10/30/2024
 Run Batch: LB133210 End Date : 10/30/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1218	100		100		100		100		100	
S02	1225	98		98		96		96		97	
S03	1228	98		98		96		98		99	
S04	1231	96		98		97		100		100	
S05	1234	95		97		96		103		103	
S06	1237	93		96		94		103		105	
S07	1240	93		96		94		104		106	
S08	1243	95		97		90		103		105	
ICV007	1259	91		93		93		95		96	
ICB007	1304	98		100		100		102		103	
ICSA007	1307	95		98		93		102		105	
ICSAB007	1310	98		101		98		105		108	
CCV035	1313	96		98		92		102		103	
CCB035	1316	100		102		101		105		105	
PBS277	1319	99		102		102		104		105	
LCS277	1323	98		100		101		103		106	
LCSD277	1326	99		102		102		105		105	
MYDBB7	1329	101		103		99		104		105	
MYDBB7D	1333	100		102		99		104		104	
MYDBB7L	1336	96		100		98		102		104	
MYDBB7S	1339	95		98		98		101		102	
MYDBB7SRE	1342	95		99		98		102		102	
MYDBB7S	1348	100		103		99		104		104	
MYDBB7SRE	1351	98		103		99		103		105	
MYDBB7ARE	1354	104		108		98		105		105	
CCV036	1357	99		100		95		106		108	
CCB036	1400	99		102		101		103		106	