

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard	M5801,MP82590,M5819,MP82908		
LCS Standard			
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE006	TUNE	10/29/24 17:38		Jaswal	OK
2	S00	S0	CAL1	10/29/24 18:13		Jaswal	OK
3	S02	S02	CAL3	10/29/24 18:19		Jaswal	OK
4	S03	S03	CAL4	10/29/24 18:22		Jaswal	OK
5	S04	S04	CAL5	10/29/24 18:26		Jaswal	OK
6	S05	S05	CAL6	10/29/24 18:29		Jaswal	OK
7	S06	S06	CAL7	10/29/24 18:31		Jaswal	OK
8	S07	S07	CAL8	10/29/24 18:34		Jaswal	OK
9	S08	S08	CAL9	10/29/24 18:37		Jaswal	OK
10	ICV006	ICV006	ICV	10/29/24 18:42		Jaswal	OK
11	ICB006	ICB006	ICB	10/29/24 18:45		Jaswal	OK
12	ICSA006	ICSA006	ICSA	10/29/24 18:49		Jaswal	OK
13	ICSAB006	ICSAB006	ICSAB	10/29/24 18:52		Jaswal	OK
14	CCV029	CCV029	CCV	10/29/24 18:55		Jaswal	OK
15	CCB029	CCB029	CCB	10/29/24 18:58		Jaswal	OK
16	PB164339BL	PBW339	MB	10/29/24 19:01		Jaswal	OK
17	PB164339BS	LCS339	LCS	10/29/24 19:04		Jaswal	OK
18	P4322-01	MYD0L6	SAM	10/29/24 19:07		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard	M5801,MP82590,M5819,MP82908		
LCS Standard			
Chk Standard	MP82587,MP82586		

19	P4322-02	MYD0Y6	SAM	10/29/24 19:11		Jaswal	OK
20	P4322-03	MYD0Y8	SAM	10/29/24 19:14		Jaswal	OK
21	P4322-04	MYD035	SAM	10/29/24 19:17		Jaswal	OK
22	P4322-05	MYD036	SAM	10/29/24 19:21		Jaswal	OK
23	P4322-06	MYD038	SAM	10/29/24 19:24		Jaswal	OK
24	P4322-07	MYDD00	SAM	10/29/24 19:27		Jaswal	OK
25	P4322-08	MYDD01	SAM	10/29/24 19:30		Jaswal	OK
26	P4322-09	MYDD02	SAM	10/29/24 19:34		Jaswal	OK
27	P4322-10	MYDD03	SAM	10/29/24 19:37		Jaswal	OK
28	CCV030	CCV030	CCV	10/29/24 19:40		Jaswal	OK
29	CCB030	CCB030	CCB	10/29/24 19:43		Jaswal	OK
30	P4314-10DL	MYDB51	SAM	10/29/24 19:47	4X for Ba,Co,Ni,V	Jaswal	Confirms
31	P4314-11DL	MYDB51D	DUP	10/29/24 19:50	4X for Ba,Co,Ni,V	Jaswal	Confirms
32	P4314-10LDL	MYDB51L	SD	10/29/24 19:53	20X for Ba,Co,Ni,V	Jaswal	Confirms
33	P4314-12DL	MYDB51S	MS	10/29/24 19:56	4X for Ba,Co,Ni,V	Jaswal	Confirms
34	P4314-10MSDL	MYDB51SRE	MS	10/29/24 20:00	4X for Ba,Co,Ni,V	Jaswal	Confirms
35	P4314-10A	MYDB51ARE	PS	10/29/24 20:06	PS for As-0.0047ml(M5801),Pb-0.0038ml(M6026),Ag-0.02ml(MP82908),Ti-0.021ml(MP82960)-10ml	Jaswal	OK
36	P4316-16DL	MYDB88	SAM	10/29/24 20:09	4X for Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	M5801,MP82590,M5819,MP82908
LCS Standard	
Chk Standard	MP82587,MP82586

37	P4316-17DL	MYDB88D	DUP	10/29/24 20:12	4X for Ba,Co,Ni,V	Jaswal	Confirms
38	P4316-16LDL	MYDB88L	SD	10/29/24 20:16	20X for Ba,Co,Ni,V	Jaswal	Confirms
39	P4316-18DL	MYDB88S	MS	10/29/24 20:19	4X for Ba,Co,Ni,V	Jaswal	Confirms
40	P4316-16MSDL	MYDB88SRE	MS	10/29/24 20:22	4X for Ba,Co,Ni,V	Jaswal	Confirms
41	P4316-16A	MYDB88ARE	PS	10/29/24 20:25	PS for Ag-0.02ml(MP82908)-10ml	Jaswal	OK
42	P4317-09DL	MYDBA0	SAM	10/29/24 20:28	4X for Ba,Co,Ni,V	Jaswal	Confirms
43	P4317-10DL	MYDBA0D	DUP	10/29/24 20:32	4X for Ba,Co,Ni,V	Jaswal	Confirms
44	P4317-09LDL	MYDBA0L	SD	10/29/24 20:35	20X for Ba,Co,Ni,V	Jaswal	Confirms
45	P4317-11DL	MYDBA0S	MS	10/29/24 20:38	4X for Ba,Co,Ni,V	Jaswal	Confirms
46	P4317-09MSDL	MYDBA0SRE	MS	10/29/24 20:41	4X for Ba,Co,Ni,V	Jaswal	Confirms
47	P4317-08DL	MYDB99	SAM	10/29/24 20:45	Not Used	Jaswal	Not Ok
48	P4317-09A	MYDBA0ARE	PS	10/29/24 20:48	PS for Pb-0.0034ml(M6026),Ag-0.02ml(MP82908)-10ml	Jaswal	OK
49	P4318-20DL	MYDB72	SAM	10/29/24 20:51	4X for Ba,Co,Ni,V	Jaswal	Confirms
50	P4318-21DL	MYDB72D	DUP	10/29/24 20:54	4X for Ba,Co,Ni,V	Jaswal	Confirms
51	P4318-20LDL	MYDB72L	SD	10/29/24 20:57	4X for Ba,Co,Ni,V	Jaswal	Confirms
52	P4318-22DL	MYDB72S	MS	10/29/24 21:01	4X for Ba,Co,Ni,V	Jaswal	Confirms
53	P4318-20MSDL	MYDB72SRE	MS	10/29/24 21:04	4X for Ba,Co,Ni,V	Jaswal	Confirms
54	P4318-20A	MYDB72ARE	PS	10/29/24 21:07	PS for Ag-0.02ml(MP82908)-10ml	Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard	M5801,MP82590,M5819,MP82908		
LCS Standard			
Chk Standard	MP82587,MP82586		

55	P4317-08	MYDB99	SAM	10/29/24 21:10		Jaswal	OK
56	CCV031	CCV031	CCV	10/29/24 21:13		Jaswal	OK
57	CCB031	CCB031	CCB	10/29/24 21:16		Jaswal	OK
58	PB164279BS	LCS279	LCS	10/29/24 21:20		Jaswal	OK
59	PB164279BL	PBS279	MB	10/29/24 21:33		Jaswal	OK
60	PB164279BSD	LCSD279	LCSD	10/29/24 21:36		Jaswal	OK
61	P4331-10	MYDBC1	SAM	10/29/24 21:40		Jaswal	OK
62	P4331-11	MYDBC2	SAM	10/29/24 21:43		Jaswal	OK
63	P4331-12	MYDBC3	SAM	10/29/24 21:46		Jaswal	OK
64	P4331-13	MYDBC4	SAM	10/29/24 21:49		Jaswal	OK
65	P4331-14	MYDBC5	SAM	10/29/24 21:52		Jaswal	OK
66	P4331-15	MYDBC6	SAM	10/29/24 21:55		Jaswal	OK
67	P4331-16	MYDBC7	SAM	10/29/24 21:58		Jaswal	OK
68	P4331-17	MYDBC8	SAM	10/29/24 22:02		Jaswal	OK
69	P4331-18	MYDBC9	SAM	10/29/24 22:05		Jaswal	OK
70	P4331-19	MYDBD0	SAM	10/29/24 22:08		Jaswal	OK
71	P4331-20	MYDBD1	SAM	10/29/24 22:11		Jaswal	OK
72	P4331-21	MYDBD2	SAM	10/29/24 22:14		Jaswal	OK
73	P4331-22	MYDBD3	SAM	10/29/24 22:17		Jaswal	OK
74	PB164281BL	PBS281	MB	10/29/24 22:20		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	M5801,MP82590,M5819,MP82908
LCS Standard	
Chk Standard	MP82587,MP82586

75	PB164281BS	LCS281	LCS	10/29/24 22:23		Jaswal	OK
76	CCV032	CCV032	CCV	10/29/24 23:07		Jaswal	OK
77	CCB032	CCB032	CCB	10/29/24 23:10		Jaswal	OK
78	P4332-08	MYDBE1	SAM	10/29/24 23:13	INT_STD 209Bi(1) Fail	Jaswal	Dilution
79	P4332-09	MYDBE2	SAM	10/29/24 23:16		Jaswal	OK
80	P4332-10	MYDBE3	SAM	10/29/24 23:19		Jaswal	OK
81	P4332-11	MYDBE4	SAM	10/29/24 23:22	INT_STD 209Bi(1) Fail	Jaswal	Dilution
82	P4332-12	MYDBE5	SAM	10/29/24 23:25	INT_STD 209Bi(1) Fail	Jaswal	Dilution
83	P4332-13	MYDBE6	SAM	10/29/24 23:28	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
84	P4332-14	MYDBE6D	DUP	10/29/24 23:31	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
85	P4332-13L	MYDBE6L	SD	10/29/24 23:34	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
86	P4332-15	MYDBE6S	MS	10/29/24 23:38	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
87	P4332-13MS	MYDBE6SRE	MS	10/29/24 23:41	Ba,Co,Ni,V High	Jaswal	Dilution
88	P4332-16	MYDBE7	SAM	10/29/24 23:44		Jaswal	OK
89	P4332-17	MYDBE8	SAM	10/29/24 23:47		Jaswal	OK
90	P4332-18	MYDBE9	SAM	10/29/24 23:50		Jaswal	OK
91	P4332-19	MYDBF0	SAM	10/29/24 23:53		Jaswal	OK
92	P4332-20	MYDBF1	SAM	10/29/24 23:56	INT_STD 209Bi(1) Fail	Jaswal	Dilution
93	P4332-21	MYDBF2	SAM	10/29/24 23:59		Jaswal	OK
94	P4332-22	MYDBF3	SAM	10/30/24 00:02	INT_STD 209Bi(1) Fail	Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard	M5801,MP82590,M5819,MP82908		
LCS Standard			
Chk Standard	MP82587,MP82586		

95	P4333-01	MYDBF4	SAM	10/30/24 00:05		Jaswal	OK
96	P4333-02	MYDBF5	SAM	10/30/24 00:08	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
97	P4333-03	MYDBF5D	DUP	10/30/24 00:11	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
98	P4333-02L	MYDBF5L	SD	10/30/24 00:14	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
99	P4333-04	MYDBF5S	MS	10/30/24 00:17	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
100	P4333-02MS	MYDBF5SRE	MS	10/30/24 00:20	Ba,Co,ni,V High	Jaswal	Dilution
101	P4332-13DL	MYDBE6	SAM	10/30/24 00:23	4X For Ba,Co,Ni,V	Jaswal	Confirms
102	P4332-14DL	MYDBE6D	DUP	10/30/24 00:26	4X For Ba,Co,Ni,V	Jaswal	Confirms
103	P4332-13LDL	MYDBE6L	SD	10/30/24 00:29	20X For Ba,Co,Ni,V	Jaswal	Confirms
104	CCV033	CCV033	CCV	10/30/24 00:33		Jaswal	OK
105	CCB033	CCB033	CCB	10/30/24 00:36		Jaswal	OK
106	P4332-15DL	MYDBE6S	MS	10/30/24 00:39	4X For Ba,Co,Ni,V	Jaswal	Confirms
107	P4332-13MSDL	MYDBE6SRE	MS	10/30/24 00:42	4X For Ba,Co,Ni,V	Jaswal	Confirms
108	P4333-02DL	MYDBF5	SAM	10/30/24 00:45	4X For Ba,Co,Ni,V	Jaswal	Confirms
109	P4333-03DL	MYDBF5D	DUP	10/30/24 00:48	4X For Ba,Co,Ni,V	Jaswal	Confirms
110	P4333-02LDL	MYDBF5L	SD	10/30/24 00:51	20X For Ba,Co,Ni,V	Jaswal	Confirms
111	P4333-04DL	MYDBF5S	MS	10/30/24 00:55	4X For Ba,Co,Ni,V	Jaswal	Confirms
112	P4333-02MSDL	MYDBF5SRE	MS	10/30/24 00:58	4X For Ba,Co,Ni,V	Jaswal	Confirms
113	PB164281BSD	LCSD281	LCSD	10/30/24 01:01		Jaswal	OK
114	P4332-01	MYDBD4	SAM	10/30/24 01:04	INT_STD 209Bi(1) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard	M5801,MP82590,M5819,MP82908		
LCS Standard			
Chk Standard	MP82587,MP82586		

115	P4332-02	MYDBD5	SAM	10/30/24 01:07	INT_STD 209Bi(1) Fail	Jaswal	Dilution
116	P4332-03	MYDBD6	SAM	10/30/24 01:10	INT_STD 209Bi(1) Fail, Cu High	Jaswal	Dilution
117	P4332-04	MYDBD7	SAM	10/30/24 01:13	INT_STD 209Bi(1) Fail	Jaswal	Dilution
118	P4332-05	MYDBD8	SAM	10/30/24 01:16	INT_STD 209Bi(1) Fail,Cu,Zn High	Jaswal	Dilution
119	P4332-06	MYDBD9	SAM	10/30/24 01:19		Jaswal	OK
120	P4332-07	MYDBE0	SAM	10/30/24 01:30		Jaswal	OK
121	P4332-08DL	MYDBE1	SAM	10/30/24 01:33	2X For INT_STD209Bi(1)	Jaswal	Confirms
122	P4332-11DL	MYDBE4	SAM	10/30/24 01:36	2X For INT_STD209Bi(1)	Jaswal	Confirms
123	P4332-12DL	MYDBE5	SAM	10/30/24 01:39	2X For INT_STD209Bi(1)	Jaswal	Confirms
124	P4332-20DL	MYDBF1	SAM	10/30/24 01:42	2X For INT_STD209Bi(1)	Jaswal	Confirms
125	P4332-22DL	MYDBF3	SAM	10/30/24 01:45	2X For INT_STD 209Bi(1), Still Fail so Report From Original	Jaswal	Not Ok
126	P4332-13A	MYDBE6ARE	PS	10/30/24 01:48	PS for As,Se,Ag,Zn - As - 0.008-M5801, Se- 0.016 - MP82590 , Ag - 0.024 - MP82908 , Zn - 0.052 - M5819, AS High	Jaswal	Dilution
127	P4332-01DL	MYDBD4	SAM	10/30/24 01:51	2X For INT_STD209Bi(1)	Jaswal	Confirms
128	P4332-02DL	MYDBD5	SAM	10/30/24 01:54	2X For INT_STD209Bi(1)	Jaswal	Confirms
129	P4332-03DL	MYDBD6	SAM	10/30/24 01:57	2X for Cu , 2X For INT_STD 209Bi(1), Still Fail so Report From Original	Jaswal	Confirms
130	P4332-04DL	MYDBD7	SAM	10/30/24 02:00	2X For INT_STD209Bi(1)	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133196

Review By	Sarabjit Jaswal	Review On	10/29/2024 8:26:02 PM
Supervise By	Mohan Bera	Supervise On	10/29/2024 8:26:44 PM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard	M5801,MP82590,M5819,MP82908		
LCS Standard			
Chk Standard	MP82587,MP82586		

131	P4332-05DL	MYDBD8	SAM	10/30/24 02:03	2X for Cu ,Zn , 2X For INT_STD 209Bi(1), Still Fail so Report From Original	Jaswal	Confirms
132	P4332-06DL	MYDBD9	SAM	10/30/24 02:06	Not Required	Jaswal	Not Ok
133	P4332-07DL	MYDBE0	SAM	10/30/24 02:09	Not Required	Jaswal	Not Ok
134	P4332-13ADL	MYDBE6ARE	PS	10/30/24 02:12	PS for As - 0.008-M5801	Jaswal	Confirms
135	CCV034	CCV034	CCV	10/30/24 02:15		Jaswal	OK
136	CCB034	CCB034	CCB	10/30/24 02:18		Jaswal	OK