

Continuing Calibration Report for SequenceID : FC101124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067426.D

Aliphatic C9-C12	8641376.000	60.000	3.175	6.457	144022.933	128354.488	-12.207
Aliphatic C12-C16	5667399.000	40.000	6.458	9.848	141684.975	130957.677	-8.191
Aliphatic C16-C21	8228693.000	60.000	9.849	13.207	137144.883	131476.102	-4.312
Aliphatic C21-C28	10542979.000	80.000	13.208	16.862	131787.238	126823.664	-3.914
Aliphatic C28-C40	13838283.000	120.000	16.863	21.712	115319.025	105215.848	-9.602
Aliphatic EPH	46918730.000	360.000	3.175	21.712	130329.806	121110.937	-7.612