

Continuing Calibration Report for SequenceID : FD101124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048454.D

Aromatic C10-C12	7047843.000	40.000	4.100	5.818	176196.075	180984.840	2.646
Aromatic C12-C16	10693300.000	60.000	5.819	8.425	178221.667	189995.313	6.197
Aromatic C16-C21	13124786.000	80.000	8.426	12.687	164059.825	178728.230	8.207
Aromatic C21-C36	26788273.000	180.000	12.688	18.089	148823.739	165360.750	10.001
Aromatic EPH	57654202.000	360.000	4.100	18.089	160150.561	174173.072	8.051