

Continuing Calibration Report for SequenceID : FD101124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048460.D

Aromatic C10-C12	7064482.000	40.000	4.100	5.818	176612.050	180984.840	2.416
Aromatic C12-C16	10873476.000	60.000	5.819	8.425	181224.600	189995.313	4.616
Aromatic C16-C21	13241236.000	80.000	8.426	12.687	165515.450	178728.230	7.393
Aromatic C21-C36	26858279.000	180.000	12.688	18.089	149212.661	165360.750	9.765
Aromatic EPH	58037473.000	360.000	4.100	18.089	161215.203	174173.072	7.440