

Continuing Calibration Report for SequenceID : FD101124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048472.D

Aromatic C10-C12	7272361.000	40.000	4.100	5.818	181809.025	180984.840	-0.455
Aromatic C12-C16	10992353.000	60.000	5.819	8.425	183205.883	189995.313	3.573
Aromatic C16-C21	13410430.000	80.000	8.426	12.687	167630.375	178728.230	6.209
Aromatic C21-C36	27444861.000	180.000	12.688	18.089	152471.450	165360.750	7.795
Aromatic EPH	59120005.000	360.000	4.100	18.089	164222.236	174173.072	5.713