

Continuing Calibration Report for SequenceID : FD101124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048475.D

Aliphatic C9-C12	9702351.000	60.000	3.024	6.263	161705.850	172191.705	6.090
Aliphatic C12-C16	6614425.000	40.000	6.264	9.634	165360.625	177348.983	6.760
Aliphatic C16-C21	9778931.000	60.000	9.635	12.975	162982.183	175681.675	7.229
Aliphatic C21-C28	12124048.000	80.000	12.976	16.615	151550.600	167581.496	9.566
Aliphatic C28-C40	13625838.000	120.000	16.616	21.317	113548.650	135864.687	16.425
Aliphatic EPH	51845593.000	360.000	3.024	21.317	144015.536	160212.901	10.110