

Continuing Calibration Report for SequenceID : FD101124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048485.D

Aliphatic C9-C12	9823013.000	60.000	3.024	6.263	163716.883	172191.705	4.922
Aliphatic C12-C16	6668002.000	40.000	6.264	9.634	166700.050	177348.983	6.005
Aliphatic C16-C21	9870135.000	60.000	9.635	12.975	164502.250	175681.675	6.363
Aliphatic C21-C28	12226799.000	80.000	12.976	16.615	152834.988	167581.496	8.800
Aliphatic C28-C40	13868377.000	120.000	16.616	21.317	115569.808	135864.687	14.938
Aliphatic EPH	52456326.000	360.000	3.024	21.317	145712.017	160212.901	9.051