

Continuing Calibration Report for SequenceID : FF101424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014664.D

Aliphatic C9-C12	8157028.000	60.000	3.093	6.705	135950.467	127380.686	-6.728
Aliphatic C12-C16	5580774.000	40.000	6.706	10.137	139519.350	129993.553	-7.328
Aliphatic C16-C21	8700186.000	60.000	10.138	13.497	145003.100	138557.475	-4.652
Aliphatic C21-C28	11095952.000	80.000	13.498	17.152	138699.400	144410.954	3.955
Aliphatic C28-C40	12003460.000	120.000	17.153	21.966	100028.833	118299.321	15.444
Aliphatic EPH	45537400.000	360.000	3.093	21.966	126492.778	130291.185	2.915