

Continuing Calibration Report for SequenceID : FF101424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014680.D

Aliphatic C9-C12	8183464.000	60.000	3.093	6.705	136391.067	127380.686	-7.074
Aliphatic C12-C16	5588021.000	40.000	6.706	10.137	139700.525	129993.553	-7.467
Aliphatic C16-C21	8696082.000	60.000	10.138	13.497	144934.700	138557.475	-4.603
Aliphatic C21-C28	11090672.000	80.000	13.498	17.152	138633.400	144410.954	4.001
Aliphatic C28-C40	12010115.000	120.000	17.153	21.966	100084.292	118299.321	15.397
Aliphatic EPH	45568354.000	360.000	3.093	21.966	126578.761	130291.185	2.849