

Continuing Calibration Report for SequenceID : FF101524AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014683.D

Aromatic C10-C12	5236803.000	40.000	4.369	6.234	130920.075	130059.006	-0.662
Aromatic C12-C16	8014629.000	60.000	6.235	8.923	133577.150	130536.771	-2.329
Aromatic C16-C21	10468566.000	80.000	8.924	13.226	130857.075	119247.507	-9.736
Aromatic C21-C36	21763493.000	180.000	13.227	18.657	120908.294	109724.219	-10.193
Aromatic EPH	45483491.000	360.000	4.369	18.657	126343.031	117568.685	-7.463