

Continuing Calibration Report for SequenceID : FF101524AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014693.D

Aromatic C10-C12	5248496.000	40.000	4.369	6.234	131212.400	130059.006	-0.887
Aromatic C12-C16	8095979.000	60.000	6.235	8.923	134932.983	130536.771	-3.368
Aromatic C16-C21	10512756.000	80.000	8.924	13.226	131409.450	119247.507	-10.199
Aromatic C21-C36	21776262.000	180.000	13.227	18.657	120979.233	109724.219	-10.258
Aromatic EPH	45633493.000	360.000	4.369	18.657	126759.703	117568.685	-7.818