

Continuing Calibration Report for SequenceID : FG101424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014434.D

Aliphatic C9-C12	7192876.000	60.000	3.081	6.702	119881.267	125507.047	4.482
Aliphatic C12-C16	4931059.000	40.000	6.703	10.146	123276.475	122478.062	-0.652
Aliphatic C16-C21	7842487.000	60.000	10.147	13.516	130708.117	124879.402	-4.667
Aliphatic C21-C28	10143074.000	80.000	13.517	17.186	126788.425	118749.663	-6.770
Aliphatic C28-C40	11443290.000	120.000	17.187	22.050	95360.750	93985.220	-1.464
Aliphatic EPH	41552786.000	360.000	3.081	22.050	115424.406	113056.969	-2.094