

Continuing Calibration Report for SequenceID : FG101424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014444.D

Aliphatic C9-C12	7807195.000	60.000	3.081	6.702	130119.917	125507.047	-3.675
Aliphatic C12-C16	5348678.000	40.000	6.703	10.146	133716.950	122478.062	-9.176
Aliphatic C16-C21	8472466.000	60.000	10.147	13.516	141207.767	124879.402	-13.075
Aliphatic C21-C28	10960338.000	80.000	13.517	17.186	137004.225	118749.663	-15.372
Aliphatic C28-C40	12361903.000	120.000	17.187	22.050	103015.858	93985.220	-9.609
Aliphatic EPH	44950580.000	360.000	3.081	22.050	124862.722	113056.969	-10.442