

Continuing Calibration Report for SequenceID : FG101424AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FG014450.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 7296251.000  | 60.000  | 3.081  | 6.702  | 121604.183 | 125507.047 | 3.110  |
| Aliphatic C12-C16 | 5008469.000  | 40.000  | 6.703  | 10.146 | 125211.725 | 122478.062 | -2.232 |
| Aliphatic C16-C21 | 7965345.000  | 60.000  | 10.147 | 13.516 | 132755.750 | 124879.402 | -6.307 |
| Aliphatic C21-C28 | 10332557.000 | 80.000  | 13.517 | 17.186 | 129156.963 | 118749.663 | -8.764 |
| Aliphatic C28-C40 | 11670513.000 | 120.000 | 17.187 | 22.050 | 97254.275  | 93985.220  | -3.478 |
| Aliphatic EPH     | 42273135.000 | 360.000 | 3.081  | 22.050 | 117425.375 | 113056.969 | -3.864 |