

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>CHEM02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>P4368</u>	SAS No.:	<u>P4368</u>	SDG NO.:	<u>P4368</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>10/07/2024</u>	<u>10/07/2024</u>		
		Calibration Times:		<u>11:12</u>	<u>12:47</u>		

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:	RT 200 =	<u>PS027855.D</u>	RT 500 =	<u>PS027856.D</u>	
RT 750 =	PS027857.D	RT 1000 =	PS027858.D	RT 1500 =	PS027859.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Instrument ID: ECD_S Calibration Date(s): 10/07/2024 10/07/2024

Calibration Times: 11:12 12:47

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS027855.D</u>	RT 500 = <u>PS027856.D</u>
	RT 750 = <u>PS027857.D</u>	RT 1000 = <u>PS027858.D</u>
		RT 1500 = <u>PS027859.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW	
							FROM	TO
2,4,5-T	10.16	10.16	10.16	10.16	10.16	10.16	10.06	10.26
2,4,5-TP(Silvex)	9.74	9.74	9.74	9.75	9.74	9.74	9.64	9.84
2,4-D	8.85	8.85	8.85	8.85	8.85	8.85	8.75	8.95
2,4-DB	10.72	10.72	10.72	10.72	10.72	10.72	10.62	10.82
2,4-DCAA	7.63	7.63	7.63	7.63	7.63	7.63	7.53	7.73
DICAMBA	7.82	7.82	7.82	7.82	7.82	7.82	7.72	7.92
DICHLORPROP	8.53	8.53	8.53	8.53	8.53	8.53	8.43	8.63
Dinoseb	11.09	11.09	11.09	11.10	11.10	11.09	10.99	11.19