

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** CHEM02

**Lab Code:** CHEM      **Case No.:** P4368      **SAS No.:** P4368      **SDG NO.:** P4368

**Continuing Calib Date:** 10/15/2024      **Initial Calibration Date(s):** 10/07/2024      10/07/2024

**Continuing Calib Time:** 20:32      **Initial Calibration Time(s):** 11:12      12:47

**GC Column:** RTX-CLP      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 7.27       | 7.28      | 7.18      | 7.38  | 0.01       |
| 2,4-DCAA         | 7.10       | 7.10      | 7.00      | 7.20  | 0.01       |
| DICHLORPROP      | 7.96       | 7.96      | 7.86      | 8.06  | 0.00       |
| 2,4-D            | 8.18       | 8.19      | 8.09      | 8.29  | 0.01       |
| 2,4,5-TP(Silvex) | 9.03       | 9.04      | 8.94      | 9.14  | 0.01       |
| 2,4,5-T          | 9.32       | 9.32      | 9.22      | 9.42  | 0.00       |
| 2,4-DB           | 9.88       | 9.89      | 9.79      | 9.99  | 0.01       |
| Dinoseb          | 11.06      | 11.06     | 10.96     | 11.16 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/15/2024 Initial Calibration Date(s): 10/07/2024 10/07/2024

Continuing Calib Time: 20:32 Initial Calibration Time(s): 11:12 12:47

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.82       | 7.82      | 7.72                   | 7.92  | 0.00       |
| 2,4-DCAA         | 7.62       | 7.63      | 7.53                   | 7.73  | 0.01       |
| DICHLORPROP      | 8.52       | 8.53      | 8.43                   | 8.63  | 0.01       |
| 2,4-D            | 8.85       | 8.85      | 8.75                   | 8.95  | 0.00       |
| 2,4,5-TP(Silvex) | 9.74       | 9.74      | 9.64                   | 9.84  | 0.00       |
| 2,4,5-T          | 10.15      | 10.16     | 10.06                  | 10.26 | 0.01       |
| 2,4-DB           | 10.71      | 10.72     | 10.62                  | 10.82 | 0.01       |
| Dinoseb          | 11.09      | 11.09     | 10.99                  | 11.19 | 0.00       |