

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL01 Date Analyzed: 10/15/2024

Lab Sample No.: HSTDCCC750 Data File : PS027942.D Time Analyzed: 12:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.319	9.223	9.423	748.220	712.500	5.0
2,4,5-TP(Silvex)	9.034	8.938	9.138	748.150	712.500	5.0
2,4-D	8.182	8.085	8.285	715.690	705.000	1.5
2,4-DB	9.882	9.785	9.985	730.180	712.500	2.5
2,4-DCAA	7.096	6.998	7.198	766.400	750.000	2.2
DICAMBA	7.275	7.177	7.377	730.490	705.000	3.6
DICHLORPROP	7.959	7.863	8.063	727.310	705.000	3.2
Dinoseb	11.057	10.961	11.161	753.590	705.000	6.9

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL01 Date Analyzed: 10/15/2024

Lab Sample No.: HSTDCCC750 Data File : PS027942.D Time Analyzed: 12:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.151	10.058	10.258	692.240	712.500	-2.8
2,4,5-TP(Silvex)	9.737	9.643	9.843	693.500	712.500	-2.7
2,4-D	8.847	8.752	8.952	696.520	705.000	-1.2
2,4-DB	10.713	10.619	10.819	703.630	712.500	-1.2
2,4-DCAA	7.623	7.528	7.728	722.610	750.000	-3.7
DICAMBA	7.817	7.722	7.922	684.460	705.000	-2.9
DICHLORPROP	8.522	8.428	8.628	649.960	705.000	-7.8
Dinoseb	11.087	10.993	11.193	724.580	705.000	2.8