

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL02 Date Analyzed: 10/15/2024

Lab Sample No.: HSTDCCC750 Data File : PS027951.D Time Analyzed: 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.319	9.223	9.423	763.270	712.500	7.1
2,4,5-TP(Silvex)	9.034	8.938	9.138	758.040	712.500	6.4
2,4-D	8.182	8.085	8.285	732.680	705.000	3.9
2,4-DB	9.881	9.785	9.985	741.950	712.500	4.1
2,4-DCAA	7.095	6.998	7.198	770.230	750.000	2.7
DICAMBA	7.274	7.177	7.377	750.200	705.000	6.4
DICHLORPROP	7.960	7.863	8.063	733.300	705.000	4.0
Dinoseb	11.057	10.961	11.161	755.810	705.000	7.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/07/2024 10/07/2024

Client Sample No.: CCAL02 Date Analyzed: 10/15/2024

Lab Sample No.: HSTDCCC750 Data File : PS027951.D Time Analyzed: 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.150	10.058	10.258	695.910	712.500	-2.3
2,4,5-TP(Silvex)	9.737	9.643	9.843	670.200	712.500	-5.9
2,4-D	8.847	8.752	8.952	689.030	705.000	-2.3
2,4-DB	10.713	10.619	10.819	699.920	712.500	-1.8
2,4-DCAA	7.622	7.528	7.728	653.450	750.000	-12.9
DICAMBA	7.816	7.722	7.922	696.100	705.000	-1.3
DICHLORPROP	8.522	8.428	8.628	644.100	705.000	-8.6
Dinoseb	11.087	10.993	11.193	675.790	705.000	-4.1