

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

LOD-MDL-WATER-01-Q

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Lab Sample ID: P4368-07 Date(s) Analyzed: 10/16/2024 10/16/2024

Instrument ID (1): ECD_R Instrument ID (2): ECD_R

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PR068718.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.869	4.819	4.919	0.24	0.23	0.04
COLUMN 1	2	4.891	4.841	4.941	0.22		
	3	4.955	4.905	5.005	0.23		
	4	5.058	5.008	5.108	0.22		
	5	5.372	5.322	5.422	0.24		
	1	4.103	4.053	4.153	0.23		
COLUMN 2	2	4.121	4.071	4.171	0.23	0.23	
	3	4.301	4.251	4.351	0.22		
	4	4.353	4.303	4.403	0.22		
	5	4.571	4.521	4.621	0.25		
Aroclor-1260	1	6.583	6.533	6.633	0.25	0.24	2.12
COLUMN 1	2	6.867	6.817	6.917	0.24		
	3	7.255	7.205	7.305	0.26		
	4	7.498	7.448	7.548	0.22		
	5	7.836	7.786	7.886	0.22		
	1	5.671	5.621	5.721	0.26		
COLUMN 2	2	5.877	5.827	5.927	0.25	0.23	
	3	6.036	5.986	6.086	0.22		
	4	6.545	6.495	6.595	0.23		
	5	6.81	6.76	6.86	0.21		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

LOD-MDL-WATER-01-Q

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Lab Sample ID: P4368-07 Date(s) Analyzed: 10/16/2024 10/16/2024

Instrument ID (1): ECD_R Instrument ID (2): ECD_R

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PR068719.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1221	1	3.869	3.819	3.919	0.23		
	2	3.959	3.909	4.009	0.45		
	3	4.036	3.986	4.086	0.21		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 1	1	3.205	3.155	3.255	0.18	0.30	
	2	3.286	3.236	3.336	0.38		
	3	3.367	3.317	3.417	0.21		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 2	1	3.205	3.155	3.255	0.18		
	2	3.286	3.236	3.336	0.38		
	3	3.367	3.317	3.417	0.21		
	4	0	0	0	0		
	5	0	0	0	0		
						0.25	14.66