

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/15/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 05:37 Initial Calibration Time(s): 13:03 21:11

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1016-2	(2)	5.55	5.55	5.45	5.65	0.00
Aroclor-1016-3	(3)	5.61	5.61	5.51	5.71	0.00
Aroclor-1016-4	(4)	5.71	5.71	5.61	5.81	0.01
Aroclor-1016-5	(5)	6.00	6.00	5.90	6.10	0.00
Aroclor-1260-1	(1)	7.13	7.12	7.02	7.22	-0.01
Aroclor-1260-2	(2)	7.38	7.38	7.28	7.48	0.00
Aroclor-1260-3	(3)	7.74	7.74	7.64	7.84	0.00
Aroclor-1260-4	(4)	7.97	7.97	7.87	8.07	0.00
Aroclor-1260-5	(5)	8.28	8.28	8.18	8.38	0.00
Tetrachloro-m-xylene		4.38	4.38	4.28	4.48	0.00
Decachlorobiphenyl		10.06	10.06	9.96	10.16	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/15/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 05:37 Initial Calibration Time(s): 13:03 21:11

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-3	(3)	4.92	4.93	4.83	5.03	0.01
Aroclor-1016-4	(4)	4.96	4.97	4.87	5.07	0.01
Aroclor-1016-5	(5)	5.18	5.18	5.08	5.28	0.00
Aroclor-1260-1	(1)	6.21	6.21	6.11	6.31	0.00
Aroclor-1260-2	(2)	6.40	6.40	6.30	6.50	0.01
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.01
Decachlorobiphenyl		8.64	8.65	8.55	8.75	0.01