

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/15/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 08:37 Initial Calibration Time(s): 13:03 21:11

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1016-1       | (1) | 5.52       | 5.52      | 5.42                   | 5.62  | 0.00       |
| Aroclor-1016-2       | (2) | 5.54       | 5.55      | 5.45                   | 5.65  | 0.01       |
| Aroclor-1016-3       | (3) | 5.61       | 5.61      | 5.51                   | 5.71  | 0.00       |
| Aroclor-1016-4       | (4) | 5.70       | 5.71      | 5.61                   | 5.81  | 0.01       |
| Aroclor-1016-5       | (5) | 6.00       | 6.00      | 5.90                   | 6.10  | 0.00       |
| Aroclor-1260-1       | (1) | 7.12       | 7.12      | 7.02                   | 7.22  | 0.00       |
| Aroclor-1260-2       | (2) | 7.38       | 7.38      | 7.28                   | 7.48  | 0.00       |
| Aroclor-1260-3       | (3) | 7.74       | 7.74      | 7.64                   | 7.84  | 0.00       |
| Aroclor-1260-4       | (4) | 7.97       | 7.97      | 7.87                   | 8.07  | 0.00       |
| Aroclor-1260-5       | (5) | 8.28       | 8.28      | 8.18                   | 8.38  | 0.00       |
| Tetrachloro-m-xylene |     | 4.37       | 4.38      | 4.28                   | 4.48  | 0.01       |
| Decachlorobiphenyl   |     | 10.06      | 10.06     | 9.96                   | 10.16 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4368 SAS No.: P4368 SDG NO.: P4368

Continuing Calib Date: 10/15/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 08:37 Initial Calibration Time(s): 13:03 21:11

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.73       | 4.73      | 4.63      | 4.83 | 0.01       |
| Aroclor-1016-2       | (2) | 4.75       | 4.75      | 4.65      | 4.85 | 0.01       |
| Aroclor-1016-3       | (3) | 4.92       | 4.93      | 4.83      | 5.03 | 0.01       |
| Aroclor-1016-4       | (4) | 4.96       | 4.97      | 4.87      | 5.07 | 0.01       |
| Aroclor-1016-5       | (5) | 5.18       | 5.18      | 5.08      | 5.28 | 0.00       |
| Aroclor-1260-1       | (1) | 6.21       | 6.21      | 6.11      | 6.31 | 0.00       |
| Aroclor-1260-2       | (2) | 6.39       | 6.40      | 6.30      | 6.50 | 0.01       |
| Aroclor-1260-3       | (3) | 6.55       | 6.55      | 6.45      | 6.65 | 0.00       |
| Aroclor-1260-4       | (4) | 7.02       | 7.02      | 6.92      | 7.12 | 0.00       |
| Aroclor-1260-5       | (5) | 7.26       | 7.26      | 7.16      | 7.36 | 0.00       |
| Tetrachloro-m-xylene |     | 3.65       | 3.65      | 3.55      | 3.75 | 0.01       |
| Decachlorobiphenyl   |     | 8.64       | 8.65      | 8.55      | 8.75 | 0.01       |